

Data Path : Z:\HPCHEM1\BNA_M\Data\BM072916\
 Data File : BM006734.D
 Acq On : 29 Jul 2016 11:32
 Operator : UM/SJ
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.456

Quant Time: Jul 29 12:44:26 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM072816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 28 15:44:03 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
3	Naphthalene	1.007	1.023	-1.6	106	0.00
4 SURR	2-Methylnaphthalene-d10	0.507	0.524	-3.4	108	0.00
5	2-Methylnaphthalene	0.683	0.689	-0.9	106	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
7	Acenaphthylene	2.296	2.295	0.0	105	0.00
8 C	Acenaphthene	1.390	1.432	-3.0	107	0.00
9	Fluorene	1.630	1.663	-2.0	105	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
11	Pentachlorophenol	0.050	0.041	18.0	103	0.00
12	Phenanthrene	1.112	1.144	-2.9	105	0.00
13	Anthracene	1.063	1.072	-0.8	105	0.00
14 SURR	Fluoranthene-d10	1.063	1.109	-4.3	111	0.00
15 C	Fluoranthene	1.549	1.552	-0.2	106	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
17	Pyrene	1.210	1.417	-17.1	108	0.00
18	Benzo(a)anthracene	1.097	1.267	-15.5	106	0.00
19	Chrysene	1.098	1.358	-23.7#	116	0.00
20 I	Perylene-d12	1.000	1.000	0.0	115	0.00
21	Benzo(b)fluoranthene	1.368	1.413	-3.3	114	0.00
22	Benzo(k)fluoranthene	1.301	1.447	-11.2	110	0.00
23 C	Benzo(a)pyrene	1.267	1.395	-10.1	114	0.00
24	Indeno(1,2,3-cd)pyrene	1.598	1.655	-3.6	107	0.00
25	Dibenzo(a,h)anthracene	1.198	1.386	-15.7	118	0.00
26	Benzo(g,h,i)perylene	1.287	1.411	-9.6	113	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0