


Data Path : Z:\HPCHEM1\BNA_M\Data\BM072916\
Data File : BM006731.D
Acq On : 29 Jul 2016 08:16
Operator : UM/SJ
Sample : DFTPP61
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP61

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM071416.M
Title : SVOA CALIBRATION
Last Update : Wed Jul 20 04:22:37 2016



AutoFind: Scans 2527, 2528, 2529; Background Corrected with Scan 2522

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.3	59837	PASS
68	69	0.00	2	1.5	803	PASS
69	198	0.00	100	38.5	54432	PASS
70	69	0.00	2	0.5	263	PASS
127	198	10	80	46.8	66178	PASS
197	198	0.00	2	0.5	728	PASS
198	198	100	100	100.0	141293	PASS
199	198	5	9	6.6	9267	PASS
275	198	10	60	25.2	35626	PASS
365	198	1	100	3.2	4574	PASS
441	443	0.01	100	82.9	18112	PASS
442	198	50	100	81.8	115581	PASS
443	442	15	24	18.9	21848	PASS

m/z	Abundance
37.00	362.0
38.00	885.0
39.00	4894.0
40.00	800.0
44.00	1101.0
49.00	775.0
50.00	16816.0
51.00	68216.0
52.10	3618.0
55.00	549.0
56.00	1916.0
57.00	4631.0
60.90	984.0
62.00	691.0
63.00	2499.0
64.00	363.0
65.00	1557.0
68.00	1052.0
69.00	61000.0
70.00	348.0
73.00	503.0
74.00	6380.0
75.00	10461.0
76.00	3414.0
77.10	72368.0
78.10	5371.0
79.00	3994.0
80.00	3021.0
81.00	4687.0
81.90	1363.0
83.00	1205.0
85.00	838.0
86.00	1599.0
87.00	672.0
88.10	300.0
91.00	1050.0
92.00	1169.0
93.00	8266.0
94.00	606.0
96.00	486.0
98.00	5925.0
99.00	5225.0
99.90	421.0
101.00	2596.0
103.00	860.0
104.00	1581.0
105.00	1603.0
106.00	668.0
107.00	23296.0
108.00	3548.0
109.00	632.0
110.00	36800.0
111.00	6067.0
112.00	621.0
116.00	1134.0
117.00	17744.0
118.00	1466.0
119.90	353.0
122.00	1262.0
123.00	1986.0
124.00	829.0
125.00	676.0
127.00	69464.0
128.00	5125.0
129.00	27216.0
130.00	1962.0
131.00	352.0
133.90	805.0
134.90	2018.0
136.00	1111.0
137.00	1239.0
141.00	3117.0
142.00	1094.0
142.90	776.0
144.00	354.0
146.00	739.0
147.00	1745.0
148.00	4060.0
149.00	1103.0
150.00	558.0
151.00	567.0
153.00	1014.0
154.00	799.0
155.00	2397.0
156.00	2842.0
157.10	616.0
157.90	579.0
159.00	623.0
160.10	823.0
161.00	1539.0
162.10	374.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61

165.00	1106.0
166.00	1013.0
167.00	7305.0
168.00	5198.0
169.00	969.0
171.00	337.0
172.00	605.0
173.00	770.0
174.00	1735.0
175.10	2850.0
175.90	776.0
177.00	1495.0
177.90	627.0
179.00	5338.0
180.00	4196.0
181.00	1504.0
185.00	2284.0
186.00	20032.0
187.00	5610.0
188.00	820.0
188.90	1120.0
190.90	555.0
191.90	2054.0
193.00	2005.0
193.90	413.0
196.00	3062.0
197.10	809.0
198.00	140736.0
199.00	9489.0
200.00	973.0
201.60	383.0
203.00	876.0
204.00	4980.0
205.00	8341.0
206.00	37048.0
207.10	5907.0
207.90	1427.0
208.90	626.0
210.00	716.0
211.00	1298.0
214.80	339.0
216.10	825.0
217.00	8997.0
218.00	1553.0
221.00	4788.0
221.80	1337.0
223.00	2366.0
224.00	20408.0
225.00	4918.0
226.00	405.0
227.00	9050.0
228.00	1519.0
229.00	2046.0
231.00	753.0
234.00	423.0
235.00	795.0
237.00	819.0
238.90	310.0
239.90	366.0
241.00	651.0
242.00	1400.0
243.00	1028.0
244.00	14354.0
245.00	2394.0
246.00	2926.0
246.90	643.0
249.00	436.0
253.00	545.0
254.10	515.0
255.00	78728.0
256.00	11605.0
257.00	895.0
258.00	4672.0
259.00	777.0
265.00	2027.0
273.00	2307.0
274.00	6256.0
275.00	32176.0
276.00	4525.0
277.00	2512.0
277.90	732.0
283.00	456.0
285.10	655.0
293.10	838.0
293.90	340.0
296.00	9524.0
297.00	1313.0
303.00	940.0
304.00	373.0
310.10	305.0
314.00	375.0
315.00	1175.0
316.00	666.0
323.10	2621.0
323.90	462.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61

326.90	597.0
328.10	344.0
334.00	1672.0
341.00	317.0
345.90	714.0
352.10	1070.0
352.90	499.0
354.00	1018.0
365.00	4063.0
366.00	608.0
371.90	1018.0
373.10	371.0
383.00	406.0
402.10	495.0
403.00	619.0
404.00	340.0
421.00	719.0
422.00	669.0
423.10	4493.0
424.10	1537.0
441.10	14034.0
442.10	86944.0
443.10	17488.0
444.00	1691.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61

m/z	Abundance
36.90	380.0
38.00	834.0
39.00	5028.0
40.00	1057.0
41.00	308.0
44.00	1143.0
49.00	711.0
50.00	16592.0
51.00	71552.0
52.00	3380.0
56.00	2042.0
57.00	4990.0
61.00	691.0
62.00	1134.0
63.00	2678.0
64.00	461.0
65.00	1354.0
68.00	885.0
69.00	66160.0
70.00	443.0
73.00	633.0
74.00	6910.0
75.00	10881.0
76.10	3928.0
77.10	79200.0
78.10	5190.0
79.00	4458.0
80.00	3621.0
81.00	5207.0
82.00	1254.0
83.00	1432.0
83.90	463.0
85.00	960.0
86.00	1704.0
87.10	760.0
91.00	1341.0
92.00	1137.0
93.00	8436.0
94.00	718.0
96.00	346.0
98.00	6649.0
99.00	6065.0
100.00	505.0
101.00	3390.0
102.90	1146.0
104.00	1569.0
105.00	1669.0
106.10	530.0
107.00	24864.0
108.00	3879.0
108.90	466.0
110.00	41640.0
111.00	6217.0
111.90	626.0
112.90	318.0
116.00	1167.0
117.00	19760.0
118.00	1679.0
119.00	334.0
120.00	430.0
122.00	1510.0
123.00	2340.0
124.00	1018.0
125.00	964.0
126.10	517.0
127.00	79008.0
128.00	5587.0
129.00	28440.0
130.00	2361.0
131.00	492.0
134.00	1092.0
135.00	2555.0
136.00	919.0
137.00	1411.0
138.10	397.0
139.90	443.0
141.00	3829.0
142.00	1171.0
143.00	970.0
144.10	347.0
146.00	790.0
147.00	1738.0
148.00	4861.0
149.00	968.0
151.00	752.0
152.00	326.0
152.90	1165.0
153.90	712.0
155.00	2871.0
156.00	3550.0
157.10	547.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61

157.90	637.0
158.90	568.0
160.00	1212.0
161.00	1858.0
162.00	463.0
163.90	359.0
164.90	1339.0
166.00	1106.0
167.00	7444.0
168.00	5024.0
169.00	764.0
171.00	322.0
171.90	735.0
173.00	910.0
174.00	1929.0
175.00	3470.0
176.00	966.0
177.00	1079.0
178.00	696.0
179.00	5845.0
180.00	4561.0
181.00	2340.0
184.00	676.0
185.00	3058.0
186.00	24344.0
187.00	6282.0
188.00	723.0
189.00	1907.0
191.00	910.0
191.90	1840.0
193.00	2222.0
194.00	542.0
196.00	4783.0
197.00	875.0
198.00	172032.0
199.00	11436.0
200.00	1021.0
201.30	586.0
203.00	1189.0
204.00	6231.0
205.00	10820.0
206.10	46752.0
207.00	6852.0
208.00	1834.0
208.80	415.0
210.00	729.0
210.90	1405.0
211.80	548.0
215.00	468.0
216.00	766.0
217.00	11455.0
218.00	1730.0
221.00	6132.0
221.90	1543.0
223.00	3206.0
224.00	24032.0
225.10	6301.0
226.10	549.0
227.00	10812.0
228.00	1314.0
229.00	2505.0
230.90	713.0
233.90	650.0
234.90	825.0
236.00	449.0
237.00	1143.0
239.00	373.0
241.00	479.0
242.00	1088.0
243.00	1444.0
244.10	20496.0
245.10	2560.0
246.00	3427.0
247.00	932.0
249.00	824.0
253.00	488.0
254.10	776.0
255.00	105680.0
256.00	15393.0
257.00	1111.0
258.00	5798.0
259.00	963.0
265.00	2214.0
265.80	354.0
270.90	355.0
271.90	320.0
273.00	3348.0
274.00	8147.0
275.00	43144.0
276.00	5884.0
277.00	4067.0
278.10	481.0
283.00	468.0
285.00	497.0
293.00	775.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61

296.00	13322.0
297.00	1711.0
300.80	318.0
303.00	1539.0
304.00	532.0
314.10	591.0
315.00	1444.0
316.00	780.0
321.10	354.0
323.00	3817.0
323.90	590.0
327.00	909.0
328.10	477.0
332.90	418.0
334.00	2564.0
335.00	743.0
340.90	404.0
346.00	671.0
352.00	1399.0
353.00	947.0
354.00	1123.0
365.00	5605.0
365.90	702.0
372.00	2038.0
373.00	418.0
383.00	559.0
390.10	317.0
402.10	806.0
403.00	1278.0
404.00	456.0
421.10	757.0
422.10	949.0
423.10	6690.0
424.00	2189.0
441.10	21824.0
442.10	138688.0
443.10	26320.0
444.10	2025.0

Instrument :

BNA_M

ClientSampleId :

DFTPP61

m/z	Abundance
38.00	560.0
39.00	2935.0
40.00	851.0
44.00	1222.0
49.00	621.0
50.00	9406.0
51.00	39744.0
52.00	2087.0
55.10	345.0
56.00	1222.0
57.00	2815.0
61.00	549.0
62.00	474.0
63.00	1717.0
65.00	837.0
68.00	473.0
69.00	36136.0
73.00	565.0
74.00	3976.0
75.00	6689.0
76.00	1849.0
77.00	43888.0
78.00	3166.0
79.00	2756.0
80.00	1908.0
81.00	3442.0
82.00	775.0
83.00	835.0
84.90	574.0
86.00	1231.0
86.90	408.0
90.90	839.0
92.00	958.0
93.00	4697.0
93.90	487.0
98.00	3310.0
99.00	3349.0
101.10	1735.0
102.90	350.0
104.00	1084.0
105.00	1484.0
106.10	446.0
107.00	15559.0
107.90	2023.0
109.00	442.0
110.00	25320.0
111.00	3585.0
111.90	406.0
115.90	643.0
117.00	11413.0
118.00	1073.0
120.10	331.0
122.00	687.0
122.90	1710.0
124.00	565.0
125.00	624.0
126.10	350.0
127.00	50064.0
128.00	3243.0
129.00	18912.0
130.00	1315.0
131.00	371.0
134.00	474.0
135.00	1698.0
136.00	750.0
137.00	857.0
141.00	2277.0
142.10	1004.0
143.00	358.0
145.90	426.0
147.00	1443.0
148.00	2730.0
148.90	899.0
153.00	700.0
153.90	893.0
154.90	1270.0
156.00	1906.0
157.90	582.0
159.00	395.0
160.00	860.0
161.00	933.0
164.90	1093.0
166.00	774.0
167.00	5342.0
168.00	3064.0
169.00	548.0
172.00	526.0
173.00	621.0
174.00	1280.0
175.00	2422.0
176.10	647.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61

176.90	841.0
178.00	308.0
178.90	4599.0
180.00	3200.0
181.10	1362.0
182.00	339.0
185.00	1923.0
186.00	15192.0
187.00	4566.0
188.10	638.0
189.00	930.0
190.90	674.0
192.00	1547.0
193.00	1736.0
194.00	423.0
196.00	2671.0
196.90	501.0
198.00	111112.0
199.00	6876.0
199.90	489.0
201.40	431.0
203.00	918.0
204.00	4084.0
205.10	6924.0
206.10	30344.0
207.00	4544.0
208.00	1233.0
209.00	438.0
210.00	329.0
211.00	1011.0
214.90	339.0
216.00	797.0
217.00	7902.0
218.00	1053.0
221.00	3553.0
221.90	942.0
223.00	1636.0
224.00	17520.0
225.00	4268.0
226.10	467.0
227.00	7457.0
228.00	931.0
229.00	1639.0
231.00	540.0
233.90	413.0
234.90	550.0
237.00	716.0
238.90	349.0
240.90	403.0
242.00	1098.0
243.00	933.0
244.00	13803.0
245.00	1864.0
246.00	3148.0
247.00	633.0
248.90	524.0
254.10	479.0
255.00	71248.0
256.00	9686.0
257.00	816.0
258.00	4196.0
259.00	497.0
265.00	1372.0
273.00	2294.0
274.00	5828.0
275.00	31560.0
276.00	4328.0
277.00	2714.0
277.90	442.0
285.00	437.0
292.90	611.0
296.00	9213.0
297.00	1048.0
303.00	949.0
304.00	363.0
314.10	718.0
314.90	938.0
316.00	516.0
323.10	2825.0
324.10	738.0
327.00	759.0
328.00	402.0
333.00	393.0
334.10	2088.0
335.10	518.0
341.00	312.0
345.90	537.0
352.10	810.0
353.00	706.0
354.10	962.0
365.00	4054.0
366.00	859.0
372.00	1352.0
373.00	335.0
382.90	500.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61

402.10	777.0
403.10	982.0
420.90	730.0
422.10	740.0
423.10	6366.0
424.00	1677.0
441.10	18480.0
442.10	121112.0
443.10	21736.0
444.10	2121.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61

Scan 2522 (17.421 min): BM006731.D
DFTPP61

m/z	Abundance
40.00	784.0
44.00	1468.0
49.00	361.0
73.00	497.0
83.90	309.0
207.00	990.0

Instrument :
BNA_M
ClientSampleId :
DFTPP61