

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM072919\
 Data File : BM021718.D
 Acq On : 30 Jul 2019 06:05
 Operator : HP/JU
 Sample : K3922-04
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A52F1

Manual Integrations
 APPROVED

mohammad
 7/30/2019 4:51:54 PM

Quant Time: Jul 30 07:20:54 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM072619MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 30 05:24:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.67	152	174142	20.00	ng/ul	0.00
18) Naphthalene-d8	10.46	136	703217	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.32	164	441858	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.07	188	873048	20.00	ng/ul	0.00
77) Chrysene-d12	21.27	240	746045	20.00	ng/ul	0.00
85) Perylene-d12	23.50	264	856328	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.23	96	11688	3.23	ng/uL	0.00
5) Phenol-d5	6.86	99	246416	18.81	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.03	67	145036	19.34	ng/ul	0.00
9) 2-Chlorophenol-d4	7.21	132	210694	19.39	ng/ul	0.00
13) 4-Methylphenol-d8	8.39	113	204080	18.57	ng/ul	0.00
19) Nitrobenzene-d5	8.84	128	106828	20.31	ng/ul	0.00
22) 2-Nitrophenol-d4	9.55	143	115795	19.99	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.09	165	239242	19.29	ng/ul	0.00
29) 4-Chloroaniline-d4	10.62	131	226898	19.34	ng/ul	0.00
43) Dimethylphthalate-d6	13.74	166	741255	21.74	ng/ul	0.00
46) Acenaphthylene-d8	14.02	160	854737	21.09	ng/ul	0.00
51) 4-Nitrophenol-d4	14.56	143	40106	7.75	ng/ul	0.01
57) Fluorene-d10	15.32	176	637083	21.29	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.46	200	48969	9.02	ng/ul	0.00
70) Anthracene-d10	17.17	188	947929	21.58	ng/ul	0.00
78) Pyrene-d10	19.47	212	986371	23.57	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.36	264	1035244	22.85	ng/ul	0.00

Target Compounds

					Ovalue
44) Dimethylphthalate	13.79	163	228162	6.446	ng/ul 98
69) Phenanthrene	17.12	178	921633	17.528	ng/ul 99
71) Anthracene	17.21	178	61841	1.148	ng/ul 98
74) Carbazole	17.49	167	58402	1.310	ng/ul 98
76) Fluoranthene	19.14	202	1430961	23.328	ng/ul 99
79) Pyrene	19.50	202	1082668	19.449	ng/ul 100
82) Benzo(a)anthracene	21.25	228	373537	6.876	ng/ul 98
84) Chrysene	21.30	228	484809	9.466	ng/ul 99
87) Benzo(b)fluoranthene	22.83	252	668774	11.478	ng/ul 99
88) Benzo(k)fluoranthene	22.87	252	194671m	3.418	ng/ul
90) Benzo(a)pyrene	23.40	252	396821	7.274	ng/ul 98
91) Indeno(1,2,3-cd)pyrene	25.74	276	319013	5.043	ng/ul 97
92) Dibenzo(a,h)anthracene	25.74	278	70099	1.333	ng/ul# 90
93) Benzo(g,h,i)perylene	26.43	276	316933	6.025	ng/ul 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM072919\
Data File : BM021718.D
Acq On : 30 Jul 2019 06:05
Operator : HP/JU
Sample : K3922-04
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampled :
A52F1

Manual Integrations
APPROVED

mohammad
7/30/2019 4:51:54 PM

Quant Time: Jul 30 07:20:54 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM072619MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 30 05:24:20 2019
Response via : Initial Calibration

