

(QT Reviewed)


```
Data Path   : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM072919\  
Data File  : BM021715.D  
Acq On     : 30 Jul 2019   04:16  
Operator   : HP/JU  
Sample     : K3922-01  
Misc       :  
ALS Vial   : 25      Sample Multiplier: 1
```

Instrument :
BNA_M
ClientSampleId :
A52F0

Manual Integrations
APPROVED

mohammad
7/30/2019 4:51:49 PM

Quant Time: Jul 30 06:38:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM072619MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 30 05:24:20 2019
Response via : Initial Calibration



Quantitation Report (Qedit)




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072919\
Data File : BM021714.D
Acq On : 30 Jul 2019 03:40
Operator : HP/JU
Sample : K3863-13
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A52F0

Manual Integrations
APPROVED

mohammad
7/30/2019 4:51:49 PM

Quant Time: Jul 30 04:44:40 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM072619MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jul 26 18:42:09 2019
Response via : Initial Calibration



TIC: BM021714.D

(51) 4-Nitrophenol-d4 (S)

14.592min (+0.035) 0.14ng/ul

response 781

Ion	Exp%	Act%
143.00	100	100
113.00	78.90	79.06
41.00	40.00	49.41#
42.00	27.20	0.00#

Quantitation Report (Qedit)




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072919\
 Data File : BM021714.D
 Acq On : 30 Jul 2019 03:40
 Operator : HP/JU
 Sample : K3863-13
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A52F0

Manual Integrations
 APPROVED

mohammad
 7/30/2019 4:51:49 PM

Quant Time: Jul 30 04:44:40 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM072619MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 26 18:42:09 2019
 Response via : Initial Calibration



TIC: BM021714.D

(51) 4-Nitrophenol-d4 (S)

14.592min (+0.035) 0.26ng/ul m

JU
 07/31/19

response 1385

Ion	Exp%	Act%
143.00	100	100
113.00	78.90	79.06
41.00	40.00	49.41#
42.00	27.20	0.00#

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM072919\
 Data File : BM021715.D
 Acq On : 30 Jul 2019 04:16
 Operator : HP/JU
 Sample : K3922-01
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
ClientSampleId :
 A52F0

Manual Integrations
APPROVED

mohammad
 7/30/2019 4:51:49 PM

Quant Time: Jul 30 06:26:19 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM072619MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 30 05:24:20 2019
 Response via : Initial Calibration



TIC: BM021715.D

(88) Benzo(k)fluoranthene

22.833min (-0.047) 5.65ng/ul

response 384748

Ion	Exp%	Act%
252.00	100	100
253.00	22.10	21.95
125.00	7.90	8.69
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM072919\
 Data File : BM021715.D
 Acq On : 30 Jul 2019 04:16
 Operator : HP/JU
 Sample : K3922-01
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A52F0

Manual Integrations
 APPROVED

mohammad
 7/30/2019 4:51:49 PM

Quant Time: Jul 30 06:26:19 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM072619MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 30 05:24:20 2019
 Response via : Initial Calibration



TIC: BM021715.D

(88) Benzo(k)fluoranthene *JU*
 22.874min (-0.006) 1.79ng/ul m *07/31/19*

response 122036

Ion	Exp%	Act%
252.00	100	100
253.00	22.10	23.50
125.00	7.90	8.98
0.00	0.00	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM072919\
 Data File : BM021715.D
 Acq On : 30 Jul 2019 04:16
 Operator : HP/JU
 Sample : K3922-01
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A52F0

Manual Integrations
 APPROVED

mohammad
 7/30/2019 4:51:49 PM

Quant Time: Jul 30 06:38:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM072619MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 30 05:24:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.67	152	160254	20.00	ng/ul	0.00
18) Naphthalene-d8	10.46	136	611761	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.32	164	367339	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.07	188	772328	20.00	ng/ul	0.00
77) Chrysene-d12	21.27	240	851063	20.00	ng/ul	0.00
85) Perylene-d12	23.50	264	1024218	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.23	96	8942	2.68	ng/uL	0.00
5) Phenol-d5	6.85	99	163730	13.58	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.03	67	102071	14.79	ng/ul	0.00
9) 2-Chlorophenol-d4	7.21	132	147253	14.72	ng/ul	0.00
13) 4-Methylphenol-d8	8.39	113	133849	13.23	ng/ul	0.00
19) Nitrobenzene-d5	8.84	128	69963	15.29	ng/ul	0.00
22) 2-Nitrophenol-d4	9.56	143	70463	13.98	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.09	165	159836	14.81	ng/ul	0.00
29) 4-Chloroaniline-d4	10.62	131	134585	13.19	ng/ul	0.00
43) Dimethylphthalate-d6	13.74	166	501005	17.67	ng/ul	0.00
46) Acenaphthylene-d8	14.02	160	582919	17.30	ng/ul	0.00
51) 4-Nitrophenol-d4	14.57	143	23311	5.42	ng/ul	0.02
57) Fluorene-d10	15.32	176	446409	17.95	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.46	200	29374	6.12	ng/ul	0.00
70) Anthracene-d10	17.17	188	693718	17.85	ng/ul	0.00
78) Pyrene-d10	19.47	212	857081	17.95	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.36	264	1022276	18.86	ng/ul	0.00

Target Compounds

					Qvalue
44) Dimethylphthalate	13.79	163	177364	6.027 ng/ul	99
69) Phenanthrene	17.12	178	185769	3.994 ng/ul	99
76) Fluoranthene	19.14	202	523951	9.655 ng/ul	99
79) Pyrene	19.50	202	432645	6.813 ng/ul	100
82) Benzo(a)anthracene	21.25	228	198105	3.196 ng/ul	98
84) Chrysene	21.30	228	254986	4.364 ng/ul	100
87) Benzo(b)fluoranthene	22.83	252	384748	5.521 ng/ul	99
88) Benzo(k)fluoranthene	22.87	252	122036m >	1.791 ng/ul	→ 99 07/31/19
90) Benzo(a)pyrene	23.40	252	219583	3.365 ng/ul	99
91) Indeno(1,2,3-cd)pyrene	25.74	276	181312	2.396 ng/ul#	95
93) Benzo(g,h,i)perylene	26.43	276	178883	2.843 ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed