

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM072919\
 Data File : BM021720.D
 Acq On : 30 Jul 2019 07:17
 Operator : HP/JU
 Sample : PB121643BL
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK43

Quant Time: Jul 30 08:15:18 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM072619MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 30 05:24:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.67	152	180985	20.00	ng/ul	0.00
18) Naphthalene-d8	10.46	136	740384	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.32	164	484102	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.07	188	1014416	20.00	ng/ul	0.00
77) Chrysene-d12	21.27	240	799633	20.00	ng/ul	0.00
85) Perylene-d12	23.50	264	866684	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.23	96	20783	5.52	ng/uL	0.00
5) Phenol-d5	6.86	99	407592	29.93	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.03	67	245872	31.55	ng/ul	0.00
9) 2-Chlorophenol-d4	7.21	132	357891	31.68	ng/ul	0.00
13) 4-Methylphenol-d8	8.39	113	349174	30.56	ng/ul	0.00
19) Nitrobenzene-d5	8.84	128	178458	32.23	ng/ul	0.00
22) 2-Nitrophenol-d4	9.55	143	201793	33.09	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.08	165	397624	30.45	ng/ul	0.00
29) 4-Chloroaniline-d4	10.61	131	488868	39.58	ng/ul	0.00
43) Dimethylphthalate-d6	13.74	166	1289259	34.51	ng/ul	0.00
46) Acenaphthylene-d8	14.02	160	1462590	32.93	ng/ul	0.00
51) 4-Nitrophenol-d4	14.55	143	140589	24.79	ng/ul	0.00
57) Fluorene-d10	15.32	176	1097258	33.48	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.46	200	133903	21.23	ng/ul	0.00
70) Anthracene-d10	17.17	188	1712629	33.55	ng/ul	0.00
78) Pyrene-d10	19.47	212	1736216	38.71	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.36	264	1572133	34.28	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM072919\
Data File : BM021720.D
Acq On : 30 Jul 2019 07:17
Operator : HP/JU
Sample : PB121643BL
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK43

Quant Time: Jul 30 08:15:18 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM072619MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 30 05:24:20 2019
Response via : Initial Calibration

