

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072922\
 Data File : BM036001.D
 Acq On : 29 Jul 2022 20:20
 Operator : CG/JU
 Sample : N3790-18
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DBUL4

Quant Time: Jul 30 00:24:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 30 00:20:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.875	152	316769	20.000	ng/ul	0.00
20) Naphthalene-d8	10.680	136	1304537	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.515	164	716768	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.251	188	1323344	20.000	ng/ul	0.00
79) Chrysene-d12	21.427	240	1222961	20.000	ng/ul	0.00
88) Perylene-d12	23.797	264	1245376	20.000	ng/ul	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.269	96	37116	4.409	ng/uL	0.00
4) Pyridine-d5	3.693	84	463198	20.085	ng/ul	0.00
7) Phenol-d5	7.040	99	644630	24.267	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.210	67	393779	22.236	ng/ul	0.00
11) 2-Chlorophenol-d4	7.404	132	539088	24.449	ng/ul	0.00
15) 4-Methylphenol-d8	8.586	113	505534	23.491	ng/ul	0.00
21) Nitrobenzene-d5	9.045	128	250720	24.254	ng/ul	0.00
24) 2-Nitrophenol-d4	9.763	143	257887	25.606	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.304	165	456349	25.053	ng/ul	0.00
31) 4-Chloroaniline-d4	10.822	131	694628	22.526	ng/ul	0.00
46) Dimethylphthalate-d6	13.927	166	1287857	26.377	ng/ul	0.00
49) Acenaphthylene-d8	14.210	160	1598106	25.654	ng/ul	0.00
54) 4-Nitrophenol-d4	14.710	143	171543	17.110	ng/ul	0.00
60) Fluorene-d10	15.504	176	1117178	25.321	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.621	200	116669	16.401	ng/ul	0.00
73) Anthracene-d10	17.351	188	1587377	26.532	ng/ul	0.00
81) Pyrene-d10	19.639	212	1742755	25.712	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.644	264	1688613	26.708	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM072922\
 Data File : BM036001.D
 Acq On : 29 Jul 2022 20:20
 Operator : CG/JU
 Sample : N3790-18
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DBUL4

Quant Time: Jul 30 00:24:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 30 00:20:59 2022
 Response via : Initial Calibration

