

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM073016\
 Data File : BM006754.D
 Acq On : 30 Jul 2016 02:41
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02059

Quant Time: Jul 30 03:16:10 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM072016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 30 01:37:26 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	94	0.00
2 S	1,4-Dioxane-d8	8.000	8.416	-5.2	98	0.00
3 S	Phenol-d5	20.000	20.651	-3.3	96	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	20.000	21.032	-5.2	96	0.00
5 S	2-Chlorophenol-d4	20.000	20.369	-1.8	96	0.00
6 S	4-Methylphenol-d8	20.000	20.897	-4.5	96	0.00
7 I	Naphthalene-d8	20.000	20.000	0.0	98	0.00
8 S	Nitrobenzene-d5	20.000	19.944	0.3	99	0.00
9 S	2-Nitrophenol-d4	20.000	20.863	-4.3	102	0.00
10 S	2,4-Dichlorophenol-d3	20.000	20.253	-1.3	99	0.00
11	Naphthalene	20.000	19.522	2.4	97	0.00
12 S	4-Chloroaniline-d4	20.000	21.861	-9.3	87	0.00
13	2-Methylnaphthalene	20.000	19.960	0.2	97	0.00
14	1-Methylnaphthalene	20.000	20.038	-0.2	98	0.00
15 I	Acenaphthene-d10	20.000	20.000	0.0	99	0.00
16 S	Dimethylphthalate-d6	20.000	20.286	-1.4	100	0.00
17 S	Acenaphthylene-d8	20.000	20.411	-2.1	101	0.00
18	Acenaphthylene	20.000	19.926	0.4	98	0.00
19 C	Acenaphthene	20.000	19.850	0.7	98	0.00
20 S	4-Nitrophenol-d4	20.000	19.098	4.5	96	0.00
21 S	Fluorene-d10	20.000	20.181	-0.9	100	0.00
22	Fluorene	20.000	20.486	-2.4	99	0.00
23 I	Phenanthrene-d10	20.000	20.000	0.0	100	0.00
24 S	4,6-Dinitro-2-methylphenol-	20.000	20.272	-1.4	102	0.00
25	Phenanthrene	20.000	19.547	2.3	96	0.00
26 S	Anthracene-d10	20.000	20.093	-0.5	98	0.00
27	Anthracene	20.000	19.859	0.7	97	0.00
28 C	Fluoranthene	20.000	20.188	-0.9	98	0.00
29 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
30 S	Pyrene-d10	20.000	19.846	0.8	99	0.00
31	Pyrene	20.000	19.640	1.8	97	0.00
32	Benzo(a)anthracene	20.000	19.724	1.4	96	0.00
33	Chrysene	20.000	19.780	1.1	97	0.00
34 I	Perylene-d12	20.000	20.000	0.0	100	0.00
35	Benzo(b)fluoranthene	20.000	19.774	1.1	98	0.00
36	Benzo(k)fluoranthene	20.000	19.566	2.2	96	0.00
37 S	Benzo(a)pyrene-d12	20.000	19.914	0.4	98	0.00
38 C	Benzo(a)pyrene	20.000	19.810	1.0	97	0.00
39	Indeno(1,2,3-cd)pyrene	20.000	19.952	0.2	98	-0.01
40	Dibenzo(a,h)anthracene	20.000	20.151	-0.8	98	-0.01
41	Benzo(g,h,i)perylene	20.000	19.824	0.9	98	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM073016\
Data File : BM006754.D
Acq On : 30 Jul 2016 02:41
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02059

Quant Time: Jul 30 03:16:10 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM072016.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Jul 30 01:37:26 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)
----------	--------	-------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0