

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM073016\
 Data File : BM006754.D
 Acq On : 30 Jul 2016 02:41
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02059

Quant Time: Jul 30 03:16:10 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM072016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 30 01:37:26 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2 S	1,4-Dioxane-d8	0.495	0.520	-5.1	98	0.00
3 S	Phenol-d5	1.702	1.757	-3.2	96	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	1.054	1.108	-5.1	96	0.00
5 S	2-Chlorophenol-d4	1.360	1.385	-1.8	96	0.00
6 S	4-Methylphenol-d8	1.320	1.379	-4.5	96	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.151	0.151	0.0	99	0.00
9 S	2-Nitrophenol-d4	0.161	0.168	-4.3	102	0.00
10 S	2,4-Dichlorophenol-d3	0.304	0.308	-1.3	99	0.00
11	Naphthalene	1.031	1.007	2.3	97	0.00
12 S	4-Chloroaniline-d4	0.335	0.366	-9.3	87	0.00
13	2-Methylnaphthalene	0.757	0.755	0.3	97	0.00
14	1-Methylnaphthalene	0.722	0.723	-0.1	98	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
16 S	Dimethylphthalate-d6	1.568	1.591	-1.5	100	0.00
17 S	Acenaphthylene-d8	2.002	2.043	-2.0	101	0.00
18	Acenaphthylene	2.144	2.136	0.4	98	0.00
19 C	Acenaphthene	1.378	1.367	0.8	98	0.00
20 S	4-Nitrophenol-d4	0.272	0.260	4.4	96	0.00
21 S	Fluorene-d10	1.402	1.415	-0.9	100	0.00
22	Fluorene	1.552	1.590	-2.4	99	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
24 S	4,6-Dinitro-2-methylphenol-	0.104	0.105	-1.0	102	0.00
25	Phenanthrene	1.125	1.099	2.3	96	0.00
26 S	Anthracene-d10	0.950	0.954	-0.4	98	0.00
27	Anthracene	1.158	1.150	0.7	97	0.00
28 C	Fluoranthene	1.321	1.334	-1.0	98	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
30 S	Pyrene-d10	0.911	0.904	0.8	99	0.00
31	Pyrene	1.211	1.189	1.8	97	0.00
32	Benzo(a)anthracene	1.212	1.195	1.4	96	0.00
33	Chrysene	1.118	1.105	1.2	97	0.00
34 I	Perylene-d12	1.000	1.000	0.0	100	0.00
35	Benzo(b)fluoranthene	1.199	1.185	1.2	98	0.00
36	Benzo(k)fluoranthene	1.159	1.134	2.2	96	0.00
37 S	Benzo(a)pyrene-d12	0.920	0.916	0.4	98	0.00
38 C	Benzo(a)pyrene	1.167	1.156	0.9	97	0.00
39	Indeno(1,2,3-cd)pyrene	1.351	1.348	0.2	98	-0.01
40	Dibenzo(a,h)anthracene	1.120	1.129	-0.8	98	-0.01
41	Benzo(g,h,i)perylene	1.158	1.148	0.9	98	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM073016\
Data File : BM006754.D
Acq On : 30 Jul 2016 02:41
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02059

Quant Time: Jul 30 03:16:10 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM072016.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Jul 30 01:37:26 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0