


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM073016\
 Data File : BM006735.D
 Acq On : 29 Jul 2016 12:14
 Operator : UM/SJ
 Sample : DFTPP62
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP62

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM072016.M
 Title : SVOA CALIBRATION
 Last Update : Mon Aug 01 18:04:00 2016



AutoFind: Scans 2527, 2528, 2529; Background Corrected with Scan 2522

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.8	53005	PASS
68	69	0.00	2	1.5	726	PASS
69	198	0.00	100	40.0	48381	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.5	57458	PASS
197	198	0.00	2	0.3	412	PASS
198	198	100	100	100.0	120882	PASS
199	198	5	9	6.8	8264	PASS
275	198	10	60	25.4	30749	PASS
365	198	1	100	3.1	3787	PASS
441	443	0.01	100	80.4	15173	PASS
442	198	50	100	79.3	95877	PASS
443	442	15	24	19.7	18866	PASS

m/z	Abundance
37.00	301.0
38.10	488.0
39.10	3980.0
40.00	1075.0
44.00	1415.0
49.00	631.0
50.00	13683.0
51.00	57376.0
52.00	2694.0
54.90	525.0
56.00	1547.0
57.00	3602.0
61.00	577.0
61.90	778.0
63.00	1935.0
65.10	1014.0
67.90	771.0
69.00	50480.0
73.00	675.0
74.00	5011.0
75.00	8132.0
76.00	2943.0
77.00	59528.0
78.00	3653.0
79.00	3403.0
79.90	2436.0
81.00	3722.0
82.00	1248.0
83.00	805.0
84.00	328.0
84.90	597.0
86.00	987.0
87.00	569.0
91.00	848.0
92.00	1052.0
93.00	6640.0
94.00	625.0
98.00	4817.0
99.00	4990.0
100.00	400.0
101.00	2462.0
102.90	571.0
103.90	1307.0
104.90	1184.0
106.10	455.0
107.00	18464.0
108.00	2695.0
109.10	422.0
110.00	29808.0
111.00	4203.0
111.90	452.0
116.00	904.0
117.00	14042.0
118.00	1017.0
121.90	963.0
123.00	1347.0
123.90	721.0
124.90	775.0
126.10	306.0
127.00	56672.0
128.00	3970.0
129.00	21440.0
130.00	2061.0
130.90	301.0
134.00	551.0
135.00	2374.0
135.90	933.0
137.00	1232.0
141.00	2718.0
141.90	864.0
143.00	744.0
146.10	596.0
147.00	1380.0
148.00	3360.0
149.00	883.0
150.90	324.0
151.90	402.0
153.00	772.0
153.90	699.0
155.00	1618.0
156.00	2175.0
157.10	574.0
157.90	487.0
160.00	803.0
161.00	1510.0
162.00	349.0
164.90	1086.0
166.00	872.0
167.00	5720.0
168.00	4477.0
168.90	640.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

171.90	632.0
173.00	901.0
174.00	1231.0
175.00	2301.0
175.90	463.0
177.00	1343.0
177.80	337.0
179.00	3922.0
180.00	3256.0
181.00	1578.0
181.90	347.0
184.00	308.0
185.00	1786.0
186.00	15987.0
187.00	4213.0
188.00	497.0
189.00	992.0
191.00	737.0
192.00	1445.0
193.00	1531.0
194.00	361.0
196.00	2853.0
196.90	559.0
198.00	111496.0
199.00	8221.0
199.90	468.0
201.50	334.0
203.00	791.0
204.00	4240.0
205.00	6799.0
206.00	28232.0
207.00	5512.0
208.00	944.0
208.90	413.0
210.10	432.0
211.10	1045.0
215.00	354.0
216.00	644.0
217.00	7575.0
218.00	918.0
221.00	3545.0
221.80	1043.0
222.90	1744.0
224.00	16167.0
225.00	4275.0
226.10	753.0
227.00	6532.0
228.00	1258.0
228.90	1256.0
231.00	459.0
234.00	454.0
237.00	1128.0
242.10	908.0
243.10	929.0
244.00	11350.0
245.00	1416.0
245.90	1998.0
247.00	602.0
249.00	459.0
252.90	301.0
254.00	367.0
255.00	62048.0
256.00	8891.0
257.00	807.0
258.00	3421.0
259.00	549.0
265.00	1330.0
273.00	2051.0
274.00	4302.0
275.00	26128.0
276.00	3745.0
277.00	2040.0
278.00	542.0
285.10	536.0
293.00	700.0
296.00	6808.0
297.00	964.0
303.00	750.0
315.00	788.0
316.00	396.0
323.00	2211.0
324.00	489.0
327.00	506.0
327.90	354.0
334.00	1298.0
334.90	406.0
346.10	575.0
352.00	665.0
353.00	487.0
353.90	548.0
365.00	2875.0
366.00	364.0
372.00	1497.0
372.90	343.0
401.90	420.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

402.90	562.0
421.10	348.0
422.00	476.0
423.00	3379.0
424.00	1429.0
441.10	11415.0
442.10	67464.0
443.00	13719.0
444.00	1279.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

m/z	Abundance
38.00	645.0
39.10	4472.0
40.00	1047.0
44.00	1393.0
49.00	793.0
50.00	15878.0
51.00	61528.0
52.00	3377.0
55.00	484.0
56.00	1912.0
57.00	4046.0
60.90	839.0
62.00	996.0
63.00	2471.0
65.00	1167.0
68.00	884.0
69.00	56288.0
73.00	661.0
74.00	5091.0
75.00	8424.0
76.00	2630.0
77.00	66256.0
78.00	4763.0
79.00	4299.0
80.00	3147.0
81.00	4950.0
81.90	1277.0
82.90	1166.0
85.00	767.0
86.00	1396.0
86.90	780.0
91.00	1163.0
92.00	1250.0
93.00	7663.0
94.00	457.0
96.00	537.0
98.00	5160.0
99.00	5061.0
99.90	438.0
100.90	2941.0
102.90	664.0
103.90	1432.0
104.90	1924.0
105.90	330.0
107.00	20808.0
108.00	3169.0
109.00	546.0
110.00	35048.0
111.00	6050.0
112.00	653.0
115.90	1093.0
117.00	17248.0
117.90	1239.0
122.00	1362.0
122.90	1977.0
123.90	830.0
125.00	894.0
127.00	66976.0
128.00	5116.0
129.00	24664.0
130.00	2093.0
131.10	392.0
133.90	719.0
135.00	2262.0
136.00	565.0
137.00	1410.0
137.80	323.0
141.00	3504.0
142.00	1079.0
143.00	734.0
145.90	608.0
146.90	1684.0
148.00	3850.0
148.90	765.0
149.90	371.0
151.10	570.0
151.90	363.0
152.90	1089.0
154.00	1067.0
155.00	2038.0
156.00	2764.0
157.00	508.0
158.00	559.0
159.00	495.0
160.00	1066.0
161.00	1544.0
161.80	473.0
164.90	1318.0
165.90	1077.0
167.00	6185.0
168.00	4932.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

168.90	665.0
169.90	339.0
171.00	303.0
171.90	547.0
173.00	1066.0
174.10	1418.0
175.00	3077.0
176.10	511.0
176.90	948.0
177.90	591.0
179.00	5382.0
179.90	3915.0
180.90	1894.0
183.90	490.0
185.00	2325.0
186.00	20792.0
187.00	5577.0
187.90	438.0
188.90	1173.0
190.90	710.0
192.00	1690.0
193.00	2216.0
193.90	466.0
195.10	379.0
196.00	3252.0
197.10	607.0
198.00	141952.0
198.90	9286.0
200.00	636.0
201.60	547.0
202.90	1234.0
204.00	5455.0
205.00	8327.0
206.00	38328.0
207.00	6131.0
208.00	1451.0
209.00	522.0
209.90	646.0
210.90	1566.0
214.80	385.0
216.00	1028.0
217.00	9684.0
218.00	1351.0
221.00	5251.0
221.80	1004.0
223.00	2235.0
224.00	21264.0
225.00	5419.0
226.00	1001.0
227.00	8944.0
227.90	1195.0
229.00	2001.0
231.00	972.0
233.90	411.0
234.90	526.0
235.90	412.0
237.00	796.0
239.00	398.0
241.00	563.0
242.00	1061.0
243.00	1159.0
244.00	16289.0
245.10	2028.0
246.00	3364.0
246.90	780.0
249.00	650.0
253.00	317.0
254.00	656.0
255.00	86760.0
256.00	12575.0
257.00	1078.0
258.00	4953.0
259.00	747.0
265.00	2134.0
273.00	2720.0
274.00	6389.0
275.00	36024.0
276.00	4724.0
277.00	3207.0
278.00	810.0
282.90	441.0
284.10	363.0
285.00	488.0
293.10	491.0
296.00	10205.0
297.00	1709.0
303.00	1069.0
304.00	384.0
314.10	399.0
315.00	644.0
316.00	772.0
321.10	383.0
323.10	4008.0
324.00	410.0
326.90	711.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

328.10	319.0
333.00	300.0
334.00	1861.0
335.00	471.0
346.00	630.0
352.00	758.0
353.00	750.0
354.00	1230.0
365.00	4151.0
366.00	681.0
372.00	1687.0
372.90	509.0
383.00	338.0
401.90	521.0
403.00	901.0
404.00	323.0
421.00	847.0
422.00	1166.0
423.00	5606.0
424.10	1668.0
441.00	16488.0
442.10	111104.0
443.10	21648.0
444.10	2018.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

m/z	Abundance
37.90	463.0
39.10	2578.0
40.00	874.0
44.00	1743.0
48.90	613.0
50.00	10112.0
51.00	40112.0
52.10	1779.0
55.00	300.0
55.90	1251.0
57.00	2609.0
60.90	771.0
62.10	520.0
63.00	1338.0
65.10	524.0
68.00	523.0
69.00	38376.0
73.00	483.0
74.00	3630.0
75.00	7169.0
76.00	1889.0
77.00	45624.0
78.00	2906.0
79.00	2669.0
80.00	2242.0
81.00	3510.0
82.10	814.0
82.90	566.0
85.00	644.0
85.90	931.0
86.90	592.0
91.00	753.0
92.00	889.0
93.00	5702.0
94.10	406.0
96.10	324.0
98.00	4435.0
99.00	3530.0
101.00	1789.0
103.00	541.0
103.90	1133.0
104.90	1047.0
107.00	15515.0
108.00	2431.0
109.10	511.0
110.00	24432.0
111.00	3778.0
111.90	709.0
115.90	819.0
117.00	12399.0
117.90	640.0
122.00	844.0
122.90	1041.0
124.00	548.0
125.00	559.0
127.00	48728.0
128.00	3113.0
129.00	18104.0
129.90	1729.0
130.90	304.0
134.00	472.0
134.90	1336.0
136.00	716.0
137.00	989.0
140.90	2503.0
142.10	646.0
143.00	817.0
146.00	592.0
147.00	1377.0
148.00	2645.0
148.90	650.0
151.10	357.0
152.90	996.0
153.90	666.0
155.00	1434.0
156.00	2249.0
157.10	301.0
158.00	501.0
159.00	386.0
160.00	789.0
160.90	935.0
162.00	468.0
164.90	1093.0
166.00	693.0
167.00	5748.0
168.00	3113.0
169.00	444.0
172.90	620.0
173.90	781.0
175.00	2165.0
176.00	635.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

177.00	847.0
179.00	4037.0
180.00	3388.0
181.00	1487.0
185.00	2300.0
186.00	14588.0
187.00	4588.0
188.00	456.0
189.00	787.0
190.90	566.0
191.90	1108.0
193.00	1503.0
193.90	343.0
196.00	3059.0
197.00	629.0
198.00	109200.0
199.00	7287.0
200.00	728.0
201.50	422.0
203.00	871.0
204.00	4566.0
205.00	6954.0
206.00	29000.0
207.00	4302.0
208.00	1067.0
208.90	464.0
211.00	1337.0
216.00	616.0
217.00	8186.0
217.90	973.0
221.00	3046.0
221.90	1072.0
223.00	1619.0
224.00	17480.0
225.00	4018.0
226.10	391.0
227.00	6788.0
228.00	983.0
228.90	1270.0
231.10	467.0
233.90	491.0
234.90	576.0
235.90	319.0
237.00	931.0
240.00	328.0
241.00	426.0
242.00	905.0
243.00	1245.0
244.00	13237.0
245.00	1777.0
246.00	2925.0
247.00	409.0
248.90	607.0
254.10	639.0
255.00	69448.0
256.00	9848.0
257.00	590.0
258.00	3894.0
259.00	650.0
265.00	1833.0
273.00	2644.0
274.00	5374.0
275.00	30096.0
276.00	3977.0
277.00	2351.0
277.90	409.0
283.00	411.0
285.00	460.0
293.00	538.0
296.00	8867.0
297.00	1170.0
303.00	1272.0
304.00	407.0
314.00	378.0
315.00	1287.0
316.10	367.0
320.90	419.0
323.00	3289.0
324.00	650.0
326.90	671.0
328.00	425.0
334.00	2106.0
334.90	504.0
340.90	305.0
345.90	684.0
352.00	954.0
352.90	521.0
354.10	950.0
365.00	4336.0
366.10	597.0
371.10	395.0
372.10	1379.0
373.00	356.0
383.00	489.0
402.00	774.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

403.00	1078.0
421.00	784.0
421.90	1107.0
423.00	5577.0
424.00	1573.0
441.10	17616.0
442.10	109064.0
443.10	21232.0
444.10	1388.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62

Scan 2522 (17.421 min): BM006735.D
DFTPP62

m/z	Abundance
39.90	857.0
44.00	1590.0
48.90	405.0
83.90	355.0
207.00	959.0

Instrument :
BNA_M
ClientSampleId :
DFTPP62