

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM073116\
 Data File : BM006797.D
 Acq On : 31 Jul 2016 20:26
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02063

Quant Time: Aug 01 06:44:18 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM072016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 30 01:37:26 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2 S	1,4-Dioxane-d8	0.495	0.478	3.4	106	0.00
3 S	Phenol-d5	1.702	1.761	-3.5	114	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	1.054	1.107	-5.0	113	0.00
5 S	2-Chlorophenol-d4	1.360	1.380	-1.5	113	0.00
6 S	4-Methylphenol-d8	1.320	1.383	-4.8	114	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
8 S	Nitrobenzene-d5	0.151	0.154	-2.0	118	0.00
9 S	2-Nitrophenol-d4	0.161	0.170	-5.6	121	0.00
10 S	2,4-Dichlorophenol-d3	0.304	0.312	-2.6	116	0.00
11	Naphthalene	1.031	1.017	1.4	114	0.00
12 S	4-Chloroaniline-d4	0.335	0.391	-16.7	108	0.00
13	2-Methylnaphthalene	0.757	0.758	-0.1	114	0.00
14	1-Methylnaphthalene	0.722	0.720	0.3	114	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
16 S	Dimethylphthalate-d6	1.568	1.593	-1.6	118	0.00
17 S	Acenaphthylene-d8	2.002	2.036	-1.7	118	0.00
18	Acenaphthylene	2.144	2.110	1.6	114	0.00
19 C	Acenaphthene	1.378	1.338	2.9	112	0.00
20 S	4-Nitrophenol-d4	0.272	0.271	0.4	117	0.00
21 S	Fluorene-d10	1.402	1.418	-1.1	118	0.00
22	Fluorene	1.552	1.585	-2.1	115	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
24 S	4,6-Dinitro-2-methylphenol-	0.104	0.108	-3.8	122	0.00
25	Phenanthrene	1.125	1.111	1.2	115	0.00
26 S	Anthracene-d10	0.950	0.953	-0.3	115	0.00
27	Anthracene	1.158	1.151	0.6	115	0.00
28 C	Fluoranthene	1.321	1.351	-2.3	117	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
30 S	Pyrene-d10	0.911	0.936	-2.7	117	0.00
31	Pyrene	1.211	1.232	-1.7	115	0.00
32	Benzo(a)anthracene	1.212	1.219	-0.6	112	0.00
33	Chrysene	1.118	1.120	-0.2	112	0.00
34 I	Perylene-d12	1.000	1.000	0.0	112	0.00
35	Benzo(b)fluoranthene	1.199	1.224	-2.1	113	0.00
36	Benzo(k)fluoranthene	1.159	1.136	2.0	107	0.00
37 S	Benzo(a)pyrene-d12	0.920	0.938	-2.0	113	-0.01
38 C	Benzo(a)pyrene	1.167	1.173	-0.5	111	0.00
39	Indeno(1,2,3-cd)pyrene	1.351	1.362	-0.8	111	0.00
40	Dibenzo(a,h)anthracene	1.120	1.136	-1.4	110	-0.01
41	Benzo(g,h,i)perylene	1.158	1.158	0.0	111	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM073116\
Data File : BM006797.D
Acq On : 31 Jul 2016 20:26
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02063

Quant Time: Aug 01 06:44:18 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM072016.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Jul 30 01:37:26 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0