

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM073116\
 Data File : BM006825.D
 Acq On : 01 Aug 2016 14:05
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02065

Quant Time: Aug 01 16:00:44 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM072016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jul 30 01:37:26 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	207#	0.00
2 S	1,4-Dioxane-d8	0.495	0.451	8.9	187#	0.00
3 S	Phenol-d5	1.702	1.767	-3.8	213#	0.01
4 S	Bis-(2-Chloroethyl)ether-d8	1.054	1.075	-2.0	205#	0.00
5 S	2-Chlorophenol-d4	1.360	1.380	-1.5	211#	0.00
6 S	4-Methylphenol-d8	1.320	1.351	-2.3	208#	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	207#	0.00
8 S	Nitrobenzene-d5	0.151	0.161	-6.6	225#	0.00
9 S	2-Nitrophenol-d4	0.161	0.174	-8.1	225#	0.00
10 S	2,4-Dichlorophenol-d3	0.304	0.316	-3.9	214#	0.00
11	Naphthalene	1.031	1.011	1.9	206#	0.00
12 S	4-Chloroaniline-d4	0.335	0.395	-17.9	199#	0.00
13	2-Methylnaphthalene	0.757	0.731	3.4	200#	0.00
14	1-Methylnaphthalene	0.722	0.687	4.8	197#	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	195#	0.00
16 S	Dimethylphthalate-d6	1.568	1.569	-0.1	195#	0.01
17 S	Acenaphthylene-d8	2.002	2.063	-3.0	200#	0.00
18	Acenaphthylene	2.144	2.149	-0.2	194#	0.00
19 C	Acenaphthene	1.378	1.355	1.7	190#	0.00
20 S	4-Nitrophenol-d4	0.272	0.266	2.2	193#	0.02
21 S	Fluorene-d10	1.402	1.383	1.4	192#	0.00
22	Fluorene	1.552	1.528	1.5	186#	0.01
23 I	Phenanthrene-d10	1.000	1.000	0.0	183#	0.01
24 S	4,6-Dinitro-2-methylphenol-	0.104	0.064	38.5#	113	0.02
25	Phenanthrene	1.125	1.117	0.7	180#	0.01
26 S	Anthracene-d10	0.950	0.960	-1.1	181#	0.01
27	Anthracene	1.158	1.157	0.1	180#	0.00
28 C	Fluoranthene	1.321	1.278	3.3	172#	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	150	0.01
30 S	Pyrene-d10	0.911	1.026	-12.6	170#	0.00
31	Pyrene	1.211	1.343	-10.9	165#	0.01
32	Benzo(a)anthracene	1.212	1.215	-0.2	147	0.01
33	Chrysene	1.118	1.139	-1.9	151#	0.01
34 I	Perylene-d12	1.000	1.000	0.0	155#	0.02
35	Benzo(b)fluoranthene	1.199	1.156	3.6	148	0.02
36	Benzo(k)fluoranthene	1.159	1.136	2.0	148	0.02
37 S	Benzo(a)pyrene-d12	0.920	0.916	0.4	152#	0.02
38 C	Benzo(a)pyrene	1.167	1.148	1.6	150	0.02
39	Indeno(1,2,3-cd)pyrene	1.351	1.350	0.1	152#	0.04
40	Dibenzo(a,h)anthracene	1.120	1.123	-0.3	151#	0.03
41	Benzo(g,h,i)perylene	1.158	1.174	-1.4	155#	0.04

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM073116\
Data File : BM006825.D
Acq On : 01 Aug 2016 14:05
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02065

Quant Time: Aug 01 16:00:44 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM072016.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Jul 30 01:37:26 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0