


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM073120\
Data File : BM026952.D
Acq On : 31 Jul 2020 11:13
Operator : JU/CG
Sample : DFTPP17
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP17

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM072720MA.M
Title : SVOA CALIBRATION
Last Update : Fri Jul 31 17:10:07 2020



AutoFind: Scans 2510, 2511, 2512; Background Corrected with Scan 2505

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.1	93840	PASS
68	69	0.00	2	1.9	2853	PASS
69	198	0.00	100	61.9	148355	PASS
70	69	0.00	2	0.6	886	PASS
127	198	10	80	52.2	125237	PASS
197	198	0.00	2	0.7	1634	PASS
198	198	100	100	100.0	239808	PASS
199	198	5	9	7.0	16835	PASS
275	198	10	60	28.4	68205	PASS
365	198	1	100	4.8	11418	PASS
441	198	0.01	100	11.7	28004	PASS
442	198	50	100	74.9	179504	PASS
443	442	15	24	19.6	35104	PASS

Scan 2505 (17.471 min): BM026952.D
DFTPP17

m/z	Abundance
39.00	208.0
40.00	527.0
41.00	344.0
43.00	272.0
44.00	1291.0
54.90	306.0
68.90	234.0
72.90	350.0
207.10	495.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

m/z	Abundance
36.00	176.0
37.00	457.0
38.10	1458.0
39.10	6449.0
40.00	462.0
41.00	344.0
42.00	153.0
42.90	288.0
44.00	1224.0
49.00	610.0
50.00	20672.0
51.00	73720.0
52.00	4374.0
53.00	352.0
55.00	691.0
56.00	2831.0
57.00	5603.0
58.00	259.0
61.00	1451.0
62.10	1359.0
63.10	4028.0
64.00	599.0
65.10	1976.0
66.00	200.0
66.90	302.0
68.10	2470.0
69.00	115528.0
70.00	756.0
73.00	1093.0
74.00	12302.0
75.00	18192.0
76.00	5855.0
77.00	111048.0
78.10	8235.0
79.00	8655.0
80.00	6396.0
81.00	9112.0
82.00	2252.0
82.90	1467.0
84.00	382.0
85.00	1714.0
86.00	2585.0
87.00	1233.0
87.90	382.0
89.00	337.0
91.00	1975.0
92.00	2208.0
93.00	14933.0
94.00	1086.0
96.00	538.0
97.00	315.0
98.00	11466.0
99.00	8491.0
100.10	514.0
101.00	4254.0
101.90	241.0
102.90	1193.0
103.90	2779.0
105.00	2609.0
106.10	981.0
107.00	31392.0
108.00	4306.0
109.10	952.0
110.00	50832.0
111.00	7867.0
112.00	1083.0
113.00	536.0
116.00	1901.0
117.00	31008.0
118.00	2484.0
118.90	166.0
120.00	448.0
122.10	1788.0
123.00	2907.0
124.00	1471.0
125.00	1296.0
126.20	540.0
127.00	88800.0
128.00	7412.0
129.00	38472.0
130.00	3374.0
130.90	645.0
131.90	192.0
132.80	204.0
134.00	1082.0
135.00	3043.0
136.00	1073.0
137.00	1947.0
138.10	500.0
140.00	569.0
141.00	5499.0
142.10	2143.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

142.90	1502.0
144.00	378.0
144.90	333.0
146.00	1136.0
147.00	3150.0
148.00	7034.0
149.00	1630.0
149.90	360.0
151.10	823.0
152.00	152.0
153.00	2021.0
154.00	1083.0
155.00	2994.0
156.00	4486.0
157.10	839.0
158.00	1125.0
159.00	916.0
160.00	1436.0
161.00	3135.0
162.00	910.0
162.90	179.0
164.10	291.0
164.90	2113.0
166.00	1661.0
167.00	11774.0
168.00	6227.0
169.00	1053.0
169.90	496.0
170.90	475.0
172.00	1025.0
173.00	1254.0
174.00	2045.0
175.00	3842.0
176.00	1078.0
177.00	1913.0
178.00	885.0
179.00	8846.0
180.00	5435.0
181.00	2607.0
182.20	380.0
183.00	279.0
183.90	702.0
185.00	4081.0
186.00	26432.0
187.00	7905.0
188.00	805.0
189.00	1752.0
189.90	420.0
191.00	833.0
192.00	2216.0
193.00	2824.0
193.90	628.0
195.00	300.0
196.00	4185.0
197.00	1123.0
197.90	156928.0
198.90	12059.0
200.00	995.0
201.50	829.0
203.00	1657.0
204.00	6658.0
205.00	12085.0
206.00	47400.0
207.00	6061.0
208.00	1854.0
209.00	743.0
210.20	1074.0
211.10	1679.0
211.90	365.0
215.00	723.0
216.10	1094.0
217.00	14135.0
218.00	1706.0
219.00	214.0
221.00	5051.0
221.90	1583.0
223.00	2755.0
224.00	24192.0
225.10	6632.0
225.90	736.0
227.00	13165.0
228.00	2099.0
229.00	2387.0
229.90	265.0
230.90	677.0
233.00	188.0
233.90	568.0
235.00	856.0
235.90	833.0
237.00	1307.0
237.90	172.0
239.00	378.0
240.10	355.0
240.90	519.0
242.00	1395.0
243.00	1454.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

244.00	17824.0
245.00	2778.0
246.00	4115.0
246.90	879.0
249.00	752.0
249.90	163.0
250.90	201.0
251.90	301.0
253.00	354.0
254.10	847.0
255.00	96336.0
256.00	14548.0
257.00	1442.0
258.00	6584.0
258.90	1123.0
260.00	167.0
260.90	195.0
265.00	2694.0
266.00	336.0
271.00	368.0
271.80	200.0
273.00	2622.0
274.00	6874.0
275.00	40480.0
276.00	5633.0
277.00	4087.0
278.00	964.0
281.00	235.0
282.00	151.0
283.00	447.0
283.90	265.0
285.00	665.0
289.90	240.0
292.00	154.0
292.90	711.0
293.80	225.0
294.80	171.0
296.00	13290.0
297.00	2474.0
301.00	158.0
301.90	332.0
303.00	1499.0
304.00	325.0
307.80	245.0
309.00	192.0
314.10	534.0
314.90	1632.0
315.90	638.0
320.90	346.0
322.00	242.0
323.00	3354.0
324.10	828.0
327.00	894.0
327.90	626.0
332.00	271.0
333.00	612.0
334.00	2547.0
335.00	674.0
340.90	246.0
345.90	1127.0
352.00	1152.0
353.00	724.0
354.10	1136.0
355.00	194.0
365.00	6432.0
366.00	935.0
371.00	197.0
372.10	1693.0
373.10	371.0
383.00	409.0
391.30	169.0
401.10	221.0
402.10	626.0
403.10	744.0
404.00	366.0
421.00	787.0
422.00	706.0
423.00	5031.0
424.00	1194.0
425.20	154.0
441.00	13677.0
442.10	84624.0
443.10	17280.0
444.00	1625.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

m/z	Abundance
37.00	701.0
38.10	1702.0
39.10	9089.0
40.00	831.0
40.90	458.0
43.10	569.0
44.00	1537.0
44.90	523.0
49.10	1120.0
50.10	31696.0
51.10	112376.0
52.00	5804.0
53.00	415.0
55.00	905.0
56.00	4254.0
57.00	7994.0
57.90	325.0
60.00	312.0
61.00	1973.0
62.00	2639.0
63.10	5935.0
64.10	939.0
65.00	2985.0
66.00	150.0
67.10	279.0
68.10	3040.0
69.00	179712.0
70.00	1069.0
71.10	241.0
73.10	1748.0
74.10	18648.0
75.00	28456.0
76.10	8816.0
77.10	178432.0
78.00	12091.0
79.00	13870.0
80.00	10255.0
81.00	14637.0
82.00	3724.0
83.00	2678.0
84.00	373.0
85.00	2239.0
86.00	3975.0
87.00	1589.0
87.90	656.0
88.90	316.0
91.00	3150.0
92.00	3648.0
93.00	25056.0
94.00	1842.0
95.00	417.0
95.90	707.0
96.70	177.0
98.00	17272.0
99.00	12050.0
100.00	1066.0
101.00	5973.0
102.00	386.0
103.00	2076.0
104.00	4707.0
105.00	3937.0
106.00	1202.0
107.00	51296.0
108.00	8450.0
109.10	1225.0
110.00	80344.0
111.00	14760.0
111.90	1744.0
113.00	662.0
115.10	295.0
116.00	2863.0
117.00	51592.0
118.00	3813.0
118.90	360.0
119.90	589.0
121.10	196.0
122.00	3322.0
123.00	6029.0
123.90	2545.0
125.00	1931.0
126.20	1010.0
127.00	151104.0
128.00	11687.0
129.00	63360.0
130.00	4725.0
131.00	1099.0
132.00	447.0
133.00	350.0
134.00	1616.0
135.00	5087.0
136.00	2109.0
137.00	3153.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

138.00	741.0
138.90	204.0
140.10	1037.0
141.00	9537.0
142.00	2580.0
142.90	1864.0
144.10	474.0
145.00	532.0
145.90	1510.0
147.00	4829.0
148.00	11702.0
149.00	2367.0
149.90	541.0
150.90	1135.0
151.90	483.0
152.90	2455.0
154.00	1966.0
155.00	5021.0
156.00	7393.0
157.00	1247.0
157.80	1631.0
159.00	1191.0
160.00	2930.0
161.00	4605.0
162.00	1223.0
163.00	445.0
164.00	304.0
165.00	3642.0
166.00	2459.0
167.00	19992.0
168.00	10112.0
169.00	1741.0
169.90	618.0
170.90	1050.0
171.90	1844.0
173.00	2185.0
174.00	4065.0
175.00	6959.0
176.00	1599.0
177.00	3262.0
178.00	1345.0
179.00	14676.0
180.00	9311.0
181.00	4178.0
182.10	692.0
183.10	283.0
184.00	1173.0
185.00	5833.0
186.00	47248.0
187.00	14355.0
188.10	1360.0
189.00	3049.0
190.00	549.0
191.00	1474.0
192.00	4366.0
193.00	4626.0
194.00	1017.0
195.00	617.0
196.00	8363.0
197.10	1678.0
198.00	281664.0
199.00	19024.0
200.00	1710.0
201.50	1215.0
202.10	362.0
202.90	2596.0
204.00	13004.0
205.00	21176.0
206.00	81168.0
207.10	11621.0
208.00	3102.0
209.00	819.0
209.90	1537.0
211.00	3411.0
214.90	977.0
216.00	2230.0
217.00	24712.0
218.00	2836.0
219.00	436.0
219.80	404.0
221.00	10682.0
221.90	3154.0
223.00	5310.0
224.00	47528.0
225.00	12088.0
226.00	1208.0
227.00	23896.0
228.00	3444.0
229.00	4264.0
229.90	544.0
231.00	1629.0
232.00	175.0
233.10	417.0
234.00	1445.0
235.00	1499.0
236.00	1066.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

237.00	1814.0
237.90	329.0
239.00	1032.0
240.00	753.0
241.00	1085.0
242.00	2654.0
243.00	2643.0
244.00	33992.0
245.00	4102.0
246.00	8451.0
247.00	1853.0
248.00	404.0
249.00	1407.0
249.90	240.0
251.00	411.0
251.90	384.0
252.90	985.0
254.10	1311.0
255.00	185216.0
256.00	27232.0
257.00	2374.0
258.00	12353.0
259.00	1745.0
260.00	446.0
260.90	206.0
264.00	363.0
265.00	5703.0
265.90	602.0
270.00	275.0
271.10	368.0
272.00	575.0
273.00	5916.0
274.00	13836.0
275.00	80376.0
276.00	10466.0
277.00	8379.0
277.90	1513.0
278.90	329.0
281.00	264.0
281.90	187.0
283.00	985.0
283.90	661.0
284.90	1303.0
286.00	251.0
288.90	386.0
289.90	466.0
290.90	346.0
292.00	275.0
293.00	1622.0
293.80	326.0
295.00	434.0
296.00	27464.0
296.90	4023.0
298.10	291.0
300.90	272.0
302.10	406.0
303.00	2912.0
304.00	445.0
305.10	180.0
307.90	445.0
310.10	192.0
312.90	266.0
313.90	1110.0
315.00	3173.0
315.90	1424.0
317.00	215.0
321.00	825.0
321.90	357.0
323.00	8055.0
324.10	1361.0
324.90	156.0
326.90	1869.0
328.00	719.0
328.90	221.0
332.00	802.0
332.90	764.0
334.00	4417.0
335.00	1302.0
341.00	667.0
341.90	209.0
346.00	1966.0
347.00	363.0
351.00	245.0
352.00	2327.0
353.10	1382.0
354.00	2298.0
354.90	443.0
365.00	13376.0
366.00	1928.0
369.90	293.0
371.00	562.0
372.00	3469.0
373.00	877.0
383.00	1094.0
384.10	215.0
390.00	497.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

391.00	250.0
392.10	173.0
401.10	159.0
402.00	1388.0
403.10	2695.0
404.10	689.0
405.10	280.0
421.00	2019.0
422.10	1907.0
423.00	11786.0
424.00	2763.0
425.00	273.0
441.10	30928.0
442.10	197824.0
443.10	38880.0
444.00	3460.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

m/z	Abundance
37.00	492.0
38.00	1627.0
39.10	7967.0
40.00	765.0
41.00	408.0
42.10	208.0
43.00	314.0
44.00	1177.0
45.00	508.0
49.00	1170.0
50.10	26880.0
51.10	95424.0
52.00	4700.0
53.00	357.0
54.90	913.0
56.00	3362.0
57.00	7096.0
57.90	451.0
59.10	152.0
61.00	1630.0
62.00	1946.0
63.00	5642.0
63.90	836.0
65.00	2520.0
66.00	178.0
67.10	269.0
68.10	3051.0
69.00	150528.0
70.00	834.0
71.10	245.0
73.00	1749.0
74.00	16784.0
75.00	24160.0
76.00	7948.0
77.10	153856.0
78.00	9714.0
79.00	11357.0
80.00	9291.0
81.00	12314.0
82.00	2815.0
83.00	2147.0
83.90	390.0
85.00	2158.0
86.00	3728.0
87.00	1466.0
87.90	316.0
88.80	301.0
91.00	2895.0
92.10	2932.0
93.00	19448.0
94.00	1259.0
95.00	260.0
95.90	691.0
98.00	16536.0
99.00	11119.0
99.90	1184.0
101.00	6110.0
101.90	332.0
102.90	1807.0
104.00	4324.0
105.00	3537.0
106.00	1426.0
107.00	45512.0
108.00	6471.0
109.10	1141.0
110.00	72432.0
111.00	11781.0
112.00	1644.0
113.00	404.0
115.00	238.0
116.00	2532.0
117.00	46312.0
118.00	2963.0
118.90	315.0
119.90	482.0
120.90	233.0
122.10	3269.0
123.00	4641.0
124.00	2179.0
125.00	1870.0
126.10	1032.0
127.00	135808.0
128.00	10026.0
129.00	60144.0
130.00	4982.0
131.00	812.0
131.90	419.0
132.80	303.0
133.90	1969.0
135.00	5074.0
135.90	1771.0
137.00	2570.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

138.00	777.0
138.80	363.0
140.00	819.0
141.00	8945.0
142.00	2516.0
143.00	1821.0
144.00	768.0
145.00	231.0
145.90	1459.0
147.00	4676.0
148.00	11965.0
149.00	2387.0
150.00	494.0
151.00	1196.0
151.90	574.0
153.00	2887.0
154.00	2234.0
155.00	5072.0
156.00	6245.0
157.00	1123.0
158.00	1677.0
159.00	1277.0
160.00	2460.0
161.00	4204.0
161.90	1155.0
163.10	317.0
164.00	376.0
165.00	3841.0
166.00	2425.0
167.00	18080.0
168.00	8919.0
168.90	1204.0
169.90	476.0
171.00	798.0
172.10	1575.0
173.00	2249.0
174.00	3318.0
175.00	6609.0
175.90	1392.0
177.00	3587.0
178.00	1194.0
179.00	13727.0
180.00	8986.0
180.90	4056.0
181.90	807.0
182.90	411.0
184.00	1122.0
185.00	6856.0
186.00	47576.0
187.00	13036.0
188.00	1266.0
189.00	3285.0
189.90	657.0
191.00	1749.0
192.00	4203.0
193.00	4688.0
194.00	840.0
194.90	461.0
196.00	8039.0
197.10	2102.0
197.90	280832.0
198.90	19424.0
199.90	2241.0
201.70	912.0
203.00	2910.0
204.00	12983.0
205.00	21968.0
206.00	84720.0
207.00	11827.0
208.00	3472.0
209.00	955.0
210.00	1476.0
211.00	3211.0
212.00	474.0
214.90	1242.0
216.00	1838.0
217.00	24368.0
218.00	3199.0
219.00	241.0
221.00	9938.0
221.80	2904.0
223.00	5340.0
224.00	49080.0
225.00	12417.0
226.10	1147.0
227.00	24632.0
228.00	3916.0
229.00	4511.0
229.90	575.0
231.00	1599.0
232.00	341.0
233.00	384.0
234.00	1800.0
235.00	1269.0
236.00	930.0
237.00	1698.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

238.00	334.0
239.00	1393.0
240.00	729.0
241.00	1411.0
242.00	2593.0
243.00	3190.0
244.00	36176.0
245.00	4369.0
246.00	9392.0
246.90	1671.0
247.90	354.0
248.90	1546.0
250.00	318.0
250.90	441.0
252.10	285.0
253.00	1011.0
254.00	1322.0
255.00	198336.0
256.00	28824.0
257.10	2095.0
258.00	14316.0
259.00	2296.0
260.00	350.0
260.90	276.0
263.90	344.0
265.00	5455.0
266.00	587.0
271.00	537.0
271.80	420.0
272.10	395.0
273.00	5606.0
274.00	15737.0
275.00	83760.0
276.00	10806.0
277.00	8579.0
277.90	1708.0
278.90	338.0
281.00	166.0
282.00	296.0
283.00	1010.0
284.00	569.0
285.00	1109.0
286.00	271.0
289.10	313.0
290.00	209.0
290.90	232.0
292.10	503.0
292.90	1776.0
293.90	557.0
295.10	809.0
296.00	30672.0
297.00	3976.0
298.00	288.0
300.80	461.0
302.00	643.0
303.00	3402.0
304.00	791.0
308.00	426.0
309.10	257.0
309.90	472.0
312.90	238.0
314.00	1387.0
315.00	3874.0
316.00	1786.0
317.00	386.0
321.00	893.0
322.10	492.0
323.00	8127.0
324.10	1796.0
324.90	301.0
326.90	1946.0
327.90	817.0
328.90	168.0
331.90	645.0
332.90	876.0
334.00	5034.0
335.00	1425.0
336.00	251.0
340.90	1257.0
342.00	295.0
346.00	2153.0
347.00	554.0
352.00	2417.0
353.10	1904.0
354.00	2593.0
355.10	588.0
358.90	207.0
364.20	221.0
365.00	14447.0
366.00	2147.0
366.90	167.0
370.10	241.0
371.00	559.0
372.00	4602.0
373.10	1015.0
374.10	213.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

383.00	1309.0
384.00	317.0
390.00	396.0
391.10	346.0
391.90	175.0
400.90	276.0
402.00	1697.0
403.00	2715.0
404.00	970.0
421.00	2288.0
422.00	2544.0
423.10	15159.0
424.10	3422.0
425.00	360.0
440.10	327.0
441.10	39408.0
442.10	256064.0
443.10	49152.0
444.10	4977.0
445.00	202.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17