

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM080118\
 Data File : BM016140.D
 Acq On : 01 Aug 2018 14:29
 Operator : SJ/JU
 Sample : J4175-04
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AC3

Manual Integrations
 APPROVED

Sohil
 8/2/2018 2:19:10 PM

Quant Time: Aug 02 08:24:22 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM071318MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 27 23:04:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.20	152	47478	20.00	ng/ul	0.00
18) Naphthalene-d8	11.02	136	215133	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.82	164	124340	20.00	ng/ul	-0.01
61) Phenanthrene-d10	17.56	188	265413m	20.00	ng/ul	0.00
77) Chrysene-d12	21.67	240	274183	20.00	ng/ul	0.00
85) Perylene-d12	24.03	264	252344	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.46	96	7864	7.29	ng/uL	0.00
5) Phenol-d5	7.36	99	29207	6.73	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.52	67	87410	32.02	ng/ul	-0.01
9) 2-Chlorophenol-d4	7.72	132	97279	27.61	ng/ul	0.00
13) 4-Methylphenol-d8	8.91	113	63983	17.22	ng/ul	-0.01
19) Nitrobenzene-d5	9.38	128	56100	36.32	ng/ul	-0.01
22) 2-Nitrophenol-d4	10.10	143	68368	38.93	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.64	165	109578	32.22	ng/ul	0.00
29) 4-Chloroaniline-d4	0.00	131	0	0.00	ng/ul	
43) Dimethylphthalate-d6	14.23	166	418108	36.89	ng/ul	0.00
46) Acenaphthylene-d8	14.52	160	484709	37.50	ng/ul	0.00
51) 4-Nitrophenol-d4	15.06	143	8606m	4.38	ng/ul	0.01
57) Fluorene-d10	15.81	176	339855	35.35	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.95	200	55160	31.14	ng/ul	0.00
70) Anthracene-d10	17.65	188	519188	38.21	ng/ul	0.00
78) Pyrene-d10	19.92	212	563826	44.37	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.89	264	505260	36.41	ng/ul	0.00

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM080118\
Data File : BM016140.D
Acq On : 01 Aug 2018 14:29
Operator : SJ/JU
Sample : J4175-04
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AC3

Quant Time: Aug 02 08:24:22 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM071318MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jul 27 23:04:38 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
8/2/2018 2:19:10 PM

