

Data Path : Z:\HPCHEM1\BNA M\DATA\BM080217\  
 Data File : BM011099.D  
 Acq On : 02 Aug 2017 18:02  
 Operator : SJ/JU  
 Sample : I4540-03  
 Misc :  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 72-12003

Quant Time: Aug 03 04:04:32 2017  
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM072617.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Jul 26 17:29:06 2017  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.38  | 152  | 172155   | 20.00  | ng    | -0.01    |
| 21) Naphthalene-d8          | 10.14 | 136  | 592912   | 20.00  | ng    | -0.01    |
| 38) Acenaphthene-d10        | 14.04 | 164  | 408272   | 20.00  | ng    | -0.01    |
| 63) Phenanthrene-d10        | 16.80 | 188  | 1076860  | 20.00  | ng    | -0.01    |
| 75) Chrysene-d12            | 21.02 | 240  | 1153529  | 20.00  | ng    | -0.01    |
| 86) Perylene-d12            | 23.13 | 264  | 946736   | 20.00  | ng    | -0.01    |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.03  | 112  | 1337071  | 131.29 | ng    | 0.00     |
| 7) Phenol-d6                | 6.58  | 99   | 1724003  | 119.15 | ng    | -0.01    |
| 23) Nitrobenzene-d5         | 8.53  | 82   | 1727579  | 109.36 | ng    | -0.01    |
| 41) 2,4,6-Tribromophenol    | 15.54 | 330  | 992023   | 151.60 | ng    | -0.01    |
| 44) 2-Fluorobiphenyl        | 12.65 | 172  | 3155187  | 96.92  | ng    | -0.01    |
| 78) Terphenyl-d14           | 19.46 | 244  | 5431059  | 93.26  | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM080217\  
Data File : BM011099.D  
Acq On : 02 Aug 2017 18:02  
Operator : SJ/JU  
Sample : I4540-03  
Misc :  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
72-12003

Quant Time: Aug 03 04:04:32 2017  
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM072617.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Jul 26 17:29:06 2017  
Response via : Initial Calibration

