

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080322\
 Data File : BM036090.D
 Acq On : 03 Aug 2022 22:33
 Operator : CG/JU
 Sample : N3930-14
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 RVE-106

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 08/04/2022
 Supervised By : Jagrut Upadhyay 08/05/2022

Quant Time: Aug 04 01:32:28 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM080122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 01 17:50:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.863	152	207544	20.00	ng	0.00
21) Naphthalene-d8	10.669	136	983856	20.00	ng	0.00
39) Acenaphthene-d10	14.504	164	635986	20.00	ng	0.00
64) Phenanthrene-d10	17.245	188	1395763	20.00	ng	0.00
76) Chrysene-d12	21.421	240	1523848	20.00	ng	0.00
86) Perylene-d12	23.786	264	1620984	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.428	112	1380266	107.82	ng	0.00
7) Phenol-d6	7.034	99	1883338	115.62	ng	0.00
23) Nitrobenzene-d5	9.034	82	1208713	68.77	ng	0.00
42) 2,4,6-Tribromophenol	15.992	330	850733	114.02	ng	0.00
45) 2-Fluorobiphenyl	13.127	172	2934223	68.12	ng	-0.01
79) Terphenyl-d14	19.868	244	4895079	63.36	ng	0.00
Target Compounds						
75) Fluoranthene	19.298	202	217963	2.74	ng	97
78) Pyrene	19.662	202	235618	2.49	ng	99
88) Benzo(b)fluoranthene	23.062	252	309099m	3.25	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080322\
Data File : BM036090.D
Acq On : 03 Aug 2022 22:33
Operator : CG/JU
Sample : N3930-14
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
RVE-106

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 08/04/2022
Supervised By : Jagrut Upadhyay 08/05/2022

Quant Time: Aug 04 01:32:28 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM080122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 01 17:50:14 2022
Response via : Initial Calibration

