

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080324\
 Data File : BM046958.D
 Acq On : 03 Aug 2024 10:08
 Operator : MA/JU
 Sample : P3276-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E1ZS4

Quant Time: Aug 05 00:54:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM080224.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Aug 03 02:51:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.398	152	105929	20.000	ng/ul	0.00
20) Naphthalene-d8	10.157	136	392737	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.057	164	243449	20.000	ng/ul	0.00
64) Phenanthrene-d10	16.815	188	513278	20.000	ng/ul	0.00
79) Chrysene-d12	21.050	240	508962	20.000	ng/ul	0.00
88) Perylene-d12	23.768	264	647679	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.022	96	6699	2.505	ng/uL	0.00
4) Pyridine-d5	3.416	84	84435	11.577	ng/ul	0.00
7) Phenol-d5	6.604	99	105596	13.123	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.757	67	71236	13.454	ng/ul	0.00
11) 2-Chlorophenol-d4	6.945	132	87082	12.775	ng/ul	0.00
15) 4-Methylphenol-d8	8.116	113	81857	12.841	ng/ul	0.00
21) Nitrobenzene-d5	8.545	128	40241	12.769	ng/ul	0.00
24) 2-Nitrophenol-d4	9.257	143	42725	12.414	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.792	165	78822	11.929	ng/ul	0.00
31) 4-Chloroaniline-d4	10.310	131	108494	12.493	ng/ul	0.00
46) Dimethylphthalate-d6	13.486	166	252675	13.574	ng/ul	0.00
49) Acenaphthylene-d8	13.745	160	277013	13.332	ng/ul	0.00
54) 4-Nitrophenol-d4	14.298	143	33539	9.319	ng/ul	0.00
60) Fluorene-d10	15.063	176	208725	13.108	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.198	200	31182	9.616	ng/ul	0.00
73) Anthracene-d10	16.915	188	325458	13.375	ng/ul	0.00
81) Pyrene-d10	19.227	212	385723	13.239	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.580	264	447221	13.113	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080324\
Data File : BM046958.D
Acq On : 03 Aug 2024 10:08
Operator : MA/JU
Sample : P3276-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
E1ZS4

Quant Time: Aug 05 00:54:35 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM080224.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Aug 03 02:51:11 2024
Response via : Initial Calibration

