

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080324\
 Data File : BM046977.D
 Acq On : 03 Aug 2024 23:29
 Operator : MA/JU
 Sample : P3276-08
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E20H7

Quant Time: Aug 05 00:58:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM080224.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Aug 03 02:51:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.398	152	184102	20.000	ng/u1	0.00
20) Naphthalene-d8	10.157	136	719265	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.057	164	452792	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.815	188	933666	20.000	ng/u1	0.00
79) Chrysene-d12	21.050	240	877289	20.000	ng/u1	0.00
88) Perylene-d12	23.762	264	1070371	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.022	96	11472	2.469	ng/uL	0.00
4) Pyridine-d5	3.410	84	131330	10.361	ng/u1	0.00
7) Phenol-d5	6.604	99	174702	12.492	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.751	67	115934	12.599	ng/u1	0.00
11) 2-Chlorophenol-d4	6.940	132	138537	11.694	ng/u1	0.00
15) 4-Methylphenol-d8	8.116	113	137471	12.408	ng/u1	0.00
21) Nitrobenzene-d5	8.540	128	65565	11.360	ng/u1	0.00
24) 2-Nitrophenol-d4	9.257	143	72580	11.515	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.792	165	133181	11.005	ng/u1	0.00
31) 4-Chloroaniline-d4	10.304	131	186781	11.744	ng/u1	0.00
46) Dimethylphthalate-d6	13.480	166	426986	12.333	ng/u1	0.00
49) Acenaphthylene-d8	13.739	160	488768	12.648	ng/u1	0.00
54) 4-Nitrophenol-d4	14.292	143	64526	9.639	ng/u1	0.00
60) Fluorene-d10	15.057	176	369766	12.485	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.198	200	49176	8.337	ng/u1	0.00
73) Anthracene-d10	16.910	188	572374	12.932	ng/u1	0.00
81) Pyrene-d10	19.227	212	648477	12.913	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.580	264	719330	12.763	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080324\
Data File : BM046977.D
Acq On : 03 Aug 2024 23:29
Operator : MA/JU
Sample : P3276-08
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
E20H7

Quant Time: Aug 05 00:58:54 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM080224.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Aug 03 02:51:11 2024
Response via : Initial Calibration

