


Data Path : Z:\HPCHEM1\BNA M\Data\BM080416\
Data File : BM006857.D
Acq On : 04 Aug 2016 07:09
Operator : UM/SJ
Sample : DFTPP36
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP36

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM072715.M
Title : SVOA CALIBRATION
Last Update : Thu Aug 06 07:44:33 2015



AutoFind: Scans 2521, 2522, 2523; Background Corrected with Scan 2515

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	69.7	219968	PASS
68	69	0.00	2	2.0	2912	PASS
69	198	0.00	100	46.3	146013	PASS
70	69	0.00	2	0.9	1286	PASS
127	198	10	80	46.3	146045	PASS
197	198	0.00	2	0.7	2348	PASS
198	198	100	100	100.0	315392	PASS
199	198	5	9	7.4	23357	PASS
275	198	10	60	29.7	93568	PASS
365	198	1	100	6.2	19460	PASS
441	443	0.01	100	78.3	44434	PASS
442	198	50	100	89.3	281578	PASS
443	442	15	24	20.1	56725	PASS

m/z	Abundance
39.50	402.0
40.00	1193.0
41.00	457.0
43.00	1013.0
44.00	1813.0
45.10	576.0
48.90	729.0
52.90	518.0
56.90	688.0
58.90	327.0
65.30	510.0
73.00	4823.0
74.00	481.0
76.80	350.0
83.20	524.0
84.00	808.0
88.80	329.0
91.20	316.0
95.00	510.0
96.80	514.0
132.90	499.0
135.00	465.0
147.00	1265.0
147.60	371.0
148.10	329.0
149.10	1056.0
177.20	357.0
191.20	1182.0
193.10	920.0
207.00	6680.0
208.10	1364.0
209.10	1175.0
221.10	588.0
267.00	517.0
280.90	1407.0
282.10	636.0
355.20	641.0
549.20	541.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

m/z	Abundance
35.90	429.0
37.10	1087.0
38.10	3576.0
39.10	16536.0
40.10	1954.0
41.20	1266.0
42.90	780.0
44.00	2693.0
45.00	384.0
49.10	3517.0
50.00	73776.0
51.10	236928.0
52.00	11990.0
53.10	686.0
54.90	1572.0
56.00	5853.0
57.00	13679.0
58.00	1233.0
59.00	534.0
60.90	2875.0
62.10	3360.0
63.00	9938.0
64.00	2116.0
65.00	2789.0
68.10	2872.0
69.00	152960.0
70.20	1752.0
73.10	6100.0
74.00	26184.0
75.00	28776.0
76.10	10189.0
77.10	174016.0
78.00	11509.0
79.00	16336.0
80.10	11752.0
81.00	13030.0
81.90	3876.0
83.10	2652.0
85.00	3086.0
86.10	4476.0
87.00	2417.0
87.90	882.0
88.70	320.0
91.00	3260.0
92.10	3541.0
93.00	30064.0
94.10	2320.0
95.10	556.0
96.10	841.0
96.80	410.0
98.00	29208.0
99.10	14930.0
99.90	1763.0
101.00	6038.0
101.90	482.0
103.00	2524.0
104.00	4387.0
105.00	3707.0
106.00	1981.0
107.00	64112.0
108.00	8367.0
108.90	1653.0
110.00	76720.0
111.00	14029.0
111.90	2358.0
112.90	351.0
113.10	357.0
115.00	768.0
116.00	4931.0
117.00	90720.0
118.00	5615.0
118.90	1065.0
119.90	504.0
121.20	490.0
121.90	4193.0
123.00	5055.0
124.00	4125.0
125.00	2720.0
126.10	540.0
127.00	156480.0
128.00	10686.0
129.00	66864.0
130.00	5573.0
131.00	2110.0
131.90	614.0
133.00	1218.0
133.90	2251.0
135.00	5627.0
135.90	2569.0
136.80	1791.0
137.20	2542.0
137.80	622.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

138.90	626.0
139.70	801.0
140.20	1295.0
141.00	11461.0
142.00	4348.0
143.00	1971.0
144.00	859.0
145.90	2387.0
147.00	6173.0
148.00	19152.0
149.00	3827.0
150.00	888.0
151.00	802.0
151.90	798.0
153.10	3744.0
154.10	2931.0
155.10	4924.0
156.10	10941.0
156.90	1635.0
158.00	1519.0
159.00	1105.0
159.80	3316.0
161.10	4342.0
162.00	1757.0
162.90	489.0
164.10	898.0
165.00	4545.0
166.00	3137.0
167.00	27072.0
168.00	15687.0
169.00	1123.0
170.00	322.0
171.20	1622.0
172.00	1976.0
172.90	3668.0
174.10	5058.0
175.00	7676.0
176.00	1999.0
176.70	2935.0
177.10	3715.0
178.10	1718.0
179.00	18752.0
180.00	10840.0
181.00	5907.0
182.10	644.0
184.10	2219.0
185.00	8992.0
186.00	56648.0
187.10	17512.0
187.90	1263.0
189.00	4722.0
190.00	1146.0
191.00	2428.0
192.00	5852.0
193.00	6130.0
194.20	919.0
195.00	528.0
196.10	7599.0
197.10	2798.0
198.00	301312.0
199.00	22568.0
199.90	2607.0
201.00	773.0
201.80	561.0
203.10	2599.0
204.10	13524.0
205.10	22376.0
206.10	92824.0
207.00	16440.0
208.00	6821.0
208.90	2604.0
210.10	2112.0
211.10	5492.0
211.90	1148.0
213.00	306.0
215.00	895.0
216.10	3878.0
217.00	29104.0
217.90	3241.0
219.00	958.0
220.20	674.0
221.00	8569.0
221.80	4048.0
223.10	6273.0
224.00	54024.0
225.10	13184.0
226.10	1397.0
227.00	28648.0
228.00	5130.0
229.00	5729.0
230.20	1860.0
231.00	2353.0
232.00	896.0
234.00	2582.0
235.00	2617.0
236.10	1641.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

237.10	2998.0
238.00	886.0
239.00	1391.0
240.20	1203.0
240.80	967.0
242.00	2864.0
243.00	3405.0
244.10	35472.0
245.00	6350.0
245.90	10629.0
247.00	2804.0
247.90	803.0
249.00	1580.0
250.20	538.0
251.00	646.0
251.90	790.0
253.00	1534.0
253.90	2337.0
255.00	208000.0
256.00	30112.0
257.10	2200.0
258.10	20536.0
259.00	2581.0
262.90	468.0
265.00	7089.0
265.90	748.0
267.00	1117.0
270.20	344.0
271.20	368.0
272.00	341.0
273.00	5777.0
274.00	17320.0
275.10	82880.0
276.00	11525.0
277.00	10838.0
278.10	2701.0
278.90	1228.0
281.10	2100.0
282.10	789.0
283.00	1479.0
284.00	1170.0
285.10	1714.0
286.10	477.0
288.90	318.0
293.00	1232.0
294.10	570.0
296.10	33416.0
297.10	3233.0
301.00	524.0
302.10	455.0
303.00	3029.0
304.20	603.0
308.00	948.0
309.20	593.0
310.00	820.0
313.80	873.0
314.20	627.0
315.00	4086.0
315.90	1779.0
319.90	353.0
321.00	524.0
322.10	367.0
323.10	8529.0
324.20	957.0
325.90	368.0
326.90	2390.0
327.90	1294.0
329.00	373.0
331.90	1079.0
333.00	1470.0
334.10	5852.0
335.10	2077.0
340.90	1078.0
342.20	718.0
346.00	2110.0
352.00	2008.0
353.00	1179.0
353.90	2690.0
355.10	1249.0
356.10	875.0
357.10	454.0
360.30	323.0
365.00	16252.0
366.10	2165.0
372.10	4033.0
372.80	768.0
382.90	1211.0
390.30	649.0
401.20	452.0
402.10	1370.0
403.00	2424.0
404.10	1246.0
405.20	457.0
421.00	1861.0
422.00	1588.0
423.10	8720.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

424.00	2936.0
424.80	992.0
441.10	31544.0
442.10	198976.0
443.10	44800.0
444.10	4686.0
445.10	938.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

m/z	Abundance
37.00	642.0
38.00	3269.0
39.10	15936.0
40.10	2118.0
41.10	1802.0
43.00	682.0
44.00	3435.0
45.00	2118.0
47.80	370.0
49.10	2686.0
50.10	75888.0
51.00	253696.0
52.00	12795.0
53.00	773.0
55.00	1847.0
56.00	6784.0
57.00	14288.0
58.00	520.0
58.90	1109.0
61.10	1661.0
62.00	3533.0
63.00	11231.0
63.90	1186.0
64.90	2828.0
65.90	644.0
66.90	437.0
67.90	3985.0
69.00	171264.0
70.00	974.0
73.10	6050.0
74.00	22576.0
75.10	34432.0
76.00	12643.0
77.10	199616.0
78.10	15709.0
79.00	17888.0
80.00	10331.0
81.00	15566.0
82.00	3348.0
83.00	2769.0
83.90	541.0
85.00	2365.0
86.00	6516.0
87.00	2601.0
88.00	869.0
90.90	5517.0
92.00	5752.0
93.00	39888.0
94.00	2269.0
95.00	781.0
96.10	1358.0
96.90	674.0
98.00	29608.0
99.10	18720.0
100.00	1545.0
100.90	7815.0
101.90	925.0
102.90	1898.0
104.00	4972.0
105.00	6344.0
106.10	1947.0
107.00	75440.0
108.00	8904.0
109.10	1910.0
110.00	93784.0
111.10	14332.0
111.90	2495.0
113.00	771.0
114.80	427.0
116.00	4993.0
117.00	98872.0
118.00	8190.0
119.10	730.0
120.10	906.0
120.90	436.0
122.00	4935.0
122.90	7473.0
124.00	3694.0
125.10	2057.0
125.90	851.0
127.00	163520.0
128.00	12419.0
129.00	83432.0
130.00	6047.0
130.90	1871.0
132.00	1743.0
133.00	1967.0
134.10	3657.0
135.00	5962.0
136.00	1659.0
136.90	3549.0
137.90	527.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

139.00	1509.0
139.90	2085.0
141.00	12503.0
142.10	4102.0
143.10	2696.0
143.90	328.0
144.90	847.0
146.00	2506.0
147.10	6793.0
148.10	21536.0
149.00	2469.0
150.00	1337.0
151.20	1080.0
151.90	647.0
153.00	5929.0
154.10	2975.0
155.00	7189.0
156.00	11426.0
157.00	1947.0
158.00	1551.0
159.00	1819.0
159.80	4588.0
161.00	6815.0
161.90	1861.0
163.80	440.0
164.20	368.0
165.00	5819.0
166.00	5217.0
167.00	36016.0
168.10	17296.0
168.90	1413.0
169.90	730.0
170.90	568.0
172.00	2430.0
173.10	3375.0
174.10	6002.0
175.00	12427.0
176.00	1762.0
176.90	5288.0
178.00	2023.0
179.00	23320.0
180.00	13797.0
180.90	4917.0
182.00	1362.0
182.80	750.0
184.10	2435.0
185.00	10302.0
186.10	64504.0
187.10	19472.0
188.00	2367.0
189.00	4747.0
190.00	1509.0
190.90	4138.0
191.90	8149.0
193.00	8064.0
194.00	2566.0
194.90	699.0
196.00	9121.0
196.90	2776.0
198.00	360256.0
198.90	27456.0
200.00	2879.0
200.90	424.0
201.80	878.0
203.10	3817.0
204.10	20408.0
205.00	29784.0
206.10	114608.0
207.00	19848.0
208.00	8079.0
209.00	3126.0
210.00	2446.0
210.90	3474.0
212.00	482.0
215.00	2208.0
216.00	4726.0
217.00	39904.0
218.00	5085.0
219.10	415.0
221.10	8425.0
222.00	4581.0
223.10	10016.0
224.10	73240.0
225.00	17928.0
226.10	1822.0
227.00	38528.0
228.10	5665.0
229.00	7427.0
230.10	691.0
231.10	1912.0
231.90	352.0
233.10	849.0
234.10	2710.0
235.00	2957.0
236.10	2969.0
237.10	3439.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

239.10	1910.0
240.00	821.0
241.00	2527.0
242.10	2557.0
243.00	4607.0
244.10	53224.0
245.00	9545.0
246.00	13739.0
247.00	2822.0
248.90	2391.0
250.10	305.0
250.80	445.0
251.10	498.0
252.00	608.0
253.10	2043.0
254.00	2719.0
255.00	274624.0
256.10	41120.0
256.90	3758.0
258.10	23352.0
259.00	3748.0
259.90	553.0
265.00	10381.0
266.00	1897.0
266.90	776.0
269.20	326.0
270.10	554.0
270.60	388.0
271.10	414.0
271.90	1090.0
273.00	7866.0
274.00	24792.0
275.00	113360.0
276.00	12698.0
277.00	13944.0
278.00	2910.0
279.00	707.0
281.20	3397.0
282.00	810.0
282.90	2424.0
284.20	1092.0
285.00	2065.0
289.00	840.0
290.20	652.0
293.00	2658.0
294.00	1097.0
295.20	439.0
296.00	41040.0
297.10	6661.0
298.10	468.0
300.90	597.0
302.10	567.0
303.00	3977.0
304.10	1196.0
308.00	713.0
310.00	794.0
312.70	496.0
313.10	356.0
314.10	1570.0
315.00	5257.0
316.10	2690.0
317.00	482.0
319.90	341.0
321.00	1136.0
322.00	755.0
323.10	13183.0
324.00	2292.0
327.00	3107.0
328.00	1433.0
329.00	851.0
331.80	482.0
333.20	866.0
334.10	7271.0
335.00	2887.0
341.00	2381.0
342.00	376.0
346.10	2511.0
347.00	855.0
352.00	3489.0
353.10	2644.0
354.10	2433.0
355.10	2123.0
356.20	304.0
365.00	20896.0
366.00	4038.0
370.00	556.0
372.00	7280.0
373.00	2285.0
373.80	306.0
383.10	1769.0
384.10	790.0
390.20	801.0
391.20	370.0
401.10	473.0
402.00	2859.0
403.10	4977.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

404.00	894.0
405.10	318.0
415.00	668.0
416.30	429.0
421.10	3385.0
422.10	2738.0
423.00	16504.0
424.00	5480.0
424.90	702.0
429.20	421.0
441.10	51096.0
442.10	339392.0
443.10	63432.0
444.10	6433.0
445.20	336.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

m/z	Abundance
35.90	556.0
36.80	881.0
38.10	2365.0
39.10	10780.0
40.00	1577.0
41.00	313.0
43.00	773.0
44.00	1372.0
44.90	490.0
49.10	2336.0
50.10	46600.0
51.00	169280.0
52.10	9271.0
53.20	825.0
55.20	1586.0
56.00	4528.0
57.00	10728.0
57.90	360.0
59.00	343.0
61.10	1611.0
62.10	1861.0
63.10	4961.0
63.90	829.0
65.10	2888.0
66.10	303.0
67.90	1881.0
69.00	113816.0
70.00	1134.0
73.00	5665.0
74.10	15748.0
75.00	20624.0
76.00	9287.0
77.10	134272.0
78.10	8946.0
79.00	12213.0
80.10	8683.0
81.00	9273.0
81.90	2490.0
83.10	1153.0
83.80	790.0
85.00	1838.0
86.00	4014.0
87.00	1728.0
88.00	921.0
91.00	2490.0
92.10	2668.0
93.00	25328.0
94.00	2155.0
96.00	1446.0
96.80	360.0
98.00	20528.0
99.00	13675.0
100.00	1473.0
100.90	6153.0
101.80	332.0
102.90	1474.0
103.90	2809.0
105.10	3592.0
106.10	469.0
107.00	46384.0
108.00	7022.0
109.10	858.0
110.00	61168.0
110.90	11950.0
112.10	2246.0
112.90	520.0
116.00	1981.0
117.00	73152.0
118.00	5216.0
119.10	794.0
120.10	682.0
121.20	516.0
122.00	5021.0
122.90	4138.0
123.80	3083.0
125.00	2181.0
126.10	946.0
127.00	118136.0
128.00	9686.0
129.00	57640.0
130.00	5253.0
131.00	1265.0
132.00	971.0
133.10	1177.0
133.90	2381.0
134.90	5325.0
136.10	1849.0
137.00	3537.0
138.00	536.0
139.00	685.0
140.00	2174.0
140.90	8149.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

142.00	1928.0
143.10	2148.0
144.00	999.0
146.00	2042.0
147.10	7148.0
148.00	15801.0
149.00	2612.0
150.00	813.0
151.10	1238.0
151.90	692.0
152.80	1909.0
153.20	1703.0
154.10	1772.0
155.00	5792.0
156.00	8512.0
157.00	1636.0
158.00	1159.0
158.90	1957.0
160.00	3520.0
161.00	5288.0
162.10	1274.0
162.80	378.0
165.00	3765.0
166.00	4727.0
167.00	27744.0
168.10	11577.0
169.00	1739.0
169.80	793.0
170.90	1688.0
172.00	2019.0
172.90	2171.0
174.00	4433.0
175.00	7964.0
175.90	2419.0
177.00	4664.0
177.90	1136.0
178.90	15090.0
180.10	8942.0
181.00	3725.0
181.80	406.0
182.70	828.0
184.00	928.0
185.10	6027.0
186.00	51432.0
187.00	14810.0
188.10	1493.0
189.00	4676.0
190.00	454.0
191.00	1808.0
192.00	6043.0
193.00	6471.0
194.00	989.0
194.60	311.0
196.00	7367.0
197.00	1470.0
198.00	284608.0
199.00	20048.0
199.90	1488.0
201.20	809.0
203.00	3207.0
204.00	12420.0
205.00	18136.0
206.10	92240.0
207.00	16568.0
208.00	4268.0
209.00	2236.0
209.90	1787.0
211.00	5050.0
211.80	917.0
215.10	1954.0
215.90	3397.0
217.00	27448.0
218.00	4391.0
219.80	599.0
221.00	8719.0
221.90	3375.0
223.00	6299.0
224.10	54168.0
225.10	14424.0
226.20	1423.0
227.00	31000.0
227.90	3981.0
229.00	4182.0
230.10	641.0
231.20	2257.0
231.90	370.0
232.90	1221.0
234.00	2427.0
235.00	2152.0
235.90	2131.0
237.10	3093.0
237.90	452.0
239.10	1124.0
239.90	781.0
241.10	1008.0
242.00	4089.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

243.20	5351.0
244.10	40480.0
245.10	5589.0
246.00	11818.0
247.10	1657.0
248.00	473.0
249.00	1617.0
251.00	437.0
251.70	541.0
252.20	832.0
253.10	1910.0
254.20	2328.0
255.00	220544.0
256.00	31576.0
256.90	1521.0
258.00	18312.0
259.10	2680.0
259.80	627.0
261.10	432.0
265.10	8447.0
266.00	1561.0
267.60	322.0
269.00	531.0
271.00	507.0
272.00	969.0
273.10	6756.0
274.00	22688.0
275.00	84464.0
276.10	11297.0
277.00	11032.0
278.00	2512.0
279.00	897.0
281.00	4248.0
282.10	818.0
283.10	1874.0
284.00	661.0
285.00	2508.0
286.00	373.0
289.00	644.0
289.80	352.0
291.00	334.0
292.80	1537.0
293.20	1547.0
294.00	1058.0
295.00	979.0
296.00	43200.0
297.00	5432.0
297.90	614.0
300.90	855.0
303.00	4927.0
303.90	821.0
307.90	414.0
308.80	747.0
310.00	545.0
313.10	864.0
313.90	1073.0
314.90	3114.0
316.00	2199.0
316.90	383.0
321.10	830.0
322.00	468.0
323.10	10006.0
324.10	1991.0
327.00	2118.0
327.90	856.0
332.10	1126.0
333.10	1019.0
334.10	7444.0
334.90	1378.0
341.00	1425.0
342.10	481.0
346.10	2525.0
347.00	616.0
351.90	2778.0
353.00	1319.0
354.10	3428.0
355.00	1221.0
359.20	397.0
365.00	21232.0
366.10	2993.0
367.20	372.0
370.10	600.0
370.90	464.0
372.10	4752.0
373.10	911.0
373.70	312.0
383.00	1534.0
383.90	326.0
389.90	641.0
390.90	397.0
392.00	305.0
401.10	630.0
402.10	2114.0
403.00	2227.0
403.90	761.0
414.90	300.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36

421.00	2954.0
422.10	3606.0
423.00	17368.0
424.10	4535.0
425.00	323.0
441.10	50664.0
442.10	306368.0
443.10	61944.0
444.10	4367.0
445.10	384.0

Instrument :
BNA_M
ClientSampleId :
DFTPP36