

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM080516\
 Data File : BM006923.D
 Acq On : 06 Aug 2016 05:09
 Operator : UM/SJ
 Sample : H4278-14
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BDJS5

Quant Time: Aug 06 05:54:31 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080316.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Aug 06 00:33:11 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.56	152	158879	20.00	ng/ul	0.00
18) Naphthalene-d8	10.33	136	653489	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.21	164	388971	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.97	188	919057	20.00	ng/ul	0.00
75) Chrysene-d12	21.18	240	1052046	20.00	ng/ul	0.00
83) Perylene-d12	23.37	264	1064411	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.10	96	5401	1.87	ng/uL	0.00
5) Phenol-d5	6.78	99	200042	17.11	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.91	67	146350	18.73	ng/ul	0.00
9) 2-Chlorophenol-d4	7.11	132	168118	18.19	ng/ul	0.00
13) 4-Methylphenol-d8	8.30	113	170809	17.96	ng/ul	0.00
19) Nitrobenzene-d5	8.72	128	76825	17.19	ng/ul	0.00
22) 2-Nitrophenol-d4	9.45	143	85867	16.14	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.00	165	148295	13.48	ng/ul	0.00
29) 4-Chloroaniline-d4	10.51	131	201380	17.72	ng/ul	0.00
43) Dimethylphthalate-d6	13.63	166	625887	19.53	ng/ul	0.00
46) Acenaphthylene-d8	13.90	160	744999	19.04	ng/ul	0.00
51) 4-Nitrophenol-d4	14.54	143	14802	2.11	ng/ul	0.01
57) Fluorene-d10	15.22	176	548751	18.45	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.39	200	50157	8.36	ng/ul	0.01
70) Anthracene-d10	17.07	188	831295	18.83	ng/ul	0.00
76) Pyrene-d10	19.38	212	978837	18.19	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.23	264	968588	18.69	ng/ul	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
44) Dimethylphthalate	13.69	163	237609	7.46	ng/ul#	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM080516\
Data File : BM006923.D
Acq On : 06 Aug 2016 05:09
Operator : UM/SJ
Sample : H4278-14
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BDJS5

Quant Time: Aug 06 05:54:31 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080316.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Aug 06 00:33:11 2016
Response via : Initial Calibration





#44

Dimethylphthalate

Concen: 7.46 ng/ul

RT: 13.69 min Scan# 1887

Delta R.T. 0.00 min

Lab File: BM006923.D

Acq: 06 Aug 2016 05:09

Instrument :

BNA_M

ClientSampleId :

BDJS5


Tgt Ion:163 Resp: 237609

Ion Ratio Lower Upper

163 100

194 5.1 3.1 4.7#

164 10.4 8.6 12.8



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E

