

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM080516\
 Data File : BM006900.D
 Acq On : 05 Aug 2016 15:15
 Operator : UM/SJ
 Sample : PB92544BL
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK44

Manual Integrations
 APPROVED

sohil
 8/8/2016 7:01:32 PM

Quant Time: Aug 06 01:05:48 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080316.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Aug 06 00:33:11 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.56	152	119002	20.00	ng/ul	0.00
18) Naphthalene-d8	10.33	136	498570	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.22	164	301719	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.97	188	707323	20.00	ng/ul	0.00
75) Chrysene-d12	21.19	240	790656	20.00	ng/ul	0.00
83) Perylene-d12	23.37	264	729649	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.10	96	7059	3.26	ng/uL	0.00
5) Phenol-d5	6.78	99	261742	29.88	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.91	67	191575	32.74	ng/ul	0.00
9) 2-Chlorophenol-d4	7.11	132	223331	32.27	ng/ul	0.00
13) 4-Methylphenol-d8	8.31	113	220354	30.93	ng/ul	0.00
19) Nitrobenzene-d5	8.73	128	108382	31.80	ng/ul	0.00
22) 2-Nitrophenol-d4	9.45	143	116418	28.68	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.00	165	205841	24.53	ng/ul	0.00
29) 4-Chloroaniline-d4	10.51	131	280314	32.34	ng/ul	0.00
43) Dimethylphthalate-d6	13.64	166	794167	31.95	ng/ul	0.00
46) Acenaphthylene-d8	13.90	160	939862	30.97	ng/ul	0.00
51) 4-Nitrophenol-d4	14.53	143	94065m	17.26	ng/ul	0.00
57) Fluorene-d10	15.22	176	682282	29.57	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.37	200	97256m	21.07	ng/ul	0.00
70) Anthracene-d10	17.07	188	1097587	32.31	ng/ul	0.00
76) Pyrene-d10	19.38	212	1256461	31.06	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.23	264	1148986	32.34	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM080516\
Data File : BM006900.D
Acq On : 05 Aug 2016 15:15
Operator : UM/SJ
Sample : PB92544BL
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK44

Quant Time: Aug 06 01:05:48 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080316.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Aug 06 00:33:11 2016
Response via : Initial Calibration

Manual Integrations
APPROVED

sohil
8/8/2016 7:01:32 PM

