

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080521\
 Data File : BM031494.D
 Acq On : 05 Aug 2021 22:37
 Operator : CG/JU
 Sample : M3162-10
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BGEW1

Manual Integrations
 APPROVED

mohammad
 8/6/2021 2:49:24 PM

Quant Time: Aug 06 01:40:28 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 05 11:03:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.587	152	1303	0.400	ng/ul	0.00
4) Naphthalene-d8	10.347	136	4987	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.220	164	3091	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.968	188	5823	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.164	240	4315	0.400	ng/ul	0.00
23) Perylene-d12	23.319	264	4421	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.186	96	6067	4.192	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.944	152	2788	0.332	ng/ul	0.00
18) Fluoranthene-d10	19.005	212	5408	0.376	ng/ul	0.00
Target Compounds						
					Qvalue	
15) Phenanthrene	17.013	178	528	0.029	ng/ul#	76
19) Fluoranthene	19.031	202	1034	0.056	ng/ul#	80
20) Pyrene	19.394	202	1293	0.069	ng/ul#	76
21) Benzo(a)anthracene	21.149	228	551	0.032	ng/ul#	71
22) Chrysene	21.200	228	620	0.034	ng/ul#	84
24) Benzo(b)fluoranthene	22.684	252	695m	0.035	ng/ul	
26) Benzo(a)pyrene	23.224	252	486	0.028	ng/ul#	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080521\
Data File : BM031494.D
Acq On : 05 Aug 2021 22:37
Operator : CG/JU
Sample : M3162-10
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BGEW1

Manual Integrations
APPROVED

mohammad
8/6/2021 2:49:24 PM

Quant Time: Aug 06 01:40:28 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 05 11:03:39 2021
Response via : Initial Calibration

