

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080521\
 Data File : BM031519.D
 Acq On : 06 Aug 2021 17:13
 Operator : CG/JU
 Sample : PB138020BL
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK020

Quant Time: Aug 06 17:49:28 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 06 17:49:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.587	152	959	0.400	ng/ul	0.00
4) Naphthalene-d8	10.347	136	3510	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.216	164	2100	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.964	188	3788	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.166	240	2848	0.400	ng/ul	0.00
23) Perylene-d12	23.319	264	3214	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.189	96	7144	6.707	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.944	152	2088	0.353	ng/ul	0.00
18) Fluoranthene-d10	19.001	212	3789	0.399	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080521\
Data File : BM031519.D
Acq On : 06 Aug 2021 17:13
Operator : CG/JU
Sample : PB138020BL
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK020

Quant Time: Aug 06 17:49:28 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 06 17:49:17 2021
Response via : Initial Calibration

