

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080521\
 Data File : BM031520.D
 Acq On : 06 Aug 2021 17:50
 Operator : CG/JU
 Sample : PB138020BS
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS020

Quant Time: Aug 06 18:20:43 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 06 17:49:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.587	152	777	0.400	ng/ul	0.00
4) Naphthalene-d8	10.352	136	2820	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.220	164	1540	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.971	188	2919	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.164	240	2264	0.400	ng/ul	0.00
23) Perylene-d12	23.321	264	2095	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.192	96	617	0.715	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.944	152	1697	0.357	ng/ul	0.00
18) Fluoranthene-d10	19.001	212	3075	0.407	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.223	88	1686	1.934	ng/ul#	80
5) Naphthalene	10.397	128	2923	0.348	ng/ul	99
7) 2-Methylnaphthalene	12.020	142	1954	0.340	ng/ul	99
8) 1-Methylnaphthalene	12.241	142	1784	0.315	ng/ul	99
10) Acenaphthylene	13.941	152	2384	0.363	ng/ul	97
11) Acenaphthene	14.284	153	1939	0.351	ng/ul	99
12) Fluorene	15.278	166	2134	0.336	ng/ul	97
14) Pentachlorophenol	16.628	266	619	0.662	ng/ul	99
15) Phenanthrene	17.013	178	3310	0.357	ng/ul	97
16) Anthracene	17.103	178	3077	0.355	ng/ul	98
19) Fluoranthene	19.031	202	3702	0.383	ng/ul	96
20) Pyrene	19.397	202	3766	0.384	ng/ul	98
21) Benzo(a)anthracene	21.149	228	3405	0.372	ng/ul	96
22) Chrysene	21.200	228	3626	0.383	ng/ul	99
24) Benzo(b)fluoranthene	22.682	252	3596	0.381	ng/ul	90
25) Benzo(k)fluoranthene	22.723	252	3562	0.365	ng/ul#	91
26) Benzo(a)pyrene	23.229	252	3099	0.375	ng/ul#	89
27) Indeno(1,2,3-cd)pyrene	25.461	276	3931	0.368	ng/ul#	98
28) Dibenzo(a,h)anthracene	25.475	278	3033	0.352	ng/ul	93
29) Benzo(g,h,i)perylene	26.107	276	3527	0.387	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080521\
Data File : BM031520.D
Acq On : 06 Aug 2021 17:50
Operator : CG/JU
Sample : PB138020BS
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS020

Quant Time: Aug 06 18:20:43 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 06 17:49:17 2021
Response via : Initial Calibration

