

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080521\
 Data File : BM031586.D
 Acq On : 08 Aug 2021 13:33
 Operator : CG/JU
 Sample : M3163-03
 Misc :
 ALS Vial : 112 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BGEX3

Manual Integrations
 APPROVED

mohammad
 8/9/2021 12:15:12 PM

Quant Time: Aug 09 04:14:37 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 06 18:39:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.580	152	1237	0.400	ng/ul	0.00
4) Naphthalene-d8	10.338	136	4572	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.212	164	2630	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.961	188	4995	0.400	ng/ul	#-0.01
17) Chrysene-d12	21.159	240	4031	0.400	ng/ul	# 0.00
23) Perylene-d12	23.312	264	4355	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.186	96	4992	3.634	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.935	152	1903	0.247	ng/ul	0.00
18) Fluoranthene-d10	18.993	212	3834	0.285	ng/ul	0.00
Target Compounds						
					Qvalue	
15) Phenanthrene	17.002	178	554	0.035	ng/ul#	76
19) Fluoranthene	19.028	202	1364	0.079	ng/ul#	87
20) Pyrene	19.386	202	1135	0.065	ng/ul#	72
21) Benzo(a)anthracene	21.142	228	448	0.028	ng/ul#	68
22) Chrysene	21.193	228	634	0.038	ng/ul#	77
24) Benzo(b)fluoranthene	22.674	252	911m	0.046	ng/ul	
26) Benzo(a)pyrene	23.217	252	351	0.020	ng/ul#	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080521\
Data File : BM031586.D
Acq On : 08 Aug 2021 13:33
Operator : CG/JU
Sample : M3163-03
Misc :
ALS Vial : 112 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BGEX3

Manual Integrations
APPROVED

mohammad
8/9/2021 12:15:12 PM

Quant Time: Aug 09 04:14:37 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM072821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 06 18:39:21 2021
Response via : Initial Calibration

