

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM080616\
 Data File : BM006973.D
 Acq On : 07 Aug 2016 12:53
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02024

Quant Time: Aug 08 03:56:20 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080616.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 08 03:51:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	0.00
2 S	1,4-Dioxane-d8	0.438	0.384	12.3	107	0.00
3 S	Phenol-d5	1.896	1.888	0.4	133	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	1.189	1.204	-1.3	128	0.00
5 S	2-Chlorophenol-d4	1.354	1.417	-4.7	132	0.00
6 S	4-Methylphenol-d8	1.583	1.555	1.8	131	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	133	0.00
8 S	Nitrobenzene-d5	0.149	0.157	-5.4	137	0.00
9 S	2-Nitrophenol-d4	0.178	0.184	-3.4	129	0.00
10 S	2,4-Dichlorophenol-d3	0.337	0.343	-1.8	132	0.00
11	Naphthalene	1.027	1.046	-1.9	133	0.00
12 S	4-Chloroaniline-d4	0.403	0.429	-6.5	131	0.00
13	2-Methylnaphthalene	0.822	0.820	0.2	130	0.00
14	1-Methylnaphthalene	0.790	0.778	1.5	128	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	126	0.00
16 S	Dimethylphthalate-d6	1.760	1.712	2.7	120	0.00
17 S	Acenaphthylene-d8	2.036	2.069	-1.6	125	0.00
18	Acenaphthylene	2.136	2.172	-1.7	126	0.00
19 C	Acenaphthene	1.393	1.402	-0.6	126	0.00
20 S	4-Nitrophenol-d4	0.255	0.202	20.8#	120	-0.01
21 S	Fluorene-d10	1.496	1.476	1.3	124	0.00
22	Fluorene	1.761	1.721	2.3	122	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
24 S	4,6-Dinitro-2-methylphenol-	0.083	0.115	-38.6#	227#	0.00
25	Phenanthrene	1.119	1.118	0.1	118	0.00
26 S	Anthracene-d10	0.966	0.976	-1.0	119	0.00
27	Anthracene	1.171	1.170	0.1	118	0.00
28 C	Fluoranthene	1.361	1.377	-1.2	116	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
30 S	Pyrene-d10	0.953	0.981	-2.9	114	0.00
31	Pyrene	1.243	1.272	-2.3	114	0.00
32	Benzo(a)anthracene	1.187	1.209	-1.9	108	0.00
33	Chrysene	1.045	1.057	-1.1	105	0.00
34 I	Perylene-d12	1.000	1.000	0.0	89	0.00
35	Benzo(b)fluoranthene	1.305	1.306	-0.1	93	0.00
36	Benzo(k)fluoranthene	1.260	1.273	-1.0	90	0.00
37 S	Benzo(a)pyrene-d12	0.953	0.953	0.0	88	0.00
38 C	Benzo(a)pyrene	1.184	1.169	1.3	88	0.00
39	Indeno(1,2,3-cd)pyrene	1.201	1.281	-6.7	95	-0.01
40	Dibenzo(a,h)anthracene	1.011	1.079	-6.7	96	-0.01
41	Benzo(g,h,i)perylene	0.946	1.024	-8.2	98	-0.02

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM080616\
Data File : BM006973.D
Acq On : 07 Aug 2016 12:53
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02024

Quant Time: Aug 08 03:56:20 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080616.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Aug 08 03:51:46 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0