

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080624\
 Data File : BM047028.D
 Acq On : 06 Aug 2024 18:32
 Operator : RC/JU
 Sample : PB162449BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Aug 07 01:21:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM080224.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 07 01:21:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.398	152	198274	20.000	ng/ul	0.00
20) Naphthalene-d8	10.151	136	769912	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.051	164	469248	20.000	ng/ul	0.00
64) Phenanthrene-d10	16.810	188	905944	20.000	ng/ul	0.00
79) Chrysene-d12	21.045	240	753622	20.000	ng/ul	0.00
88) Perylene-d12	23.750	264	860885	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.022	96	28380	5.671	ng/uL	0.00
4) Pyridine-d5	3.410	84	417053	30.551	ng/ul	0.00
7) Phenol-d5	6.598	99	487888	32.392	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.751	67	322629	32.555	ng/ul	0.00
11) 2-Chlorophenol-d4	6.940	132	384951	30.172	ng/ul	0.00
15) 4-Methylphenol-d8	8.116	113	380401	31.881	ng/ul	0.00
21) Nitrobenzene-d5	8.540	128	191977	31.074	ng/ul	0.00
24) 2-Nitrophenol-d4	9.257	143	210204	31.156	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.792	165	378156	29.193	ng/ul	0.00
31) 4-Chloroaniline-d4	10.304	131	520352	30.565	ng/ul	0.00
46) Dimethylphthalate-d6	13.480	166	1085318	30.249	ng/ul	0.00
49) Acenaphthylene-d8	13.739	160	1238484	30.924	ng/ul	0.00
54) 4-Nitrophenol-d4	14.292	143	179256	25.840	ng/ul	0.00
60) Fluorene-d10	15.057	176	904220	29.461	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.192	200	140108	24.479	ng/ul	0.00
73) Anthracene-d10	16.910	188	1319676	30.728	ng/ul	0.00
81) Pyrene-d10	19.221	212	1435858	33.283	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.568	264	1420829	31.343	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080624\
Data File : BM047028.D
Acq On : 06 Aug 2024 18:32
Operator : RC/JU
Sample : PB162449BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Aug 07 01:21:58 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM080224.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Aug 07 01:21:15 2024
Response via : Initial Calibration

