

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080624\
 Data File : BM047035.D
 Acq On : 06 Aug 2024 23:19
 Operator : RC/JU
 Sample : P3328-15
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Aug 07 01:24:04 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM080224.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 07 01:21:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.398	152	281971	20.000	ng/ul	0.00
20) Naphthalene-d8	10.157	136	1076781	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.057	164	678705	20.000	ng/ul	0.00
64) Phenanthrene-d10	16.815	188	1387496	20.000	ng/ul	0.00
79) Chrysene-d12	21.045	240	1100858	20.000	ng/ul	0.00
88) Perylene-d12	23.762	264	1210269	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.022	96	20078	2.821	ng/uL	0.00
4) Pyridine-d5	3.410	84	304142	15.666	ng/ul	0.00
7) Phenol-d5	6.604	99	434163	20.269	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.757	67	262406	18.619	ng/ul	0.00
11) 2-Chlorophenol-d4	6.940	132	343370	18.924	ng/ul	0.00
15) 4-Methylphenol-d8	8.116	113	353187	20.814	ng/ul	0.00
21) Nitrobenzene-d5	8.545	128	165322	19.133	ng/ul	0.00
24) 2-Nitrophenol-d4	9.257	143	184143	19.515	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.792	165	358899	19.811	ng/ul	0.00
31) 4-Chloroaniline-d4	10.304	131	476275	20.003	ng/ul	0.00
46) Dimethylphthalate-d6	13.480	166	1091282	21.029	ng/ul	0.00
49) Acenaphthylene-d8	13.745	160	1251547	21.606	ng/ul	0.00
54) 4-Nitrophenol-d4	14.298	143	170577	17.000	ng/ul	0.00
60) Fluorene-d10	15.063	176	936881	21.104	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.198	200	74072	8.450	ng/ul	0.00
73) Anthracene-d10	16.910	188	1425290	21.669	ng/ul	0.00
81) Pyrene-d10	19.227	212	1601170	25.408	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.580	264	1450729	22.764	ng/ul	0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM080624\
Data File : BM047035.D
Acq On : 06 Aug 2024 23:19
Operator : RC/JU
Sample : P3328-15
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Aug 07 01:24:04 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM080224.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Aug 07 01:21:15 2024
Response via : Initial Calibration

