

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM080715\
 Data File : BM002255.D
 Acq On : 08 Aug 2015 01:55
 Operator : TP/UM
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02013

Quant Time: Aug 08 04:42:42 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080715MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Aug 08 04:36:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
3	Naphthalene	1.053	1.079	-2.5	104	0.00
4	2-Methylnaphthalene	0.752	0.775	-3.1	104	0.00
5	1-Methylnaphthalene	0.741	0.762	-2.8	103	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
7 S	Acenaphthylene-d8	2.072	2.123	-2.5	104	0.00
8	Acenaphthylene	2.151	2.208	-2.6	103	0.00
9 C	Acenaphthene	1.447	1.481	-2.3	103	0.00
10 S	Fluorene-d10	1.441	1.484	-3.0	104	0.00
11	Fluorene	1.629	1.683	-3.3	104	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
13	Phenanthrene	1.151	1.192	-3.6	105	0.00
14 S	Anthracene-d10	0.970	0.998	-2.9	104	0.00
15	Anthracene	1.189	1.230	-3.4	105	0.00
16 C	Fluoranthene	1.275	1.345	-5.5	106	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
18 S	Pyrene-d10	0.970	0.984	-1.4	106	0.00
19	Pyrene	1.293	1.319	-2.0	106	0.00
20	Benzo(a)anthracene	1.221	1.254	-2.7	108	0.00
21	Chrysene	1.157	1.189	-2.8	109	0.00
22 I	Perylene-d12	1.000	1.000	0.0	108	0.00
23	Benzo(b)fluoranthene	1.243	1.292	-3.9	111	0.00
24	Benzo(k)fluoranthene	1.194	1.217	-1.9	105	0.00
25 S	Benzo(a)pyrene-d12	1.062	1.093	-2.9	109	0.00
26 C	Benzo(a)pyrene	1.189	1.228	-3.3	109	0.00
27	Indeno(1,2,3-cd)pyrene	1.363	1.417	-4.0	107	0.00
28	Dibenzo(a,h)anthracene	1.155	1.201	-4.0	107	0.00
29	Benzo(g,h,i)perylene	1.150	1.189	-3.4	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0