

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM080715\
 Data File : BM002265.D
 Acq On : 08 Aug 2015 07:57
 Operator : TP/UM
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02014

Quant Time: Aug 09 05:28:31 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080715MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Aug 09 05:27:43 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
3	Naphthalene	1.053	1.078	-2.4	119	0.00
4	2-Methylnaphthalene	0.752	0.769	-2.3	118	0.00
5	1-Methylnaphthalene	0.741	0.756	-2.0	117	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
7 S	Acenaphthylene-d8	2.072	2.124	-2.5	115	0.00
8	Acenaphthylene	2.151	2.221	-3.3	115	0.00
9 C	Acenaphthene	1.447	1.495	-3.3	115	0.00
10 S	Fluorene-d10	1.441	1.469	-1.9	114	0.00
11	Fluorene	1.629	1.668	-2.4	114	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
13	Phenanthrene	1.151	1.189	-3.3	112	0.00
14 S	Anthracene-d10	0.970	0.997	-2.8	111	0.00
15	Anthracene	1.189	1.231	-3.5	112	0.00
16 C	Fluoranthene	1.275	1.309	-2.7	110	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	112	0.00
18 S	Pyrene-d10	0.970	0.987	-1.8	110	0.00
19	Pyrene	1.293	1.326	-2.6	110	0.00
20	Benzo(a)anthracene	1.221	1.249	-2.3	111	0.00
21	Chrysene	1.157	1.181	-2.1	111	0.00
22 I	Perylene-d12	1.000	1.000	0.0	110	0.00
23	Benzo(b)fluoranthene	1.243	1.297	-4.3	115	0.00
24	Benzo(k)fluoranthene	1.194	1.208	-1.2	107	0.00
25 S	Benzo(a)pyrene-d12	1.062	1.090	-2.6	111	0.00
26 C	Benzo(a)pyrene	1.189	1.216	-2.3	110	0.00
27	Indeno(1,2,3-cd)pyrene	1.363	1.344	1.4	104	-0.01
28	Dibenzo(a,h)anthracene	1.155	1.149	0.5	105	0.00
29	Benzo(g,h,i)perylene	1.150	1.091	5.1	99	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0