

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM080715\
 Data File : BM002282.D
 Acq On : 08 Aug 2015 18:14
 Operator : TP/UM
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02015

Quant Time: Aug 09 05:34:36 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080715MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Aug 09 05:31:20 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	154#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	144	0.00
3	Naphthalene	1.053	1.074	-2.0	144	0.00
4	2-Methylnaphthalene	0.752	0.741	1.5	139	0.00
5	1-Methylnaphthalene	0.741	0.727	1.9	137	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	128	0.00
7 S	Acenaphthylene-d8	2.072	2.137	-3.1	128	0.00
8	Acenaphthylene	2.151	2.230	-3.7	128	0.00
9 C	Acenaphthene	1.447	1.494	-3.2	128	0.00
10 S	Fluorene-d10	1.441	1.439	0.1	124	0.00
11	Fluorene	1.629	1.637	-0.5	124	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
13	Phenanthrene	1.151	1.188	-3.2	118	0.00
14 S	Anthracene-d10	0.970	1.000	-3.1	118	0.00
15	Anthracene	1.189	1.223	-2.9	117	0.00
16 C	Fluoranthene	1.275	1.262	1.0	112	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
18 S	Pyrene-d10	0.970	1.132	-16.7	110	0.00
19	Pyrene	1.293	1.519	-17.5	111	0.00
20	Benzo(a)anthracene	1.221	1.265	-3.6	99	0.00
21	Chrysene	1.157	1.178	-1.8	97	0.00
22 I	Perylene-d12	1.000	1.000	0.0	78	0.00
23	Benzo(b)fluoranthene	1.243	1.345	-8.2	83	0.00
24	Benzo(k)fluoranthene	1.194	1.281	-7.3	80	0.00
25 S	Benzo(a)pyrene-d12	1.062	1.101	-3.7	79	0.00
26 C	Benzo(a)pyrene	1.189	1.223	-2.9	78	0.00
27	Indeno(1,2,3-cd)pyrene	1.363	1.300	4.6	71	0.00
28	Dibenzo(a,h)anthracene	1.155	1.113	3.6	71	0.00
29	Benzo(g,h,i)perylene	1.150	1.086	5.6	70	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0