

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM080715\
 Data File : BM002302.D
 Acq On : 09 Aug 2015 06:58
 Operator : TP/UM
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02016

Quant Time: Aug 09 07:29:45 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080715MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Aug 09 06:16:29 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	157#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	146	0.00
3	Naphthalene	1.053	1.078	-2.4	147	0.00
4	2-Methylnaphthalene	0.752	0.746	0.8	142	0.00
5	1-Methylnaphthalene	0.741	0.726	2.0	139	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	130	0.00
7 S	Acenaphthylene-d8	2.072	2.138	-3.2	131	0.00
8	Acenaphthylene	2.151	2.226	-3.5	131	0.00
9 C	Acenaphthene	1.447	1.491	-3.0	130	0.00
10 S	Fluorene-d10	1.441	1.423	1.2	125	0.00
11	Fluorene	1.629	1.618	0.7	125	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	119	0.00
13	Phenanthrene	1.151	1.190	-3.4	119	0.00
14 S	Anthracene-d10	0.970	0.997	-2.8	118	0.00
15	Anthracene	1.189	1.226	-3.1	118	0.00
16 C	Fluoranthene	1.275	1.205	5.5	108	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	89	0.00
18 S	Pyrene-d10	0.970	1.178	-21.4#	105	0.00
19	Pyrene	1.293	1.571	-21.5#	104	0.00
20	Benzo(a)anthracene	1.221	1.260	-3.2	90	0.00
21	Chrysene	1.157	1.168	-1.0	88	0.00
22 I	Perylene-d12	1.000	1.000	0.0	76	0.00
23	Benzo(b)fluoranthene	1.243	1.302	-4.7	79	0.00
24	Benzo(k)fluoranthene	1.194	1.250	-4.7	76	0.00
25 S	Benzo(a)pyrene-d12	1.062	1.089	-2.5	76	0.00
26 C	Benzo(a)pyrene	1.189	1.205	-1.3	75	0.00
27	Indeno(1,2,3-cd)pyrene	1.363	1.355	0.6	72	0.00
28	Dibenzo(a,h)anthracene	1.155	1.148	0.6	72	0.00
29	Benzo(g,h,i)perylene	1.150	1.132	1.6	71	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0