


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM080715\
 Data File : BM002248.D
 Acq On : 07 Aug 2015 21:41
 Operator : TP/UM
 Sample : DFTPP91
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP91

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080715MA.M
 Title : SVOA CALIBRATION
 Last Update : Mon Aug 10 05:35:03 2015



AutoFind: Scans 2631, 2632, 2633; Background Corrected with Scan 2626

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.9	56475	PASS
68	69	0.00	2	1.8	938	PASS
69	198	0.00	100	42.6	52339	PASS
70	69	0.00	2	0.6	306	PASS
127	198	10	80	54.1	66511	PASS
197	198	0.00	2	0.7	823	PASS
198	198	100	100	100.0	122970	PASS
199	198	5	9	6.9	8464	PASS
275	198	10	60	27.1	33269	PASS
365	198	1	100	3.2	3988	PASS
441	443	0.01	100	77.0	15633	PASS
442	198	50	100	83.8	103090	PASS
443	442	15	24	19.7	20307	PASS

m/z	Abundance
36.00	595.0
36.90	356.0
38.00	630.0
39.00	2713.0
40.00	2688.0
41.10	7376.0
42.10	2417.0
43.10	10123.0
44.00	24864.0
45.10	2901.0
46.00	337.0
47.00	316.0
49.00	1260.0
50.00	536.0
51.00	965.0
52.10	485.0
53.00	1205.0
54.00	1111.0
55.10	7049.0
56.00	3331.0
57.10	8754.0
58.00	2934.0
59.10	2994.0
60.00	1363.0
61.00	413.0
65.00	667.0
66.10	363.0
67.00	2399.0
68.10	1506.0
69.10	3511.0
70.10	1836.0
71.10	4422.0
73.10	99832.0
74.10	8234.0
75.10	4030.0
76.00	379.0
77.00	1079.0
78.10	618.0
79.00	1579.0
80.10	648.0
81.00	2739.0
82.00	1902.0
83.10	3328.0
84.00	2169.0
85.10	3572.0
85.90	1036.0
87.00	882.0
91.00	1935.0
92.00	459.0
93.00	1276.0
94.10	582.0
95.10	2411.0
96.10	2386.0
97.10	2860.0
98.00	2443.0
99.00	806.0
100.00	451.0
102.90	550.0
104.00	405.0
105.00	1353.0
106.00	365.0
107.10	1180.0
108.10	401.0
109.00	1438.0
110.10	853.0
111.10	1372.0
112.10	670.0
113.10	529.0
115.00	661.0
117.10	1412.0
118.10	408.0
119.00	1434.0
120.00	340.0
121.00	1131.0
122.00	450.0
123.10	839.0
124.10	476.0
125.10	575.0
126.10	461.0
127.00	595.0
127.90	412.0
129.00	1134.0
131.00	2102.0
132.10	627.0
133.00	3289.0
134.00	783.0
135.00	7900.0
136.00	1145.0
137.00	765.0
141.10	378.0
143.10	439.0

Instrument :
BNA_M
ClientSampleId :
DFTPP91

145.00	1797.0
146.00	821.0
147.10	40336.0
148.00	7315.0
149.10	4434.0
150.00	692.0
151.10	307.0
152.10	450.0
153.10	342.0
154.90	404.0
156.90	313.0
159.10	810.0
161.10	742.0
163.10	741.0
165.00	635.0
169.00	325.0
175.00	803.0
175.90	438.0
177.00	1399.0
177.90	423.0
179.00	1312.0
180.10	399.0
180.90	357.0
189.10	613.0
191.00	5429.0
192.00	3233.0
193.00	28496.0
194.10	5633.0
195.00	3341.0
196.00	563.0
197.00	361.0
203.00	390.0
205.00	1906.0
206.00	664.0
207.10	54992.0
208.10	11692.0
209.00	6814.0
210.10	1093.0
211.00	400.0
219.10	763.0
220.20	562.0
221.10	27384.0
222.00	6645.0
223.10	4093.0
224.10	678.0
225.00	417.0
233.10	723.0
235.00	444.0
237.00	466.0
239.20	1223.0
240.20	339.0
249.00	1034.0
251.00	2723.0
252.10	846.0
253.00	3394.0
254.00	1135.0
255.10	1402.0
263.20	342.0
265.10	3036.0
266.10	983.0
267.10	9365.0
268.10	2199.0
269.00	1671.0
270.00	466.0
279.10	910.0
280.10	431.0
281.10	20824.0
282.10	5938.0
283.00	4824.0
284.00	1157.0
285.10	376.0
292.90	503.0
295.10	4767.0
296.10	1524.0
297.00	1016.0
307.00	385.0
309.10	465.0
311.00	722.0
313.10	1532.0
314.00	439.0
323.00	394.0
325.00	1930.0
326.00	760.0
327.00	5728.0
328.00	2197.0
329.00	1084.0
339.00	884.0
340.00	357.0
341.10	10293.0
342.10	3404.0
343.10	2308.0
344.00	415.0
353.10	375.0
355.10	13800.0
356.10	4733.0

Instrument :
BNA_M
ClientSampleId :
DFTPP91

357.10	3224.0
358.10	817.0
359.00	353.0
367.00	443.0
369.10	1330.0
370.10	547.0
371.10	746.0
373.10	377.0
383.00	301.0
385.00	955.0
386.00	349.0
387.10	2009.0
388.10	711.0
389.10	393.0
399.00	890.0
400.00	396.0
401.10	3751.0
402.10	1382.0
403.10	904.0
415.10	3521.0
416.10	1592.0
417.00	995.0
429.10	4236.0
430.10	1666.0
431.10	1215.0
432.10	399.0
445.10	360.0
447.00	621.0
459.00	533.0
461.10	2092.0
462.10	866.0
463.10	731.0
475.10	3577.0
476.20	1776.0
477.00	864.0
478.10	390.0
489.10	1850.0
490.10	950.0
491.10	599.0
503.10	863.0
504.00	474.0
505.00	496.0
521.00	922.0
522.10	445.0
535.10	1924.0
536.10	1134.0
537.10	743.0
538.00	391.0
549.20	1755.0

Instrument :
BNA_M
ClientSampleId :
DFTPP91

m/z	Abundance
36.00	877.0
37.00	415.0
38.10	1076.0
39.10	6238.0
40.00	2332.0
41.10	7458.0
42.10	2469.0
43.10	10436.0
44.00	23416.0
45.10	2821.0
46.00	345.0
47.00	504.0
49.00	1730.0
50.00	11627.0
51.10	45608.0
52.10	2975.0
53.00	1394.0
54.00	1223.0
55.10	7688.0
56.00	4374.0
57.10	12244.0
58.10	2783.0
59.10	2994.0
60.00	1194.0
61.00	1034.0
62.00	702.0
63.00	1945.0
64.00	568.0
65.00	1455.0
66.00	484.0
67.00	2535.0
68.10	2303.0
69.00	44672.0
70.10	2071.0
71.10	5360.0
73.10	99456.0
74.10	12193.0
75.00	10458.0
76.10	2543.0
77.10	47744.0
78.10	3791.0
79.00	4024.0
80.00	2763.0
81.00	5767.0
82.10	2613.0
83.10	3891.0
84.00	2176.0
85.10	3991.0
86.00	1582.0
87.00	1239.0
87.90	407.0
91.10	2663.0
92.00	1335.0
93.00	5957.0
94.00	1054.0
95.10	2464.0
96.00	2650.0
97.10	3087.0
98.00	6233.0
99.00	3988.0
100.00	684.0
101.00	2182.0
103.00	1082.0
104.00	1551.0
105.00	2477.0
106.10	645.0
107.00	14338.0
108.00	2401.0
109.10	1736.0
110.00	27072.0
111.00	4874.0
112.00	1249.0
113.00	735.0
115.00	719.0
116.00	1073.0
117.00	11491.0
118.00	1112.0
119.00	1493.0
120.00	534.0
121.10	1360.0
122.00	1360.0
123.00	1978.0
124.00	1061.0
125.00	1283.0
127.00	49672.0
128.00	4506.0
129.00	19656.0
129.90	1907.0
131.00	2305.0
132.00	598.0
133.00	3421.0

Instrument :
BNA_M
ClientSampleId :
DFTPP91

134.00	1378.0
135.00	9302.0
136.00	1830.0
137.00	1623.0
138.10	530.0
138.90	384.0
140.00	386.0
141.00	2486.0
142.00	913.0
142.90	810.0
145.00	1915.0
146.00	1074.0
147.10	42104.0
148.00	10166.0
149.00	5260.0
150.00	784.0
151.00	639.0
152.00	475.0
153.00	888.0
153.90	577.0
155.00	1528.0
156.00	2021.0
157.10	511.0
157.90	469.0
159.10	1245.0
160.00	1170.0
161.00	1659.0
162.00	597.0
163.00	973.0
164.00	311.0
165.00	1387.0
166.00	912.0
167.00	4295.0
167.90	2030.0
169.00	735.0
170.00	397.0
171.10	368.0
171.90	456.0
172.90	776.0
174.00	1023.0
175.00	2953.0
176.00	873.0
177.00	2056.0
178.00	674.0
179.00	4365.0
179.90	2509.0
181.00	1273.0
183.10	475.0
183.90	346.0
185.00	1799.0
186.00	11949.0
187.00	3742.0
188.00	452.0
188.90	1447.0
190.00	440.0
191.00	6034.0
192.00	4058.0
193.00	30376.0
194.00	6248.0
195.00	3101.0
196.00	3268.0
197.10	1005.0
197.90	83920.0
198.90	6106.0
200.00	603.0
201.50	580.0
203.00	936.0
204.00	3035.0
205.00	6913.0
206.00	22488.0
207.10	58824.0
208.10	12676.0
209.10	7086.0
210.10	1255.0
211.00	1294.0
215.00	462.0
216.00	430.0
217.00	5456.0
218.00	946.0
219.00	945.0
221.10	31016.0
222.10	7115.0
223.00	4806.0
224.00	11909.0
225.00	3378.0
226.00	373.0
227.00	4898.0
227.90	623.0
229.00	945.0
231.00	493.0
233.00	888.0
234.00	554.0
235.00	758.0
236.10	359.0
236.90	874.0

Instrument :
BNA_M
ClientSampleId :
DFTPP91

239.20	1346.0
240.20	337.0
241.00	548.0
242.00	509.0
243.10	622.0
244.00	8414.0
244.90	1219.0
246.00	1936.0
247.00	583.0
249.00	1488.0
250.00	492.0
251.00	3016.0
252.00	827.0
253.00	3441.0
254.10	1337.0
255.00	46864.0
256.00	6885.0
257.00	680.0
258.00	2763.0
258.90	424.0
263.00	542.0
264.10	316.0
265.00	3833.0
266.00	1115.0
267.10	9036.0
268.00	2395.0
269.00	2096.0
270.10	495.0
271.00	308.0
273.00	1206.0
274.00	3575.0
275.00	20992.0
276.00	2927.0
277.00	1572.0
279.00	1225.0
280.20	495.0
281.10	20872.0
282.10	6358.0
283.10	5347.0
284.10	1328.0
285.00	571.0
293.00	892.0
294.10	418.0
295.10	4623.0
296.00	6776.0
297.00	1796.0
298.00	399.0
302.90	717.0
307.10	370.0
309.00	447.0
311.00	800.0
313.00	1723.0
314.00	893.0
314.90	834.0
316.10	408.0
323.00	2119.0
324.00	496.0
325.00	2260.0
326.00	949.0
327.00	5981.0
328.00	1945.0
329.00	1166.0
334.00	1112.0
335.10	313.0
339.00	880.0
340.10	472.0
341.10	11216.0
342.10	3680.0
343.10	2031.0
344.00	506.0
346.10	414.0
352.00	525.0
353.00	735.0
354.00	804.0
355.10	13549.0
356.10	4903.0
357.10	3319.0
358.00	954.0
359.10	436.0
365.00	2398.0
367.00	367.0
369.10	1490.0
370.10	618.0
371.10	717.0
372.00	1045.0
373.00	442.0
383.00	490.0
385.00	913.0
386.00	398.0
387.10	2043.0
388.00	788.0
389.00	433.0
399.00	710.0
399.90	334.0
401.10	3512.0

Instrument :
BNA_M
ClientSampleId :
DFTPP91

402.00	1925.0
403.00	1520.0
404.00	481.0
415.10	3503.0
416.00	1496.0
417.10	1052.0
418.10	436.0
421.00	478.0
422.00	415.0
423.00	2964.0
424.00	661.0
429.10	4293.0
430.10	1957.0
431.10	1354.0
432.10	384.0
441.00	7820.0
442.00	51352.0
443.00	10585.0
444.00	1141.0
445.00	376.0
447.10	612.0
459.10	596.0
461.10	2077.0
462.00	870.0
462.90	457.0
475.10	3685.0
476.10	1728.0
477.10	1116.0
478.00	365.0
489.10	1824.0
490.10	1082.0
491.10	760.0
503.10	840.0
504.00	555.0
505.00	519.0
521.10	879.0
522.00	442.0
535.10	2440.0
536.10	1106.0
537.10	876.0
549.20	2061.0

Instrument :
BNA_M
ClientSampleId :
DFTPP91

m/z	Abundance
36.00	775.0
37.00	502.0
38.00	1360.0
39.10	7715.0
40.00	2616.0
41.10	7791.0
42.10	2691.0
43.10	9809.0
44.00	23472.0
45.00	2984.0
46.00	368.0
46.90	403.0
49.00	1777.0
50.00	16664.0
51.10	69024.0
52.00	3523.0
53.00	1499.0
54.00	1155.0
55.10	8119.0
56.10	4707.0
57.10	12539.0
58.00	2853.0
59.10	3200.0
60.00	1340.0
61.00	1188.0
62.00	1008.0
63.00	2806.0
64.00	581.0
65.10	1740.0
66.00	485.0
67.00	2265.0
68.10	2652.0
69.00	65904.0
70.10	2054.0
71.10	4706.0
73.10	99840.0
74.10	14054.0
75.00	13569.0
76.10	3561.0
77.10	69832.0
78.10	5479.0
79.00	5878.0
80.00	4373.0
81.00	7558.0
82.00	3005.0
83.10	4338.0
84.00	2422.0
85.10	4101.0
86.00	2414.0
87.00	1501.0
87.90	525.0
89.00	489.0
91.10	2866.0
92.00	1357.0
93.00	8682.0
94.00	1072.0
95.00	2344.0
96.10	2593.0
97.10	2820.0
98.00	8362.0
99.00	5433.0
100.00	753.0
101.00	3100.0
102.00	349.0
102.90	1459.0
104.00	1927.0
105.00	2732.0
106.10	805.0
107.00	22592.0
108.00	4030.0
109.10	2055.0
110.00	41000.0
111.00	6958.0
112.00	1517.0
113.00	837.0
114.10	336.0
115.00	826.0
116.00	1508.0
117.00	17248.0
118.00	1865.0
119.00	1523.0
120.10	679.0
121.00	1317.0
122.00	1929.0
123.00	2741.0
124.00	1381.0
125.10	1341.0
126.10	1012.0
127.00	79800.0
128.00	6268.0
129.00	30944.0

Instrument :
BNA_M
ClientSampleId :
DFTPP91

130.00	2711.0
131.00	2544.0
132.00	958.0
133.00	3827.0
134.00	1524.0
135.00	9889.0
136.00	2125.0
137.00	1906.0
138.10	512.0
139.10	458.0
140.00	613.0
141.00	3965.0
142.00	1514.0
143.00	1220.0
144.00	454.0
145.10	2058.0
146.00	1412.0
147.10	43176.0
148.00	11345.0
149.00	5370.0
150.00	845.0
151.00	1020.0
152.00	682.0
153.00	1372.0
154.00	982.0
155.00	2126.0
156.00	3236.0
156.90	1123.0
157.90	851.0
159.10	1550.0
160.00	1344.0
161.00	2218.0
162.00	744.0
163.00	1001.0
164.10	384.0
165.00	1911.0
166.00	1497.0
167.00	7266.0
168.00	3322.0
169.10	791.0
170.00	431.0
170.90	544.0
172.00	699.0
173.00	1187.0
174.00	1764.0
175.00	4031.0
176.10	1041.0
177.00	2504.0
178.00	976.0
179.00	6063.0
180.00	4173.0
181.00	1974.0
182.00	410.0
183.10	398.0
183.90	463.0
185.00	2658.0
186.00	20136.0
187.00	5916.0
187.90	538.0
188.90	1538.0
190.00	401.0
191.00	5666.0
192.00	4973.0
193.00	31248.0
194.00	6337.0
195.00	3097.0
196.00	4880.0
197.10	1408.0
197.90	143680.0
199.00	10118.0
199.90	1000.0
201.30	861.0
202.00	413.0
203.00	1478.0
204.00	5276.0
205.00	10432.0
206.00	38872.0
207.10	62880.0
208.10	12771.0
209.10	6807.0
210.10	1639.0
211.00	1836.0
215.00	551.0
216.00	867.0
217.00	8784.0
218.00	1184.0
219.00	1037.0
220.20	758.0
221.10	34144.0
222.00	7707.0
223.10	5891.0
224.00	20848.0
225.00	5859.0
226.00	771.0
227.00	8033.0

Instrument :
BNA_M
ClientSampleId :
DFTPP91

228.00	1015.0
229.00	1995.0
230.00	443.0
231.00	897.0
233.00	813.0
234.00	722.0
235.00	945.0
236.00	631.0
237.00	1062.0
239.10	1651.0
240.00	421.0
240.90	611.0
241.90	1004.0
243.00	1085.0
244.00	14667.0
245.00	2169.0
246.00	3165.0
247.00	867.0
248.90	1763.0
250.00	465.0
251.00	2881.0
252.00	987.0
253.00	3546.0
254.10	1585.0
255.00	83240.0
256.00	12361.0
257.00	1348.0
258.00	4544.0
258.90	745.0
263.10	455.0
264.10	323.0
265.10	4802.0
266.10	1173.0
267.10	9492.0
268.00	2390.0
269.10	1942.0
269.90	475.0
271.00	339.0
273.00	2609.0
274.00	6625.0
275.00	38624.0
276.00	5411.0
277.00	3538.0
277.90	620.0
279.00	1179.0
280.10	555.0
281.00	20000.0
282.10	5707.0
283.10	5076.0
284.00	1362.0
285.10	745.0
292.00	300.0
293.00	1198.0
294.00	456.0
295.10	5295.0
296.00	11151.0
297.00	2256.0
298.10	318.0
303.00	1216.0
303.90	383.0
307.10	303.0
308.00	313.0
309.10	520.0
311.00	824.0
312.00	337.0
313.10	1777.0
314.00	922.0
315.00	1318.0
316.00	624.0
321.00	424.0
323.10	3633.0
324.00	883.0
325.00	2015.0
326.00	817.0
327.00	6346.0
328.00	2272.0
329.00	1158.0
330.00	406.0
333.10	335.0
334.00	2190.0
335.00	665.0
339.00	894.0
340.10	412.0
341.10	10756.0
342.10	3822.0
343.00	2252.0
344.00	470.0
345.90	651.0
352.00	1007.0
353.00	1067.0
354.10	1508.0
355.10	13666.0
356.10	4651.0
357.10	3117.0
358.10	839.0

Instrument :
BNA_M
ClientSampleId :
DFTPP91

359.00	399.0
365.00	4398.0
366.00	739.0
367.00	385.0
369.00	1293.0
370.10	598.0
371.10	842.0
372.00	2087.0
373.10	665.0
383.00	895.0
385.00	786.0
386.00	362.0
387.00	1993.0
388.10	896.0
389.00	477.0
389.90	411.0
399.00	805.0
400.00	360.0
401.10	3477.0
402.00	2005.0
403.00	2138.0
404.00	752.0
415.10	3421.0
416.10	1301.0
417.10	936.0
418.10	453.0
421.00	798.0
422.00	992.0
423.00	6269.0
424.00	1342.0
429.10	3648.0
430.10	1837.0
431.10	1355.0
432.00	501.0
441.00	17072.0
442.00	114048.0
443.00	22728.0
444.00	2045.0
445.00	458.0
447.10	636.0
459.00	564.0
460.00	329.0
461.10	2254.0
462.20	768.0
463.10	618.0
475.10	3147.0
476.10	1668.0
477.10	1159.0
478.10	327.0
489.10	1779.0
490.10	872.0
491.20	685.0
503.10	1132.0
504.10	455.0
505.20	559.0
519.10	306.0
521.10	899.0
522.10	398.0
535.10	2224.0
536.10	1091.0
537.10	662.0
549.20	1978.0

Instrument :
BNA_M
ClientSampleId :
DFTPP91

m/z	Abundance
36.00	713.0
37.00	597.0
38.00	1190.0
39.10	6945.0
40.00	2378.0
41.10	7665.0
42.10	2548.0
43.10	9968.0
44.00	22448.0
45.00	2882.0
46.00	312.0
47.00	466.0
49.00	1717.0
50.10	14302.0
51.10	57688.0
52.00	3130.0
53.00	1329.0
54.00	1190.0
55.10	8327.0
56.10	4563.0
57.10	12571.0
58.00	2731.0
59.10	3103.0
60.00	1320.0
61.00	1047.0
62.00	791.0
63.10	2481.0
64.00	567.0
65.10	1617.0
66.00	615.0
67.10	2530.0
68.10	2378.0
69.00	56976.0
70.10	2303.0
71.10	4879.0
73.10	102352.0
74.10	14454.0
75.00	12267.0
76.10	3163.0
77.10	61592.0
78.10	4719.0
79.00	5402.0
80.00	3741.0
81.00	6700.0
82.00	3186.0
83.10	3884.0
84.00	2426.0
85.10	4191.0
86.00	2320.0
87.00	1573.0
88.00	469.0
89.00	363.0
91.00	2703.0
92.00	1465.0
93.00	7588.0
94.00	1043.0
95.10	2395.0
96.10	2546.0
97.10	2611.0
98.00	7628.0
99.00	5057.0
100.00	745.0
101.00	2443.0
103.00	1261.0
103.90	1920.0
105.10	2942.0
106.00	1022.0
107.00	19616.0
108.00	3238.0
109.10	1971.0
110.00	36912.0
111.00	6410.0
112.00	1290.0
113.00	717.0
114.00	358.0
114.90	934.0
116.00	1410.0
117.00	15895.0
118.00	1373.0
119.00	1318.0
120.00	671.0
121.10	1153.0
122.00	1776.0
123.00	2712.0
124.00	1239.0
125.10	1303.0
126.20	1289.0
127.00	71848.0
128.00	5837.0
129.00	28056.0
130.00	2480.0

Instrument :
BNA_M
ClientSampleId :
DFTPP91

131.00	2321.0
132.00	633.0
133.00	3595.0
134.00	1727.0
135.00	9183.0
136.00	2307.0
137.00	1670.0
138.00	478.0
139.00	369.0
140.00	655.0
141.00	3527.0
142.00	1348.0
143.00	1103.0
144.00	399.0
145.10	2089.0
146.10	1301.0
147.10	43176.0
148.00	11404.0
149.10	5177.0
150.00	889.0
151.00	696.0
152.00	611.0
153.00	1353.0
154.00	727.0
155.00	2237.0
156.00	3355.0
157.10	943.0
158.00	943.0
159.00	1479.0
160.00	1383.0
161.00	2388.0
161.90	677.0
163.10	993.0
164.10	389.0
165.00	1929.0
166.00	1404.0
167.00	7260.0
168.00	3368.0
169.00	961.0
170.10	335.0
170.90	663.0
172.00	632.0
173.00	1124.0
174.00	1640.0
175.00	3520.0
176.00	1123.0
177.00	2157.0
178.00	747.0
179.00	6498.0
180.00	3497.0
180.90	2048.0
182.00	494.0
183.00	358.0
183.90	562.0
185.00	2706.0
186.00	19848.0
187.00	5540.0
188.00	702.0
189.00	1847.0
190.10	519.0
191.00	5757.0
192.00	4903.0
193.00	30400.0
194.00	6731.0
195.00	3190.0
196.00	4923.0
197.10	1141.0
197.90	141312.0
199.00	9169.0
200.00	911.0
201.30	779.0
202.10	349.0
203.00	1285.0
204.00	5036.0
205.00	10602.0
206.00	38520.0
207.10	60728.0
208.10	13558.0
209.10	6949.0
210.10	1408.0
211.10	1661.0
212.00	419.0
214.90	666.0
216.00	778.0
217.00	9237.0
218.00	1290.0
219.00	914.0
220.10	712.0
221.10	33040.0
222.00	7673.0
223.00	5555.0
224.00	20024.0
225.00	5294.0
226.00	664.0
227.00	8383.0

Instrument :
BNA_M
ClientSampleId :
DFTPP91

227.90	1275.0
229.00	1696.0
229.90	345.0
231.00	908.0
233.00	966.0
234.00	775.0
234.90	822.0
236.00	645.0
237.00	1084.0
239.10	1727.0
240.10	526.0
241.00	636.0
242.00	1093.0
243.10	1285.0
244.00	15566.0
245.00	2302.0
246.00	3028.0
247.00	1049.0
249.00	1696.0
250.00	471.0
251.00	3442.0
252.00	1125.0
253.00	3541.0
254.10	1741.0
255.00	84080.0
256.00	12336.0
257.00	1042.0
258.00	4736.0
258.90	691.0
263.10	333.0
264.10	394.0
265.00	4881.0
266.00	1135.0
267.10	9496.0
268.10	2581.0
269.10	1933.0
270.00	467.0
273.00	2671.0
274.00	6835.0
275.00	40192.0
276.00	5061.0
277.00	3278.0
278.00	600.0
279.00	1004.0
280.00	422.0
281.10	20512.0
282.00	5922.0
283.10	5474.0
284.10	1350.0
285.00	881.0
293.00	1146.0
294.00	369.0
295.10	5269.0
296.00	11454.0
297.00	2390.0
298.10	375.0
303.00	1150.0
304.00	416.0
308.10	317.0
309.00	497.0
309.90	307.0
311.00	777.0
312.00	309.0
313.00	1663.0
314.10	894.0
315.00	1499.0
315.90	806.0
321.00	457.0
323.10	4145.0
323.90	660.0
325.00	2306.0
326.00	825.0
327.00	6727.0
328.00	2243.0
329.10	1391.0
330.00	397.0
333.00	475.0
334.00	2438.0
335.00	770.0
339.10	640.0
341.10	10799.0
342.10	3711.0
343.10	2214.0
344.00	575.0
346.00	776.0
352.00	1170.0
353.10	1053.0
354.00	1451.0
355.10	13462.0
356.10	5002.0
357.10	3123.0
358.00	986.0
359.00	447.0
365.00	5170.0
366.00	846.0

Instrument :
BNA_M
ClientSampleId :
DFTPP91

367.00	462.0
369.10	1483.0
370.00	583.0
371.10	1166.0
372.00	2487.0
373.00	786.0
383.00	913.0
385.00	915.0
386.00	330.0
387.00	1681.0
388.00	741.0
389.00	517.0
390.10	342.0
390.90	301.0
399.10	751.0
400.00	384.0
401.10	3435.0
402.00	2091.0
403.00	2346.0
404.00	773.0
415.10	3644.0
416.10	1453.0
417.10	1001.0
418.00	303.0
421.00	1093.0
422.00	1132.0
423.00	7907.0
424.00	1578.0
429.10	3928.0
430.00	1690.0
431.00	1347.0
432.00	364.0
441.00	22008.0
442.00	143872.0
443.00	27608.0
444.00	2377.0
445.00	459.0
447.00	808.0
459.00	496.0
461.10	1998.0
462.10	849.0
463.00	560.0
475.10	3350.0
476.10	1677.0
477.10	1095.0
478.10	351.0
489.10	1814.0
490.10	841.0
491.10	535.0
503.20	1144.0
504.20	488.0
505.00	422.0
521.10	792.0
522.10	447.0
523.10	358.0
535.10	2253.0
536.10	1239.0
537.10	958.0
538.10	352.0
549.10	1865.0

Instrument :
BNA_M
ClientSampleId :
DFTPP91