


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM080720\
 Data File : BM027009.D
 Acq On : 06 Aug 2020 18:23
 Operator : JU/CG
 Sample : DFTPP20
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP20

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM080720MA.M
 Title : SVOA CALIBRATION
 Last Update : Fri Aug 07 03:56:08 2020



AutoFind: Scans 2465, 2466, 2467; Background Corrected with Scan 2460

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.5	42872	PASS
68	69	0.00	2	1.7	1196	PASS
69	198	0.00	100	49.7	69986	PASS
70	69	0.00	2	0.5	342	PASS
127	198	10	80	45.8	64450	PASS
197	198	0.00	2	0.6	862	PASS
198	198	100	100	100.0	140794	PASS
199	198	5	9	6.5	9199	PASS
275	198	10	60	30.0	42186	PASS
365	198	1	100	5.0	7046	PASS
441	198	0.01	100	12.6	17714	PASS
442	198	50	100	84.0	118248	PASS
443	442	15	24	19.8	23467	PASS

Scan 2460 (17.457 min): BM027009.D
DFTPP20

m/z	Abundance
39.90	463.0
44.00	716.0
48.90	154.0
73.00	401.0
83.80	195.0
86.00	170.0
190.90	154.0
207.10	1061.0
208.00	219.0
209.00	169.0
281.00	327.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

m/z	Abundance
36.90	258.0
38.00	683.0
39.10	3709.0
40.10	695.0
41.10	150.0
43.90	610.0
49.00	601.0
50.10	11667.0
51.10	43256.0
52.00	1999.0
55.00	312.0
56.00	1674.0
57.00	3313.0
57.80	176.0
61.00	859.0
62.00	784.0
63.00	2227.0
64.10	424.0
65.00	933.0
68.00	1241.0
69.00	68032.0
69.90	378.0
72.90	761.0
74.00	7573.0
75.00	11224.0
76.10	3167.0
77.10	67536.0
78.00	4639.0
79.00	5643.0
80.00	4316.0
81.00	5268.0
82.00	1246.0
83.00	1154.0
84.00	275.0
84.90	729.0
86.00	1482.0
87.10	954.0
88.00	272.0
91.00	1131.0
92.00	1074.0
93.00	9264.0
93.90	562.0
94.80	210.0
96.00	494.0
98.00	6942.0
99.00	6525.0
99.90	503.0
101.00	2607.0
102.10	169.0
103.00	926.0
104.00	1821.0
105.00	2089.0
106.00	446.0
107.00	19528.0
108.00	2796.0
109.00	428.0
110.00	32328.0
111.00	5787.0
112.00	732.0
113.00	269.0
116.00	1124.0
117.00	19680.0
118.00	1535.0
118.90	152.0
120.10	242.0
121.90	1004.0
123.10	1960.0
123.90	1119.0
124.90	1018.0
126.20	334.0
127.00	57856.0
128.00	4848.0
129.00	25512.0
130.00	1878.0
130.90	506.0
132.00	187.0
133.90	803.0
135.00	1797.0
135.90	969.0
137.00	1352.0
138.00	206.0
139.10	153.0
139.90	381.0
141.00	3960.0
142.00	1280.0
143.00	846.0
144.00	221.0
144.80	321.0
146.00	784.0
147.10	1995.0
148.00	5066.0
148.90	1136.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

150.00	220.0
151.00	358.0
151.90	312.0
152.90	1346.0
153.90	967.0
155.00	1783.0
156.00	2926.0
157.00	572.0
158.10	704.0
158.90	652.0
160.00	1126.0
161.00	1743.0
161.90	615.0
163.00	261.0
164.10	206.0
165.00	1421.0
166.00	1398.0
167.00	6869.0
168.00	4643.0
168.90	728.0
170.00	253.0
171.00	528.0
171.90	869.0
173.00	905.0
174.00	1497.0
175.00	3049.0
175.90	696.0
176.90	1472.0
178.00	452.0
179.00	5651.0
180.00	4216.0
181.00	1661.0
182.10	225.0
182.80	177.0
184.00	373.0
185.00	2794.0
186.00	18232.0
187.00	5401.0
187.90	502.0
188.90	1122.0
189.90	349.0
190.90	926.0
192.00	1959.0
193.00	2297.0
194.10	419.0
194.90	176.0
196.00	3647.0
197.00	808.0
198.00	118160.0
198.90	7436.0
200.00	850.0
201.70	385.0
202.90	1014.0
204.00	4555.0
205.00	8522.0
206.00	33544.0
207.00	5790.0
208.00	1581.0
209.00	751.0
210.10	636.0
210.90	1312.0
212.10	325.0
214.90	440.0
216.00	680.0
217.00	10491.0
217.90	968.0
219.00	264.0
220.00	181.0
221.00	3717.0
222.00	1302.0
222.90	1981.0
224.00	19016.0
225.00	4893.0
226.00	570.0
227.00	9380.0
228.00	1111.0
229.00	1643.0
229.90	221.0
231.10	655.0
232.90	161.0
233.90	543.0
234.90	516.0
235.90	660.0
237.00	843.0
239.00	321.0
239.90	334.0
241.00	411.0
242.00	1032.0
243.00	1101.0
244.10	13244.0
245.00	1893.0
246.00	3096.0
247.00	434.0
248.00	181.0
248.90	454.0
252.90	389.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

254.00	568.0
255.00	75824.0
256.00	11035.0
257.00	1078.0
258.00	5081.0
259.00	947.0
265.00	2036.0
266.00	422.0
271.10	317.0
272.10	301.0
272.90	2191.0
274.00	4790.0
275.00	31096.0
276.00	4386.0
277.00	3115.0
277.90	483.0
281.10	425.0
282.00	210.0
283.00	440.0
284.10	218.0
284.90	503.0
290.10	271.0
291.90	168.0
293.00	764.0
293.80	171.0
295.00	248.0
296.00	11154.0
297.00	1586.0
301.00	194.0
301.90	165.0
303.00	1356.0
304.10	311.0
308.00	169.0
314.00	545.0
314.90	1284.0
316.00	680.0
320.80	361.0
323.10	3243.0
324.00	792.0
327.00	572.0
328.00	243.0
332.00	337.0
333.10	327.0
334.10	1796.0
335.00	529.0
340.80	327.0
346.00	666.0
352.10	962.0
353.10	489.0
353.90	891.0
355.10	200.0
365.00	5131.0
366.00	553.0
370.90	238.0
372.00	1442.0
373.00	228.0
383.00	452.0
383.80	157.0
390.10	178.0
391.10	163.0
402.00	508.0
403.00	942.0
403.90	318.0
421.00	632.0
422.00	520.0
423.10	4833.0
424.00	1305.0
425.00	169.0
441.00	10526.0
442.10	70840.0
443.10	14161.0
444.00	1519.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

m/z	Abundance
37.00	344.0
38.00	596.0
39.10	4151.0
40.00	653.0
44.00	639.0
49.00	555.0
50.00	14097.0
51.00	52608.0
52.00	2690.0
54.90	262.0
56.00	1721.0
57.00	4012.0
57.90	244.0
61.00	1020.0
62.00	1197.0
63.00	3014.0
63.90	417.0
65.00	1245.0
68.00	1469.0
69.00	85496.0
69.90	446.0
73.10	1157.0
74.00	9559.0
75.00	13910.0
76.00	4282.0
77.10	84320.0
78.00	6428.0
79.00	6000.0
80.00	5396.0
81.00	6946.0
82.00	1403.0
83.00	1360.0
83.90	379.0
85.00	1109.0
86.00	2083.0
87.00	977.0
87.90	293.0
90.90	1664.0
92.00	1497.0
93.00	11285.0
94.00	1044.0
95.00	244.0
96.00	494.0
96.90	185.0
98.00	9152.0
99.10	6579.0
100.00	575.0
101.00	3589.0
101.90	242.0
102.90	1115.0
103.90	2132.0
105.00	2246.0
106.20	632.0
107.00	26424.0
108.00	3964.0
109.10	680.0
110.00	42608.0
111.00	6787.0
112.00	840.0
112.90	270.0
115.90	1254.0
117.00	26456.0
118.00	1584.0
118.90	295.0
120.00	411.0
122.00	1942.0
123.00	2667.0
124.00	1422.0
124.90	1459.0
127.00	80704.0
128.00	5593.0
129.00	33592.0
130.00	2735.0
130.90	543.0
132.00	323.0
132.90	313.0
134.00	1199.0
135.00	3366.0
136.00	1246.0
137.00	1515.0
137.90	218.0
139.00	163.0
139.80	432.0
140.10	415.0
141.00	4401.0
142.00	1501.0
143.00	996.0
143.90	419.0
144.90	238.0
146.00	1023.0
147.00	2820.0
148.00	5896.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

149.00	1274.0
150.00	404.0
150.90	553.0
151.90	335.0
153.00	1536.0
153.90	1263.0
155.00	3145.0
156.00	3627.0
157.00	759.0
157.90	1053.0
159.00	587.0
160.00	2073.0
161.10	2529.0
161.90	618.0
163.90	266.0
165.00	2029.0
166.00	1534.0
167.00	9871.0
168.00	6077.0
169.00	1114.0
169.90	310.0
170.80	374.0
172.00	1067.0
172.90	1128.0
174.00	2312.0
175.00	3924.0
176.00	1100.0
177.00	2031.0
177.90	564.0
179.00	8284.0
180.00	5456.0
180.90	2332.0
181.90	299.0
182.70	214.0
183.90	636.0
185.00	4143.0
186.00	26856.0
187.00	8463.0
188.00	842.0
189.00	2125.0
190.00	531.0
191.00	853.0
192.00	2894.0
193.00	3116.0
194.00	539.0
194.90	172.0
196.00	5283.0
197.10	1059.0
198.00	173184.0
198.90	12216.0
199.90	1159.0
201.40	892.0
203.00	1612.0
204.00	7330.0
205.00	12003.0
206.00	49264.0
207.00	7394.0
208.00	2107.0
209.00	808.0
210.00	1117.0
211.10	2280.0
212.90	171.0
214.90	728.0
216.00	1386.0
217.00	14731.0
218.00	1572.0
219.10	251.0
221.00	6556.0
222.00	2017.0
223.00	3061.0
224.00	27784.0
225.00	7040.0
226.10	833.0
227.00	13521.0
228.00	1938.0
229.00	3047.0
230.00	317.0
231.00	1181.0
232.00	214.0
233.10	168.0
234.00	1112.0
234.90	976.0
236.00	645.0
236.90	980.0
237.90	342.0
238.90	504.0
240.00	375.0
241.00	944.0
242.10	1341.0
243.00	1388.0
244.00	21040.0
245.00	3163.0
246.00	5315.0
246.90	944.0
248.00	356.0
249.00	867.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

249.90	166.0
251.00	294.0
252.00	213.0
253.00	556.0
254.20	988.0
255.00	119344.0
256.00	17456.0
256.90	1181.0
258.00	8020.0
259.00	1623.0
260.00	251.0
260.80	274.0
264.00	206.0
265.00	3704.0
265.90	299.0
269.90	152.0
271.00	424.0
271.90	384.0
273.00	3578.0
274.00	8729.0
275.00	52960.0
276.00	6757.0
277.00	5234.0
278.00	828.0
280.90	428.0
281.90	172.0
282.90	645.0
284.00	639.0
285.00	969.0
286.00	195.0
288.90	209.0
289.90	289.0
292.00	240.0
292.90	1078.0
294.00	355.0
294.90	334.0
296.00	17232.0
297.00	2422.0
297.90	162.0
301.00	409.0
302.00	383.0
303.10	2217.0
303.90	390.0
307.90	313.0
312.90	194.0
314.00	878.0
315.00	2143.0
316.00	1239.0
316.90	169.0
321.00	513.0
322.00	299.0
323.00	4448.0
323.90	843.0
327.00	1161.0
328.00	471.0
331.90	551.0
333.00	394.0
334.00	3640.0
335.00	943.0
335.90	161.0
341.10	681.0
346.00	1351.0
346.90	199.0
352.10	1572.0
353.00	1076.0
354.10	1775.0
354.90	266.0
359.00	206.0
365.00	8204.0
366.00	955.0
370.90	350.0
372.00	2575.0
373.10	739.0
383.00	738.0
384.10	259.0
390.00	174.0
390.90	269.0
400.90	159.0
402.00	1243.0
403.00	1471.0
404.10	541.0
421.00	1321.0
422.00	1434.0
423.10	7450.0
424.00	2092.0
425.00	282.0
441.10	20696.0
442.10	141120.0
443.10	28376.0
444.00	2411.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

m/z	Abundance
38.00	436.0
39.10	2673.0
40.00	661.0
43.90	779.0
48.90	355.0
50.00	9222.0
51.10	32752.0
52.10	1544.0
55.10	314.0
56.00	1229.0
57.00	2421.0
61.00	658.0
62.00	731.0
63.00	1965.0
64.00	432.0
65.00	755.0
68.00	879.0
69.00	56432.0
70.00	202.0
73.10	807.0
74.00	6376.0
75.00	8494.0
76.00	3276.0
77.10	54736.0
78.00	3347.0
79.00	4538.0
80.00	2967.0
81.00	4570.0
81.90	1158.0
82.90	674.0
83.90	326.0
84.90	643.0
85.90	1506.0
87.00	595.0
90.90	1075.0
91.90	978.0
93.00	8124.0
93.90	305.0
95.00	278.0
96.10	259.0
98.00	5494.0
99.00	4625.0
100.10	394.0
101.00	2291.0
103.00	616.0
104.00	1701.0
105.00	1468.0
105.90	415.0
107.00	17288.0
108.00	2704.0
108.90	652.0
110.00	28144.0
111.00	4654.0
112.10	476.0
113.00	198.0
115.00	226.0
116.10	1198.0
117.00	19312.0
118.00	1538.0
120.00	206.0
121.90	1246.0
123.00	1824.0
123.90	855.0
124.90	681.0
125.90	303.0
127.00	54792.0
128.00	4057.0
129.00	23048.0
130.00	1837.0
130.90	258.0
132.00	170.0
134.00	821.0
135.00	1944.0
135.90	682.0
137.00	1331.0
138.00	171.0
140.00	429.0
141.00	3609.0
142.00	1298.0
142.90	648.0
144.10	169.0
144.90	209.0
146.00	789.0
147.00	1743.0
147.90	4293.0
148.90	1001.0
149.90	236.0
151.10	334.0
151.80	282.0
152.90	1192.0
154.00	630.0
155.00	1859.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

156.00	2629.0
157.20	486.0
157.90	727.0
158.90	509.0
160.00	1010.0
161.00	1933.0
162.00	456.0
163.00	164.0
163.80	271.0
165.00	1388.0
165.90	1038.0
167.00	8061.0
168.00	4154.0
169.00	710.0
170.00	251.0
170.90	407.0
171.90	785.0
173.00	862.0
174.00	1377.0
175.00	3220.0
176.00	561.0
176.90	1532.0
178.00	332.0
179.00	5859.0
180.00	3701.0
181.00	1781.0
181.90	523.0
182.70	306.0
184.00	631.0
185.00	3012.0
186.00	20008.0
187.00	6139.0
188.00	914.0
188.90	1474.0
190.10	208.0
191.00	956.0
192.00	1913.0
193.00	2102.0
194.00	291.0
195.10	221.0
195.90	3825.0
196.70	710.0
197.10	721.0
197.90	131040.0
198.90	7947.0
200.00	934.0
201.10	371.0
201.70	405.0
203.00	1102.0
204.00	5702.0
205.00	9695.0
206.00	39208.0
207.00	6129.0
208.00	1671.0
209.00	780.0
210.00	566.0
210.90	1751.0
211.80	330.0
212.90	222.0
215.00	443.0
216.00	868.0
217.00	12277.0
218.00	1325.0
221.00	5066.0
222.00	1514.0
223.00	2790.0
224.00	22528.0
225.00	4610.0
226.10	610.0
227.00	10926.0
228.00	1739.0
229.00	2034.0
229.90	281.0
231.00	1005.0
232.00	176.0
233.00	234.0
233.90	776.0
234.90	519.0
235.90	514.0
237.00	859.0
239.00	555.0
240.00	276.0
241.00	952.0
242.00	1458.0
243.10	1122.0
244.10	17952.0
245.00	2442.0
246.00	3994.0
247.00	762.0
247.90	249.0
249.00	844.0
251.00	238.0
252.00	190.0
252.90	379.0
254.10	590.0
255.00	94328.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20

256.00	14338.0
257.00	948.0
258.00	6816.0
259.10	728.0
263.80	238.0
265.00	2544.0
266.00	470.0
271.00	199.0
272.00	167.0
273.00	2876.0
274.00	6934.0
275.00	42504.0
276.00	5380.0
277.00	4164.0
277.90	896.0
280.90	553.0
281.90	187.0
283.00	558.0
283.90	347.0
285.10	644.0
290.00	343.0
291.90	212.0
292.90	923.0
293.90	316.0
295.00	363.0
296.00	14622.0
297.00	1853.0
298.00	207.0
300.80	221.0
302.10	341.0
303.00	1911.0
304.00	363.0
308.10	215.0
314.00	747.0
315.00	2007.0
316.00	886.0
321.00	531.0
321.90	183.0
323.00	4366.0
323.90	479.0
326.90	913.0
328.10	505.0
331.90	466.0
333.00	375.0
334.00	2785.0
335.10	519.0
335.90	174.0
340.90	563.0
342.00	291.0
345.90	1269.0
352.00	1434.0
353.10	1006.0
354.10	1736.0
355.10	344.0
358.90	190.0
365.00	7804.0
366.00	1049.0
371.00	335.0
372.00	2519.0
373.00	679.0
383.00	690.0
383.80	170.0
390.00	406.0
391.00	150.0
402.00	817.0
403.00	1548.0
404.00	443.0
421.00	1242.0
422.00	906.0
423.10	7899.0
424.00	2014.0
425.00	202.0
441.00	21920.0
442.10	142784.0
443.10	27864.0
444.10	2678.0

Instrument :
BNA_M
ClientSampleId :
DFTPP20