

Data Path : Z:\HPCHEM1\BNA M\DATA\BM080915\
 Data File : BM002313.D
 Acq On : 09 Aug 2015 13:55
 Operator : TP/UM
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD02018

Quant Time: Aug 10 05:36:32 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM080715MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 10 05:35:03 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.92	152	210251	20.00	ng/ul	0.00
2) Naphthalene-d8	11.79	136	973603	20.00	ng/ul	0.00
6) Acenaphthene-d10	15.50	164	498084	20.00	ng/ul	0.00
12) Phenanthrene-d10	18.21	188	1000430	20.00	ng/ul	0.00
17) Chrysene-d12	22.30	240	800308	20.00	ng/ul	0.00
22) Perylene-d12	25.00	264	688566	20.00	ng/ul	0.00

System Monitoring Compounds

7) Acenaphthylene-d8	15.20	160	1061310	20.57	ng/ul	0.00
10) Fluorene-d10	16.47	176	719214	20.04	ng/ul	0.00
14) Anthracene-d10	18.31	188	1004105	20.68	ng/ul	0.00
18) Pyrene-d10	20.52	212	922107	23.76	ng/ul	0.00
25) Benzo(a)pyrene-d12	24.82	264	755170	20.66	ng/ul	0.00

Target Compounds

						Ovalue
3) Naphthalene	11.85	128	1050506	20.49	ng/ul	100
4) 2-Methylnaphthalene	13.39	142	723979	19.77	ng/ul	99
5) 1-Methylnaphthalene	13.60	142	713093	19.76	ng/ul	100
8) Acenaphthylene	15.23	152	1106865	20.66	ng/ul	100
9) Acenaphthene	15.56	153	741900	20.59	ng/ul	100
11) Fluorene	16.53	166	813998	20.06	ng/ul	100
13) Phenanthrene	18.25	178	1195739	20.77	ng/ul	99
15) Anthracene	18.34	178	1229563	20.67	ng/ul	100
16) Fluoranthene	20.19	202	1226526	19.24	ng/ul	97
19) Pyrene	20.55	202	1232750	23.83	ng/ul	99
20) Benzo(a)anthracene	22.28	228	1007888	20.63	ng/ul	100
21) Chrysene	22.34	228	937528	20.25	ng/ul	100
23) Benzo(b)fluoranthene	24.16	252	894329	20.90	ng/ul	99
24) Benzo(k)fluoranthene	24.22	252	873812	21.26	ng/ul	100
26) Benzo(a)pyrene	24.87	252	836631	20.43	ng/ul	99
27) Indeno(1,2,3-cd)pyrene	27.85	276	943293	20.10	ng/ul	99
28) Dibenzo(a,h)anthracene	27.86	278	801237	20.15	ng/ul	99
29) Benzo(g,h,i)perylene	28.73	276	793438	20.04	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM080915\
Data File : BM002313.D
Acq On : 09 Aug 2015 13:55
Operator : TP/UM
Sample : SSTDCCC020EC
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD02018

Quant Time: Aug 10 05:36:32 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM080715MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Aug 10 05:35:03 2015
Response via : Initial Calibration

