

Data Path : Z:\HPCHEM1\BNA M\DATA\BM080915\
 Data File : BM002313.D
 Acq On : 09 Aug 2015 13:55
 Operator : TP/UM
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02018

Quant Time: Aug 10 05:36:32 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM080715MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 10 05:35:03 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	145	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	137	0.00
3	Naphthalene	1.053	1.079	-2.5	138	0.00
4	2-Methylnaphthalene	0.752	0.744	1.1	132	0.00
5	1-Methylnaphthalene	0.741	0.732	1.2	131	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
7 S	Acenaphthylene-d8	2.072	2.131	-2.8	123	0.00
8	Acenaphthylene	2.151	2.222	-3.3	123	0.00
9 C	Acenaphthene	1.447	1.490	-3.0	123	0.00
10 S	Fluorene-d10	1.441	1.444	-0.2	120	0.00
11	Fluorene	1.629	1.634	-0.3	120	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.00
13	Phenanthrene	1.151	1.195	-3.8	115	0.00
14 S	Anthracene-d10	0.970	1.004	-3.5	114	0.00
15	Anthracene	1.189	1.229	-3.4	114	0.00
16 C	Fluoranthene	1.275	1.226	3.8	105	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	89	0.00
18 S	Pyrene-d10	0.970	1.152	-18.8	102	0.00
19	Pyrene	1.293	1.540	-19.1	102	0.00
20	Benzo(a)anthracene	1.221	1.259	-3.1	89	0.00
21	Chrysene	1.157	1.171	-1.2	88	0.00
22 I	Perylene-d12	1.000	1.000	0.0	76	0.00
23	Benzo(b)fluoranthene	1.243	1.299	-4.5	79	0.00
24	Benzo(k)fluoranthene	1.194	1.269	-6.3	78	0.00
25 S	Benzo(a)pyrene-d12	1.062	1.097	-3.3	77	0.00
26 C	Benzo(a)pyrene	1.189	1.215	-2.2	76	0.00
27	Indeno(1,2,3-cd)pyrene	1.363	1.370	-0.5	73	0.00
28	Dibenzo(a,h)anthracene	1.155	1.164	-0.8	73	0.00
29	Benzo(g,h,i)perylene	1.150	1.152	-0.2	72	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0