


Data Path : Z:\HPCHEM1\BNA M\DATA\BM080915\
Data File : BM002303.D
Acq On : 09 Aug 2015 07:53
Operator : TP/UM
Sample : DFTPP92
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP92

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080715MA.M
Title : SVOA CALIBRATION
Last Update : Mon Aug 10 05:35:03 2015



AutoFind: Scans 2630, 2631, 2632; Background Corrected with Scan 2624

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.4	65754	PASS
68	69	0.00	2	2.0	1222	PASS
69	198	0.00	100	39.7	60201	PASS
70	69	0.00	2	0.7	423	PASS
127	198	10	80	50.9	77117	PASS
197	198	0.00	2	0.7	1014	PASS
198	198	100	100	100.0	151616	PASS
199	198	5	9	6.6	10039	PASS
275	198	10	60	27.5	41770	PASS
365	198	1	100	3.3	4998	PASS
441	443	0.01	100	80.0	21867	PASS
442	198	50	100	93.4	141586	PASS
443	442	15	24	19.3	27322	PASS

m/z	Abundance
39.10	300.0
40.00	1784.0
41.10	694.0
43.00	567.0
44.00	10588.0
55.00	823.0
57.00	424.0
69.00	351.0
73.00	730.0
91.00	378.0
146.90	324.0
193.10	342.0
207.10	2071.0
208.00	399.0
209.10	319.0
281.10	545.0

Instrument :
BNA_M
ClientSampleId :
DFTPP92

m/z	Abundance
38.10	961.0
39.10	5323.0
40.00	1959.0
41.10	847.0
43.10	692.0
44.00	9989.0
45.00	360.0
49.00	572.0
50.10	17232.0
51.10	69288.0
52.00	3970.0
55.00	1122.0
56.00	2233.0
57.00	5372.0
61.00	991.0
62.10	1054.0
63.10	2573.0
64.00	460.0
65.00	1280.0
67.00	396.0
68.10	1259.0
69.00	64808.0
70.10	411.0
73.00	1215.0
74.00	6031.0
75.00	10037.0
76.10	3290.0
77.10	72096.0
78.10	5147.0
79.00	4641.0
80.00	3451.0
81.00	5084.0
82.00	1244.0
83.00	1369.0
85.00	1017.0
86.00	1336.0
87.00	648.0
91.00	1364.0
92.00	1249.0
93.00	7656.0
94.00	624.0
95.00	319.0
96.00	677.0
97.10	365.0
98.00	5665.0
99.00	4486.0
100.00	449.0
101.00	2657.0
103.00	888.0
104.00	1597.0
105.00	1568.0
106.10	689.0
107.00	20552.0
108.00	3591.0
109.20	779.0
110.00	40536.0
111.00	5922.0
112.00	807.0
116.10	1216.0
117.00	15579.0
118.00	1170.0
119.00	350.0
120.00	345.0
122.00	1353.0
123.00	2176.0
123.90	1095.0
125.00	957.0
127.00	76576.0
128.00	6174.0
129.00	27248.0
130.00	2610.0
130.90	537.0
132.00	301.0
133.00	355.0
133.90	734.0
135.10	2309.0
136.00	877.0
137.00	1315.0
140.00	348.0
141.00	3895.0
142.10	1244.0
142.90	830.0
144.80	403.0
146.00	641.0
147.00	2225.0
148.00	3940.0
149.00	1121.0
151.20	553.0
152.00	465.0
153.00	1131.0
154.00	794.0

Instrument :
BNA_M
ClientSampleId :
DFTPP92

155.00	2080.0
156.00	3176.0
157.10	596.0
158.00	656.0
159.00	588.0
160.00	1174.0
161.00	1781.0
162.00	563.0
165.00	1295.0
166.00	1078.0
167.00	7160.0
168.00	3473.0
169.00	645.0
170.90	328.0
171.90	581.0
172.90	850.0
174.00	1634.0
175.00	2620.0
176.10	914.0
177.00	1225.0
177.90	446.0
178.90	4456.0
180.00	3505.0
181.00	1578.0
181.90	326.0
183.90	559.0
185.00	2373.0
186.00	18800.0
187.00	5755.0
188.00	648.0
189.00	1185.0
191.00	762.0
192.00	1683.0
193.00	1810.0
193.90	364.0
195.90	4058.0
196.80	811.0
197.10	848.0
198.00	138816.0
199.00	9416.0
200.00	700.0
201.50	655.0
203.00	858.0
204.00	4673.0
205.00	7804.0
206.00	36216.0
207.00	6663.0
208.00	1464.0
209.00	640.0
211.00	1315.0
214.90	361.0
216.00	780.0
217.00	8880.0
218.00	1039.0
221.00	6773.0
223.00	2022.0
224.00	19008.0
225.00	4795.0
226.00	494.0
227.00	7964.0
228.00	994.0
229.00	1552.0
231.00	649.0
234.00	504.0
234.90	516.0
236.00	399.0
236.90	679.0
239.00	369.0
241.00	376.0
242.00	825.0
243.00	1204.0
244.00	15023.0
245.00	2164.0
246.00	2780.0
247.00	581.0
249.00	546.0
253.10	344.0
255.00	76424.0
256.00	11055.0
257.00	949.0
258.00	4260.0
259.00	754.0
265.00	1675.0
273.00	2398.0
274.00	6453.0
275.00	36208.0
276.00	4628.0
277.00	2637.0
278.00	513.0
281.10	558.0
283.10	383.0
283.90	301.0
285.00	400.0
293.00	748.0
296.00	8470.0

Instrument :
BNA_M
ClientSampleId :
DFTPP92

297.00	1231.0
303.10	960.0
304.00	343.0
314.00	359.0
314.90	904.0
315.90	402.0
323.10	2854.0
324.10	515.0
327.00	701.0
328.00	386.0
333.00	327.0
334.00	1981.0
334.90	517.0
341.00	474.0
346.00	688.0
352.00	870.0
353.10	591.0
354.00	1203.0
354.90	317.0
365.00	3811.0
366.00	535.0
372.00	1685.0
373.00	390.0
383.00	499.0
402.00	768.0
403.00	1076.0
404.00	333.0
421.00	803.0
422.00	655.0
423.00	5560.0
424.00	1310.0
441.00	15427.0
442.00	100984.0
443.00	21104.0
444.00	1705.0

Instrument :
BNA_M
ClientSampleId :
DFTPP92

m/z	Abundance
37.00	365.0
38.00	1069.0
39.10	6509.0
40.00	1974.0
41.10	846.0
43.10	632.0
44.00	10051.0
45.10	353.0
49.00	664.0
50.10	19096.0
51.10	78280.0
52.10	3764.0
53.00	334.0
55.00	1036.0
56.00	2499.0
57.00	5631.0
61.00	883.0
62.00	1104.0
63.00	3155.0
64.10	577.0
65.10	1377.0
67.00	483.0
68.10	1462.0
69.00	71008.0
70.00	533.0
73.00	1220.0
74.00	7330.0
75.00	10723.0
76.10	3848.0
77.10	81344.0
78.00	6023.0
79.00	5436.0
80.00	3737.0
81.00	6129.0
81.90	1658.0
83.00	1758.0
84.00	409.0
85.00	1087.0
86.00	1451.0
87.10	863.0
91.00	1644.0
92.00	1601.0
93.00	8738.0
94.00	669.0
95.10	382.0
96.10	727.0
97.10	369.0
98.00	6510.0
99.00	5435.0
100.00	611.0
101.00	3366.0
103.00	1213.0
104.00	2007.0
105.00	2142.0
105.90	637.0
107.00	24248.0
108.00	3843.0
109.10	1052.0
110.00	46872.0
111.00	7352.0
112.00	952.0
113.00	365.0
115.00	305.0
116.00	1420.0
117.00	18168.0
118.00	1303.0
119.00	411.0
120.00	349.0
122.00	1686.0
123.00	2614.0
124.00	1239.0
125.00	1033.0
127.00	91288.0
128.00	7027.0
129.00	33864.0
130.00	2915.0
130.90	759.0
132.00	384.0
132.90	335.0
133.90	1172.0
135.00	2877.0
136.00	1046.0
137.00	1285.0
137.90	366.0
140.00	551.0
141.00	4347.0
142.00	1376.0
143.00	909.0
144.80	375.0
146.00	809.0
147.00	2308.0

Instrument :
BNA_M
ClientSampleId :
DFTPP92

148.00	4901.0
149.00	1024.0
151.00	463.0
151.40	529.0
151.80	424.0
152.90	1468.0
154.00	1136.0
155.00	2458.0
156.00	3515.0
157.00	702.0
158.00	938.0
158.90	657.0
160.00	1418.0
161.00	2218.0
161.90	709.0
164.90	1515.0
166.00	1254.0
167.00	8368.0
168.00	4180.0
169.00	683.0
170.90	466.0
172.00	739.0
173.00	946.0
174.00	1939.0
175.00	3389.0
176.00	1115.0
177.00	1484.0
178.00	500.0
179.00	6701.0
180.00	4648.0
181.00	1943.0
182.20	302.0
183.90	422.0
185.00	2941.0
186.00	25184.0
187.00	6842.0
188.10	681.0
189.00	1428.0
190.00	304.0
190.90	1050.0
192.00	2088.0
193.00	2540.0
194.00	482.0
195.00	339.0
196.00	5813.0
197.10	1240.0
198.00	182336.0
199.00	12230.0
200.00	1153.0
201.40	974.0
203.00	1247.0
204.00	5849.0
205.00	10763.0
206.00	47520.0
207.00	8070.0
208.00	1849.0
209.00	788.0
210.20	1067.0
211.10	1703.0
215.00	547.0
216.00	853.0
216.90	11307.0
217.90	1703.0
221.00	7868.0
223.00	2584.0
224.00	24520.0
225.00	6155.0
226.10	678.0
227.00	9675.0
228.00	1357.0
229.00	2170.0
230.90	1013.0
233.90	594.0
235.00	720.0
236.00	511.0
236.90	760.0
239.00	349.0
240.00	349.0
241.00	639.0
242.00	1148.0
243.00	1414.0
244.00	19712.0
245.00	2958.0
246.00	3916.0
247.00	671.0
248.90	795.0
252.90	464.0
255.00	104896.0
256.00	15123.0
257.00	1415.0
258.00	5458.0
259.00	947.0
265.00	2405.0
265.90	342.0
273.00	3336.0

Instrument :
BNA_M
ClientSampleId :
DFTPP92

274.00	8101.0
275.00	49440.0
276.00	6624.0
277.00	3970.0
277.90	713.0
281.00	545.0
282.90	455.0
284.00	323.0
285.00	840.0
292.90	907.0
296.00	11456.0
297.00	1832.0
303.10	1395.0
304.00	447.0
314.00	669.0
314.90	1463.0
315.90	660.0
321.00	494.0
323.10	4259.0
324.00	842.0
326.90	802.0
328.10	427.0
331.90	318.0
332.90	453.0
334.00	2977.0
335.00	766.0
341.10	628.0
346.00	1117.0
352.00	1330.0
353.00	868.0
354.00	1599.0
355.10	410.0
365.00	5892.0
366.00	830.0
371.00	439.0
372.00	2689.0
373.00	701.0
383.10	671.0
383.90	311.0
390.00	448.0
401.90	1084.0
403.00	1365.0
404.00	545.0
421.00	1326.0
422.00	1453.0
423.00	9018.0
424.00	2069.0
441.00	25392.0
442.00	163968.0
443.00	31000.0
444.00	2830.0

Instrument :
BNA_M
ClientSampleId :
DFTPP92

m/z	Abundance
38.00	777.0
39.10	4215.0
40.00	1807.0
41.00	846.0
43.00	731.0
44.00	9973.0
49.10	474.0
50.00	12648.0
51.10	49696.0
52.10	2583.0
55.00	1095.0
56.00	1688.0
57.00	3722.0
61.00	679.0
62.00	653.0
63.00	2162.0
65.00	1072.0
67.00	339.0
68.00	946.0
69.00	45840.0
70.10	326.0
73.00	1180.0
74.00	4679.0
75.00	7312.0
76.10	2639.0
77.10	54768.0
78.10	3805.0
79.00	3619.0
80.00	2830.0
81.00	3979.0
82.00	1053.0
83.00	1096.0
85.00	970.0
86.00	1135.0
87.00	552.0
91.00	1006.0
92.10	996.0
93.00	5723.0
93.90	519.0
95.00	344.0
96.10	656.0
97.10	426.0
98.00	4830.0
99.00	3485.0
100.10	307.0
101.00	2005.0
103.00	917.0
104.00	1427.0
105.00	1435.0
106.10	486.0
107.00	16301.0
108.00	2769.0
109.00	649.0
110.00	32696.0
111.00	5087.0
111.90	625.0
116.00	954.0
117.00	12677.0
118.00	987.0
121.90	1290.0
123.00	2100.0
124.00	760.0
125.00	905.0
127.00	63488.0
128.00	4688.0
129.00	23832.0
129.90	2020.0
131.10	454.0
132.90	342.0
133.90	742.0
135.00	2063.0
135.90	722.0
137.00	1042.0
140.00	326.0
141.00	3437.0
142.00	1090.0
142.90	732.0
146.00	628.0
147.00	1735.0
148.00	3303.0
149.00	814.0
151.00	453.0
153.00	1076.0
154.00	735.0
155.00	1723.0
156.00	2462.0
157.00	572.0
158.00	681.0
158.90	545.0
160.00	1078.0
160.90	1576.0

Instrument :
BNA_M
ClientSampleId :
DFTPP92

161.90	558.0
165.00	1348.0
166.00	884.0
167.00	6242.0
168.00	3077.0
169.00	652.0
172.00	549.0
172.90	863.0
174.00	1343.0
175.00	2203.0
175.90	911.0
177.00	1443.0
178.10	505.0
179.00	4508.0
180.00	3333.0
181.00	1716.0
183.90	556.0
185.00	2424.0
186.00	17392.0
187.00	5396.0
188.00	483.0
188.90	1256.0
191.00	763.0
192.00	1664.0
193.00	1815.0
194.00	520.0
195.00	323.0
196.00	4282.0
197.10	954.0
198.00	133696.0
199.00	8473.0
200.00	707.0
201.40	790.0
203.00	847.0
204.00	4494.0
205.00	7844.0
206.00	35792.0
207.00	6058.0
208.00	1530.0
209.00	549.0
210.30	708.0
211.00	1608.0
215.00	444.0
216.00	755.0
217.00	8317.0
217.90	1150.0
221.10	6291.0
223.00	2188.0
224.00	19392.0
225.00	4720.0
225.90	498.0
227.00	7665.0
228.00	966.0
229.00	1593.0
231.10	679.0
234.00	504.0
235.00	560.0
236.00	446.0
237.00	571.0
239.00	469.0
240.90	445.0
241.90	922.0
243.10	957.0
244.00	15323.0
245.00	1901.0
246.00	2830.0
246.90	484.0
248.90	562.0
253.00	370.0
254.10	945.0
255.00	82936.0
256.00	11719.0
257.00	927.0
258.00	4411.0
259.00	530.0
265.00	2031.0
266.00	413.0
272.00	320.0
273.00	2481.0
274.00	6788.0
275.00	39664.0
276.00	5290.0
277.00	3319.0
278.00	436.0
281.10	576.0
283.00	446.0
285.00	486.0
293.00	679.0
295.10	310.0
296.00	10700.0
297.00	1477.0
303.10	1257.0
304.00	382.0
314.00	470.0
314.90	1183.0

Instrument :
BNA_M
ClientSampleId :
DFTPP92

316.00	710.0
320.90	383.0
323.00	3881.0
324.00	763.0
327.00	874.0
328.00	422.0
333.00	380.0
334.10	2326.0
335.00	594.0
341.00	520.0
345.90	903.0
352.10	1148.0
353.00	752.0
354.00	1310.0
355.00	393.0
365.00	5293.0
366.00	788.0
370.90	339.0
372.00	2249.0
373.00	639.0
383.00	589.0
402.00	900.0
403.00	1397.0
404.00	471.0
421.00	1212.0
422.00	1348.0
423.00	8073.0
424.00	1712.0
441.00	24784.0
442.00	159808.0
443.00	29864.0
444.00	2965.0

Instrument :
BNA_M
ClientSampleId :
DFTPP92