


Data Path : Z:\HPCHEM1\BNA_M\Data\BM081015\
Data File : BM002314.D
Acq On : 10 Aug 2015 13:02
Operator : UM/IZ
Sample : DFTPP93
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP93

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080715.M
Title : SVOA CALIBRATION
Last Update : Tue Aug 11 01:34:33 2015



AutoFind: Scans 2627, 2628, 2629; Background Corrected with Scan 2622

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.0	65416	PASS
68	69	0.00	2	2.0	1270	PASS
69	198	0.00	100	38.6	63043	PASS
70	69	0.00	2	0.8	513	PASS
127	198	10	80	52.3	85450	PASS
197	198	0.00	2	0.4	718	PASS
198	198	100	100	100.0	163384	PASS
199	198	5	9	6.7	11008	PASS
275	198	10	60	26.2	42832	PASS
365	198	1	100	3.0	4911	PASS
441	443	0.01	100	78.3	20900	PASS
442	198	50	100	84.8	138509	PASS
443	442	15	24	19.3	26677	PASS

m/z	Abundance
36.90	324.0
38.00	878.0
39.00	5172.0
40.00	1150.0
41.10	1005.0
42.10	371.0
43.00	1024.0
44.00	8082.0
45.00	461.0
49.00	546.0
50.00	15758.0
51.10	62984.0
52.10	3223.0
53.00	323.0
55.00	951.0
56.00	1878.0
57.00	4398.0
60.10	353.0
61.00	744.0
62.00	845.0
63.00	2516.0
64.00	351.0
65.00	1163.0
68.00	1353.0
69.00	60680.0
70.00	493.0
73.00	1121.0
74.00	6158.0
75.00	9601.0
76.10	3446.0
77.10	70704.0
78.10	5100.0
79.00	4441.0
80.00	3639.0
81.00	5023.0
82.00	1352.0
83.00	1278.0
85.00	914.0
86.00	1219.0
87.00	693.0
91.00	1378.0
92.00	1303.0
93.00	7675.0
94.00	584.0
96.00	590.0
98.00	6062.0
99.00	4747.0
100.00	442.0
101.00	3170.0
103.00	1078.0
104.00	1954.0
105.00	1932.0
106.00	613.0
107.00	22336.0
108.00	3385.0
109.10	818.0
110.00	43576.0
111.00	6380.0
112.00	725.0
116.00	1328.0
117.00	17064.0
118.00	1321.0
119.00	339.0
120.00	454.0
122.00	1363.0
123.00	2201.0
124.00	1169.0
125.00	1078.0
127.00	88184.0
128.00	6769.0
129.00	33272.0
130.00	2903.0
131.00	630.0
132.20	316.0
132.90	394.0
134.00	1130.0
135.00	2794.0
135.90	1079.0
137.00	1445.0
137.90	388.0
140.00	386.0
141.00	4319.0
142.00	1360.0
143.00	851.0
144.00	333.0
146.00	736.0
147.00	2582.0
148.00	4878.0
149.00	942.0
151.00	691.0
152.10	374.0

Instrument :
BNA_M
ClientSampleId :
DFTPP93

153.00	1433.0
154.00	1034.0
155.00	2427.0
156.00	3179.0
157.00	642.0
157.90	797.0
158.90	639.0
160.00	1423.0
160.90	1726.0
162.00	664.0
165.00	1556.0
166.00	1192.0
167.00	8235.0
168.00	4467.0
168.90	792.0
170.00	336.0
170.90	360.0
172.00	745.0
173.00	911.0
174.00	1959.0
175.00	3122.0
176.00	1069.0
177.00	1674.0
177.90	522.0
179.00	6835.0
180.00	4639.0
181.00	2166.0
182.00	404.0
184.00	411.0
185.00	2914.0
186.00	23424.0
187.00	6850.0
188.00	598.0
188.90	1327.0
191.00	936.0
192.00	1993.0
193.00	2413.0
194.00	420.0
195.10	359.0
196.00	5285.0
197.00	1013.0
197.90	179776.0
198.90	11804.0
200.00	915.0
201.50	856.0
203.00	1199.0
204.00	6272.0
205.00	10283.0
206.00	45432.0
207.00	7594.0
208.00	1671.0
209.00	693.0
210.20	853.0
211.00	1865.0
214.90	487.0
215.90	862.0
217.00	11591.0
217.90	1673.0
221.00	7874.0
221.80	1990.0
223.00	2859.0
224.00	25568.0
225.00	6573.0
226.00	710.0
227.00	10139.0
228.00	1459.0
229.00	2053.0
229.90	315.0
231.00	1006.0
233.90	669.0
234.90	683.0
235.90	405.0
237.00	965.0
239.00	466.0
240.00	400.0
241.00	488.0
241.90	1134.0
243.00	1424.0
244.00	20144.0
245.00	2742.0
246.00	3791.0
246.90	745.0
249.00	691.0
252.90	424.0
254.00	871.0
255.00	108352.0
256.00	15191.0
257.00	1258.0
258.00	5758.0
258.90	950.0
265.00	2185.0
266.00	419.0
273.00	2953.0
274.00	8200.0
275.00	50240.0

Instrument :
BNA_M
ClientSampleId :
DFTPP93

276.00	6601.0
277.00	4017.0
278.00	780.0
281.00	496.0
283.00	633.0
283.90	363.0
285.00	688.0
292.90	907.0
295.00	331.0
296.00	13038.0
297.00	1754.0
303.00	1776.0
303.90	386.0
314.00	682.0
314.90	1433.0
316.00	1079.0
321.10	415.0
323.00	4830.0
323.90	766.0
327.00	929.0
328.00	482.0
332.10	376.0
332.90	357.0
334.00	2832.0
335.00	780.0
341.00	600.0
345.90	970.0
352.00	1567.0
353.00	879.0
354.10	1557.0
355.00	310.0
365.00	6117.0
365.90	913.0
370.90	383.0
372.00	2801.0
373.00	711.0
383.00	696.0
389.90	319.0
402.00	1217.0
403.00	1493.0
404.00	530.0
421.00	1333.0
422.00	1243.0
423.00	10272.0
424.00	1907.0
441.00	28160.0
442.00	185664.0
443.00	34968.0
444.00	2882.0

Instrument :
BNA_M
ClientSampleId :
DFTPP93

m/z	Abundance
37.00	411.0
38.00	1008.0
39.10	6357.0
40.00	1196.0
41.00	878.0
42.00	422.0
43.00	1034.0
44.00	8028.0
45.00	435.0
49.00	486.0
50.00	19272.0
51.10	78024.0
52.10	4146.0
53.00	305.0
55.00	1020.0
56.00	2553.0
57.00	5524.0
57.90	416.0
60.00	327.0
61.00	1112.0
62.00	1084.0
63.00	2937.0
64.10	389.0
65.00	1460.0
67.00	314.0
68.10	1505.0
69.00	76400.0
70.00	571.0
73.10	1299.0
74.00	8106.0
75.00	12129.0
76.10	4088.0
77.00	86608.0
78.10	6283.0
79.00	5904.0
80.00	4355.0
81.00	6351.0
82.00	1768.0
83.00	1717.0
85.00	1155.0
86.00	1781.0
87.00	885.0
91.00	1411.0
92.00	1440.0
93.00	9012.0
94.00	682.0
95.00	341.0
96.10	679.0
97.00	383.0
98.00	7547.0
99.00	6141.0
100.00	518.0
101.00	3317.0
102.90	1105.0
104.00	2187.0
105.00	2158.0
106.00	727.0
107.00	26840.0
108.00	4530.0
109.00	996.0
110.00	50400.0
111.00	7172.0
112.00	1028.0
116.00	1523.0
117.00	20944.0
117.90	1511.0
119.90	362.0
122.00	1764.0
123.00	2976.0
123.90	1283.0
125.00	1214.0
126.10	671.0
127.00	102272.0
128.00	8240.0
129.00	38320.0
130.00	3253.0
130.90	851.0
132.00	382.0
133.00	336.0
133.90	1088.0
135.00	3374.0
136.00	1364.0
137.00	1751.0
138.00	373.0
139.90	432.0
141.00	4579.0
141.90	1529.0
142.90	1107.0
144.90	341.0
146.00	815.0
147.00	2712.0

Instrument :
BNA_M
ClientSampleId :
DFTPP93

148.00	5324.0
149.00	1493.0
150.00	373.0
150.90	647.0
151.80	463.0
152.90	1456.0
153.90	1138.0
155.00	2444.0
156.00	3995.0
157.10	842.0
158.00	991.0
159.00	657.0
159.90	1420.0
161.00	2534.0
162.00	750.0
162.90	351.0
164.00	368.0
164.90	1922.0
166.00	1460.0
167.00	9103.0
168.00	4848.0
169.00	803.0
170.10	418.0
170.90	398.0
171.90	703.0
173.00	950.0
174.00	2097.0
175.00	3742.0
176.00	1039.0
177.00	1656.0
178.00	608.0
179.00	7314.0
180.00	5500.0
181.00	2252.0
182.00	459.0
183.90	632.0
185.00	3766.0
186.00	27784.0
187.00	7620.0
187.90	680.0
189.00	1489.0
191.00	964.0
191.90	2199.0
193.00	2740.0
193.90	609.0
194.90	424.0
196.00	5632.0
197.00	1141.0
197.90	191808.0
198.90	12757.0
200.00	1125.0
201.40	1045.0
202.90	1141.0
204.00	6435.0
205.00	11061.0
206.00	50544.0
207.00	8391.0
208.00	1697.0
209.00	719.0
210.20	830.0
211.00	2158.0
214.90	567.0
216.00	970.0
217.00	12188.0
218.00	1473.0
221.00	8177.0
221.80	2069.0
223.00	2548.0
224.00	26416.0
225.00	7149.0
226.00	840.0
227.00	11056.0
228.00	1407.0
228.90	2474.0
229.90	378.0
231.00	945.0
233.00	304.0
233.90	653.0
235.00	829.0
235.90	483.0
237.00	956.0
238.90	417.0
240.00	319.0
241.00	466.0
242.00	1414.0
243.00	1416.0
244.00	21136.0
245.00	2991.0
246.00	3396.0
247.00	639.0
249.00	784.0
252.90	410.0
254.10	789.0
255.00	107400.0
256.00	14962.0

Instrument :
BNA_M
ClientSampleId :
DFTPP93

256.90	1141.0
258.00	5859.0
258.90	953.0
264.90	2595.0
266.00	341.0
273.00	3149.0
274.00	8809.0
275.00	50800.0
276.00	6626.0
277.00	3657.0
278.00	689.0
280.90	470.0
283.10	449.0
284.00	315.0
285.10	604.0
292.90	801.0
294.00	433.0
294.80	355.0
296.00	12728.0
296.90	1659.0
303.00	1517.0
304.00	411.0
313.90	560.0
315.00	1479.0
316.00	775.0
321.00	489.0
323.00	4507.0
324.00	932.0
327.00	960.0
328.00	390.0
333.00	462.0
334.00	2623.0
335.00	707.0
341.00	504.0
345.90	932.0
352.00	1387.0
353.00	1033.0
354.00	1300.0
365.00	5791.0
365.90	901.0
371.00	455.0
372.00	2324.0
373.00	662.0
383.00	590.0
402.00	887.0
403.00	1331.0
404.00	575.0
421.00	1073.0
422.00	1192.0
423.00	8670.0
423.90	1891.0
441.00	23408.0
442.00	159040.0
443.00	31032.0
444.00	2566.0

Instrument :
BNA_M
ClientSampleId :
DFTPP93

m/z	Abundance
38.00	796.0
39.10	5162.0
40.00	1227.0
41.00	1029.0
41.90	303.0
43.10	902.0
44.00	8244.0
45.00	377.0
49.00	439.0
50.00	13599.0
51.00	55240.0
52.00	3278.0
55.10	810.0
56.00	1689.0
57.00	4167.0
61.00	586.0
62.00	698.0
63.00	2110.0
65.00	1008.0
68.00	954.0
69.00	53200.0
70.00	475.0
73.10	1006.0
74.00	5289.0
75.00	8187.0
76.10	2707.0
77.00	60376.0
78.00	4653.0
79.00	3993.0
80.00	2962.0
81.00	4415.0
82.00	1164.0
83.00	1049.0
83.90	322.0
85.00	850.0
86.00	1180.0
87.00	630.0
88.00	313.0
91.00	1139.0
91.90	999.0
93.00	6587.0
94.00	500.0
95.00	309.0
95.90	461.0
97.00	306.0
98.00	4860.0
99.00	4272.0
100.00	410.0
100.90	2122.0
103.00	849.0
104.00	1486.0
105.00	1342.0
105.90	487.0
107.00	17712.0
108.00	2804.0
109.00	590.0
110.00	33600.0
111.00	5330.0
112.00	762.0
113.00	302.0
116.00	1034.0
117.00	13563.0
117.90	1083.0
119.90	354.0
121.90	1086.0
123.00	1818.0
123.90	866.0
125.00	972.0
126.10	303.0
127.00	65896.0
128.00	5169.0
129.00	24024.0
130.00	2207.0
131.00	417.0
134.00	680.0
135.00	2123.0
136.00	914.0
137.00	1178.0
140.00	405.0
141.00	3027.0
142.00	995.0
142.90	693.0
145.90	654.0
146.90	1854.0
148.00	3491.0
149.00	913.0
151.00	463.0
153.00	1007.0
153.80	717.0
155.00	1712.0
156.00	2717.0

Instrument :
BNA_M
ClientSampleId :
DFTPP93

157.10	476.0
158.00	493.0
159.00	494.0
159.90	952.0
161.00	1347.0
161.90	432.0
164.90	1222.0
165.90	965.0
167.00	5786.0
168.00	3249.0
169.00	560.0
172.00	534.0
172.90	744.0
174.00	1403.0
175.00	2046.0
175.80	641.0
177.00	1355.0
178.00	444.0
178.90	4114.0
180.00	3141.0
181.00	1631.0
183.90	413.0
185.00	1831.0
186.00	16936.0
187.00	4553.0
188.00	525.0
189.00	1035.0
191.00	800.0
191.90	1217.0
193.00	1747.0
194.00	344.0
196.00	3696.0
197.90	118568.0
198.90	8465.0
200.00	689.0
201.20	472.0
201.60	414.0
202.90	690.0
204.00	4123.0
205.00	6772.0
206.00	29384.0
207.00	5355.0
207.90	1321.0
209.00	557.0
210.20	601.0
210.90	1134.0
214.90	301.0
216.00	620.0
217.00	7377.0
218.00	905.0
221.00	4815.0
221.80	1108.0
223.00	1441.0
224.00	15698.0
225.00	4286.0
225.90	454.0
227.00	6542.0
227.90	724.0
228.90	1288.0
230.90	578.0
233.80	450.0
235.00	370.0
235.90	337.0
236.90	634.0
240.90	452.0
241.90	749.0
243.00	958.0
244.00	11859.0
245.00	1591.0
246.00	2112.0
246.90	417.0
249.00	504.0
254.10	561.0
255.00	61264.0
256.00	9098.0
257.00	674.0
258.00	3505.0
258.90	523.0
265.00	1235.0
273.00	1722.0
274.00	4799.0
275.00	27456.0
276.00	3979.0
277.00	2040.0
277.90	474.0
280.90	388.0
285.00	367.0
293.00	447.0
296.00	6496.0
297.00	1027.0
302.90	934.0
313.90	358.0
315.00	810.0
316.00	440.0
323.00	2383.0

Instrument :
BNA_M
ClientSampleId :
DFTPP93

323.90	378.0
326.90	474.0
333.90	1400.0
335.00	386.0
345.80	503.0
351.90	582.0
352.90	486.0
353.90	640.0
364.90	2826.0
365.90	476.0
372.00	1236.0
383.00	336.0
401.90	473.0
403.00	705.0
421.00	701.0
421.90	567.0
423.00	4153.0
424.00	928.0
441.00	11133.0
442.00	70824.0
443.00	14031.0
444.00	1535.0

Instrument :
BNA_M
ClientSampleId :
DFTPP93

m/z	Abundance
39.00	390.0
39.90	945.0
41.10	962.0
42.00	364.0
43.10	936.0
44.00	8371.0
55.00	738.0
57.00	493.0
69.00	383.0
73.00	632.0
207.00	1753.0
207.90	339.0
280.90	423.0

Instrument :
BNA_M
ClientSampleId :
DFTPP93