

Data Path : Z:\HPCHEM1\BNA_M\Data\BM081016\
 Data File : BM007022.D
 Acq On : 11 Aug 2016 04:24
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02044

Quant Time: Aug 11 04:55:56 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080916.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 10 01:30:54 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2 S	1,4-Dioxane-d8	0.439	0.420	4.3	78	0.00
3 S	Phenol-d5	1.780	1.700	4.5	79	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	0.995	0.973	2.2	79	0.00
5 S	2-Chlorophenol-d4	1.350	1.334	1.2	82	0.00
6 S	4-Methylphenol-d8	1.507	1.427	5.3	78	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	80	0.00
8 S	Nitrobenzene-d5	0.153	0.150	2.0	77	0.00
9 S	2-Nitrophenol-d4	0.175	0.165	5.7	74	0.00
10 S	2,4-Dichlorophenol-d3	0.342	0.344	-0.6	80	0.00
11	Naphthalene	1.003	1.029	-2.6	81	0.00
12 S	4-Chloroaniline-d4	0.399	0.424	-6.3	77	0.00
13	2-Methylnaphthalene	0.794	0.803	-1.1	79	0.00
14	1-Methylnaphthalene	0.762	0.764	-0.3	79	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	77	0.00
16 S	Dimethylphthalate-d6	1.709	1.676	1.9	73	0.00
17 S	Acenaphthylene-d8	2.002	2.091	-4.4	78	0.00
18	Acenaphthylene	2.069	2.160	-4.4	78	0.00
19 C	Acenaphthene	1.338	1.380	-3.1	76	0.00
20 S	4-Nitrophenol-d4	0.307	0.293	4.6	72	0.00
21 S	Fluorene-d10	1.453	1.482	-2.0	76	0.00
22	Fluorene	1.588	1.639	-3.2	76	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	76	0.00
24 S	4,6-Dinitro-2-methylphenol-	0.108	0.099	8.3	72	0.00
25	Phenanthrene	1.072	1.114	-3.9	76	0.00
26 S	Anthracene-d10	0.932	0.968	-3.9	76	0.00
27	Anthracene	1.107	1.152	-4.1	76	0.00
28 C	Fluoranthene	1.350	1.393	-3.2	76	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	73	0.00
30 S	Pyrene-d10	0.909	0.975	-7.3	75	0.00
31	Pyrene	1.160	1.272	-9.7	76	0.00
32	Benzo(a)anthracene	1.178	1.215	-3.1	73	0.00
33	Chrysene	1.069	1.106	-3.5	73	0.00
34 I	Perylene-d12	1.000	1.000	0.0	66	0.00
35	Benzo(b)fluoranthene	1.242	1.299	-4.6	67	0.00
36	Benzo(k)fluoranthene	1.174	1.258	-7.2	68	0.00
37 S	Benzo(a)pyrene-d12	0.919	0.955	-3.9	67	0.00
38 C	Benzo(a)pyrene	1.143	1.178	-3.1	66	0.00
39	Indeno(1,2,3-cd)pyrene	1.133	1.275	-12.5	74	-0.01
40	Dibenzo(a,h)anthracene	0.938	1.071	-14.2	74	-0.01
41	Benzo(g,h,i)perylene	0.932	1.057	-13.4	75	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM081016\
Data File : BM007022.D
Acq On : 11 Aug 2016 04:24
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02044

Quant Time: Aug 11 04:55:56 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080916.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Aug 10 01:30:54 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0