

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM081018\
 Data File : BM016310.D
 Acq On : 10 Aug 2018 23:41
 Operator : SJ/JU
 Sample : J4272-08
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 JC-03-080918-D

Quant Time: Aug 11 05:37:05 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM072318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 10 18:44:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.13	152	27080	20.00	ng	0.00
21) Naphthalene-d8	10.94	136	90239	20.00	ng	-0.01
38) Acenaphthene-d10	14.76	164	40111	20.00	ng	0.00
63) Phenanthrene-d10	17.50	188	81527	20.00	ng	-0.01
75) Chrysene-d12	21.63	240	93431	20.00	ng	-0.01
86) Perylene-d12	23.96	264	111151	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.66	112	233417	143.18	ng	0.00
7) Phenol-d6	7.30	99	225852	106.08	ng	0.00
23) Nitrobenzene-d5	9.31	82	234776	121.01	ng	-0.01
41) 2,4,6-Tribromophenol	16.25	330	71264	140.08	ng	-0.01
44) 2-Fluorobiphenyl	13.38	172	399128	121.07	ng	-0.01
78) Terphenyl-d14	20.07	244	561704	125.85	ng	-0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM081018\
Data File : BM016310.D
Acq On : 10 Aug 2018 23:41
Operator : SJ/JU
Sample : J4272-08
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
JC-03-080918-D

Quant Time: Aug 11 05:37:05 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM072318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 10 18:44:22 2018
Response via : Initial Calibration

