

Data Path : Z:\HPCHEM1\BNA M\DATA\BM081215\
 Data File : BM002325.D
 Acq On : 11 Aug 2015 14:16
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02023

Quant Time: Aug 12 02:16:29 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM080715MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 10 05:35:03 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	118	-0.02
2 I	Naphthalene-d8	20.000	20.000	0.0	114	-0.02
3	Naphthalene	20.000	19.955	0.2	111	-0.02
4	2-Methylnaphthalene	20.000	19.967	0.2	111	-0.02
5	1-Methylnaphthalene	20.000	19.829	0.9	109	-0.02
6 I	Acenaphthene-d10	20.000	20.000	0.0	110	-0.02
7 S	Acenaphthylene-d8	20.000	20.176	-0.9	108	-0.02
8	Acenaphthylene	20.000	20.174	-0.9	108	-0.02
9 C	Acenaphthene	20.000	20.127	-0.6	107	-0.02
10 S	Fluorene-d10	20.000	20.002	-0.0	107	-0.01
11	Fluorene	20.000	20.238	-1.2	108	-0.02
12 I	Phenanthrene-d10	20.000	20.000	0.0	109	-0.02
13	Phenanthrene	20.000	20.090	-0.4	106	-0.02
14 S	Anthracene-d10	20.000	20.199	-1.0	106	-0.01
15	Anthracene	20.000	20.167	-0.8	106	-0.01
16 C	Fluoranthene	20.000	20.355	-1.8	106	-0.02
17 I	Chrysene-d12	20.000	20.000	0.0	111	-0.01
18 S	Pyrene-d10	20.000	20.054	-0.3	107	-0.01
19	Pyrene	20.000	20.038	-0.2	107	-0.01
20	Benzo(a)anthracene	20.000	20.341	-1.7	109	-0.01
21	Chrysene	20.000	19.963	0.2	108	-0.01
22 I	Perylene-d12	20.000	20.000	0.0	107	-0.01
23	Benzo(b)fluoranthene	20.000	20.754	-3.8	110	-0.01
24	Benzo(k)fluoranthene	20.000	19.932	0.3	103	-0.02
25 S	Benzo(a)pyrene-d12	20.000	20.283	-1.4	107	-0.02
26 C	Benzo(a)pyrene	20.000	20.279	-1.4	106	-0.01
27	Indeno(1,2,3-cd)pyrene	20.000	19.564	2.2	100	-0.02
28	Dibenzo(a,h)anthracene	20.000	19.568	2.2	100	-0.02
29	Benzo(g,h,i)perylene	20.000	19.093	4.5	97	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0