

Data Path : Z:\HPCHEM1\BNA M\Data\BM081215\
 Data File : BM002330.D
 Acq On : 11 Aug 2015 19:51
 Operator : UM/IZ
 Sample : G3162-06
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 D9C73

Quant Time: Aug 12 04:41:15 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM080715MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 12 04:37:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
3	Naphthalene	1.053	0.000	100.0#	0#	-11.82#
4	2-Methylnaphthalene	0.752	0.000	100.0#	0#	-13.36#
5	1-Methylnaphthalene	0.741	0.000	100.0#	0#	-13.58#
6 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
7 S	Acenaphthylene-d8	2.072	3.297	-59.1#	170#	0.00
8	Acenaphthylene	2.151	0.000	100.0#	0#	-15.22#
9 C	Acenaphthene	1.447	0.000	100.0#	0#	-15.54#
10 S	Fluorene-d10	1.441	2.280	-58.2#	169#	0.00
11	Fluorene	1.629	0.005	99.7#	0#	-0.05
12 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
13	Phenanthrene	1.151	0.001	99.9#	0#	0.00
14 S	Anthracene-d10	0.970	1.590	-63.9#	174#	0.00
15	Anthracene	1.189	0.000	100.0#	0#	-0.04
16 C	Fluoranthene	1.275	0.001	99.9#	0#	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	112	0.00
18 S	Pyrene-d10	0.970	1.605	-65.5#	179#	0.00
19	Pyrene	1.293	0.001	99.9#	0#	0.00
20	Benzo(a)anthracene	1.221	0.003	99.8#	0#	0.01
21	Chrysene	1.157	0.001	99.9#	0#	0.00
22 I	Perylene-d12	1.000	1.000	0.0	104	0.00
23	Benzo(b)fluoranthene	1.243	0.000	100.0#	0#	0.00
24	Benzo(k)fluoranthene	1.194	0.001	99.9#	0#	0.00
25 S	Benzo(a)pyrene-d12	1.062	1.770	-66.7#	171#	0.00
26 C	Benzo(a)pyrene	1.189	0.006	99.5#	1#	-0.06
27	Indeno(1,2,3-cd)pyrene	1.363	0.000	100.0#	0#	-27.82#
28	Dibenzo(a,h)anthracene	1.155	0.000	100.0#	0#	-27.83#
29	Benzo(g,h,i)perylene	1.150	0.000	100.0#	0#	-28.70#

(#) = Out of Range

SPCC's out = 0 CCC's out = 3