

Data Path : Z:\HPCHEM1\BNA M\DATA\BM081215\
 Data File : BM002336.D
 Acq On : 11 Aug 2015 23:30
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02024

Quant Time: Aug 12 04:49:50 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM080715MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 12 04:37:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	133	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	127	0.00
3	Naphthalene	1.053	1.053	0.0	125	0.00
4	2-Methylnaphthalene	0.752	0.744	1.1	123	0.00
5	1-Methylnaphthalene	0.741	0.734	0.9	122	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	121	0.00
7 S	Acenaphthylene-d8	2.072	2.081	-0.4	119	0.00
8	Acenaphthylene	2.151	2.175	-1.1	119	0.00
9 C	Acenaphthene	1.447	1.449	-0.1	118	0.00
10 S	Fluorene-d10	1.441	1.433	0.6	117	0.00
11	Fluorene	1.629	1.629	0.0	118	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
13	Phenanthrene	1.151	1.158	-0.6	114	0.00
14 S	Anthracene-d10	0.970	0.986	-1.6	115	0.00
15	Anthracene	1.189	1.196	-0.6	113	0.00
16 C	Fluoranthene	1.275	1.264	0.9	111	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
18 S	Pyrene-d10	0.970	1.024	-5.6	112	0.00
19	Pyrene	1.293	1.364	-5.5	111	0.00
20	Benzo(a)anthracene	1.221	1.237	-1.3	108	0.00
21	Chrysene	1.157	1.149	0.7	106	0.00
22 I	Perylene-d12	1.000	1.000	0.0	88	0.00
23	Benzo(b)fluoranthene	1.243	1.337	-7.6	94	0.00
24	Benzo(k)fluoranthene	1.194	1.304	-9.2	93	0.00
25 S	Benzo(a)pyrene-d12	1.062	1.079	-1.6	88	0.00
26 C	Benzo(a)pyrene	1.189	1.218	-2.4	88	0.00
27	Indeno(1,2,3-cd)pyrene	1.363	1.157	15.1	71	-0.01
28	Dibenzo(a,h)anthracene	1.155	0.985	14.7	72	0.00
29	Benzo(g,h,i)perylene	1.150	0.931	19.0	68	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0