

Data Path : Z:\HPCHEM1\BNA M\DATA\BM081215\
 Data File : BM002351.D
 Acq On : 12 Aug 2015 08:32
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02025

Quant Time: Aug 12 09:12:26 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM080715MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 12 04:50:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	156#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	141	0.00
3	Naphthalene	1.053	1.050	0.3	137	0.00
4	2-Methylnaphthalene	0.752	0.719	4.4	131	0.00
5	1-Methylnaphthalene	0.741	0.707	4.6	130	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	125	0.00
7 S	Acenaphthylene-d8	2.072	2.089	-0.8	122	0.00
8	Acenaphthylene	2.151	2.172	-1.0	122	0.00
9 C	Acenaphthene	1.447	1.450	-0.2	121	0.00
10 S	Fluorene-d10	1.441	1.404	2.6	118	0.00
11	Fluorene	1.629	1.595	2.1	118	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	115	0.00
13	Phenanthrene	1.151	1.157	-0.5	112	0.00
14 S	Anthracene-d10	0.970	0.980	-1.0	112	0.00
15	Anthracene	1.189	1.192	-0.3	111	0.00
16 C	Fluoranthene	1.275	1.150	9.8	99	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	82	0.00
18 S	Pyrene-d10	0.970	1.185	-22.2#	97	0.00
19	Pyrene	1.293	1.570	-21.4#	96	0.00
20	Benzo(a)anthracene	1.221	1.238	-1.4	81	0.00
21	Chrysene	1.157	1.159	-0.2	80	0.00
22 I	Perylene-d12	1.000	1.000	0.0	70	0.00
23	Benzo(b)fluoranthene	1.243	1.285	-3.4	72	0.00
24	Benzo(k)fluoranthene	1.194	1.215	-1.8	69	0.00
25 S	Benzo(a)pyrene-d12	1.062	1.070	-0.8	70	0.00
26 C	Benzo(a)pyrene	1.189	1.200	-0.9	69	0.00
27	Indeno(1,2,3-cd)pyrene	1.363	1.384	-1.5	68	0.00
28	Dibenzo(a,h)anthracene	1.155	1.180	-2.2	68	0.00
29	Benzo(g,h,i)perylene	1.150	1.165	-1.3	68	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0