

Data Path : Z:\HPCHEM1\BNA M\DATA\BM081215\
 Data File : BM002364.D
 Acq On : 12 Aug 2015 17:01
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02027

Quant Time: Aug 12 18:23:55 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM080715MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 12 17:25:11 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	171#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	151#	0.00
3	Naphthalene	1.053	1.048	0.5	147	0.01
4	2-Methylnaphthalene	0.752	0.707	6.0	138	0.01
5	1-Methylnaphthalene	0.741	0.689	7.0	136	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	126	0.00
7 S	Acenaphthylene-d8	2.072	2.097	-1.2	124	0.00
8	Acenaphthylene	2.151	2.158	-0.3	123	0.00
9 C	Acenaphthene	1.447	1.448	-0.1	122	0.00
10 S	Fluorene-d10	1.441	1.368	5.1	117	0.00
11	Fluorene	1.629	1.537	5.6	115	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
13	Phenanthrene	1.151	1.154	-0.3	99	0.01
14 S	Anthracene-d10	0.970	0.976	-0.6	99	0.00
15	Anthracene	1.189	1.186	0.3	98	0.00
16 C	Fluoranthene	1.275	1.028	19.4	79	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	69	0.00
18 S	Pyrene-d10	0.970	1.119	-15.4	77	0.00
19	Pyrene	1.293	1.479	-14.4	76	0.00
20	Benzo(a)anthracene	1.221	1.245	-2.0	68	0.00
21	Chrysene	1.157	1.173	-1.4	68	0.00
22 I	Perylene-d12	1.000	1.000	0.0	69	0.02
23	Benzo(b)fluoranthene	1.243	1.229	1.1	68	0.02
24	Benzo(k)fluoranthene	1.194	1.186	0.7	66	0.01
25 S	Benzo(a)pyrene-d12	1.062	1.059	0.3	67	0.02
26 C	Benzo(a)pyrene	1.189	1.195	-0.5	68	0.02
27	Indeno(1,2,3-cd)pyrene	1.363	1.430	-4.9	69	0.03
28	Dibenzo(a,h)anthracene	1.155	1.214	-5.1	69	0.03
29	Benzo(g,h,i)perylene	1.150	1.202	-4.5	68	0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0