

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM081215\
 Data File : BM002378.D
 Acq On : 13 Aug 2015 01:36
 Operator : UM/IZ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02028

Quant Time: Aug 13 02:12:50 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080715MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 13 00:45:03 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	122	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
3	Naphthalene	1.053	1.044	0.9	106	0.00
4	2-Methylnaphthalene	0.752	0.714	5.1	101	0.00
5	1-Methylnaphthalene	0.741	0.699	5.7	100	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
7 S	Acenaphthylene-d8	2.072	2.094	-1.1	92	0.00
8	Acenaphthylene	2.151	2.174	-1.1	92	0.00
9 C	Acenaphthene	1.447	1.456	-0.6	91	0.00
10 S	Fluorene-d10	1.441	1.386	3.8	88	0.00
11	Fluorene	1.629	1.568	3.7	88	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	80	0.00
13	Phenanthrene	1.151	1.153	-0.2	77	0.00
14 S	Anthracene-d10	0.970	0.974	-0.4	77	0.00
15	Anthracene	1.189	1.188	0.1	77	0.00
16 C	Fluoranthene	1.275	1.042	18.3	62	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	52	0.00
18 S	Pyrene-d10	0.970	1.150	-18.6	60	0.00
19	Pyrene	1.293	1.542	-19.3	60	0.00
20	Benzo(a)anthracene	1.221	1.241	-1.6	51	0.00
21	Chrysene	1.157	1.152	0.4	51	0.00
22 I	Perylene-d12	1.000	1.000	0.0	49	0.00
23	Benzo(b)fluoranthene	1.243	1.222	1.7	48	0.00
24	Benzo(k)fluoranthene	1.194	1.225	-2.6	49	0.00
25 S	Benzo(a)pyrene-d12	1.062	1.073	-1.0	49	0.00
26 C	Benzo(a)pyrene	1.189	1.190	-0.1	48	0.00
27	Indeno(1,2,3-cd)pyrene	1.363	1.419	-4.1	49	0.00
28	Dibenzo(a,h)anthracene	1.155	1.205	-4.3	49	0.00
29	Benzo(g,h,i)perylene	1.150	1.190	-3.5	49	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0