

Data Path : Z:\HPCHEM1\BNA M\DATA\BM081216\
 Data File : BM007047.D
 Acq On : 12 Aug 2016 17:28
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02060

Quant Time: Aug 15 15:53:53 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM081116.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 12 23:03:04 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2	1,4-Dioxane	0.476	0.507	-6.5	131	0.00
3 S	1,4-Dioxane-d8	0.458	0.528	-15.3	136	0.00
4	Benzaldehyde	0.974	1.104	-13.3	126	0.00
5 S	Phenol-d5	1.620	1.617	0.2	121	0.00
6	Phenol	1.707	1.686	1.2	120	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.929	0.935	-0.6	120	0.00
8	Bis(2-Chloroethyl)ether	1.259	1.275	-1.3	120	0.00
9 S	2-Chlorophenol-d4	1.290	1.331	-3.2	124	0.00
10	2-Chlorophenol	1.320	1.342	-1.7	120	0.00
11	2-Methylphenol	1.283	1.234	3.8	118	0.00
12	2,2'-oxybis(1-Chloropropane	1.443	1.432	0.8	121	0.00
13 S	4-Methylphenol-d8	1.333	1.262	5.3	117	0.00
14	Acetophenone	2.053	1.957	4.7	115	0.00
15 P	N-Nitroso-di-n-propylamine	1.068	1.006	5.8	113	0.00
16	4-Methylphenol	1.401	1.324	5.5	117	0.00
17	Hexachloroethane	0.551	0.560	-1.6	122	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
19 S	Nitrobenzene-d5	0.145	0.151	-4.1	117	0.00
20	Nitrobenzene	0.374	0.380	-1.6	116	0.00
21	Isophorone	0.700	0.696	0.6	114	0.00
22 S	2-Nitrophenol-d4	0.155	0.164	-5.8	120	0.00
23 C	2-Nitrophenol	0.170	0.183	-7.6	121	0.00
24	2,4-Dimethylphenol	0.389	0.390	-0.3	114	0.00
25	Bis(2-Chloroethoxy)methane	0.420	0.423	-0.7	116	0.00
26 S	2,4-Dichlorophenol-d3	0.325	0.332	-2.2	118	0.00
27 C	2,4-Dichlorophenol	0.324	0.328	-1.2	116	0.00
28	Naphthalene	0.991	0.974	1.7	113	0.00
29 S	4-Chloroaniline-d4	0.387	0.388	-0.3	110	0.00
30	4-Chloroaniline	0.397	0.391	1.5	107	0.00
31 C	Hexachlorobutadiene	0.239	0.244	-2.1	117	0.00
32	Caprolactam	0.110	0.104	5.5	112	0.00
33 C	4-Chloro-3-methylphenol	0.354	0.351	0.8	114	0.00
34	2-Methylnaphthalene	0.764	0.741	3.0	111	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
36	1,2,4,5-Tetrachlorobenzene	0.707	0.757	-7.1	111	0.00
37	Hexachlorocyclopentadiene	0.450	0.475	-5.6	110	0.00
38 C	2,4,6-Trichlorophenol	0.454	0.489	-7.7	112	0.00
39	2,4,5-Trichlorophenol	0.475	0.505	-6.3	110	0.00
40	1,1'-Biphenyl	1.605	1.682	-4.8	110	0.00
41	2-Chloronaphthalene	1.259	1.311	-4.1	109	0.00
42	2-Nitroaniline	0.361	0.382	-5.8	110	0.00
43 S	Dimethylphthalate-d6	1.653	1.658	-0.3	106	0.00
44	Dimethylphthalate	1.652	1.654	-0.1	106	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM081216\
 Data File : BM007047.D
 Acq On : 12 Aug 2016 17:28
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02060

Quant Time: Aug 15 15:53:53 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM081116.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 12 23:03:04 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.321	0.328	-2.2	108	0.00
46 S	Acenaphthylene-d8	1.985	2.060	-3.8	108	0.00
47	Acenaphthylene	2.043	2.083	-2.0	106	0.00
48	3-Nitroaniline	0.317	0.296	6.6	94	0.00
49 C	Acenaphthene	1.327	1.324	0.2	105	0.00
50	2,4-Dinitrophenol	0.190	0.157	17.4	101	0.00
51 S	4-Nitrophenol-d4	0.296	0.290	2.0	105	0.00
52	4-Nitrophenol	0.301	0.286	5.0	102	0.00
53	Dibenzofuran	1.969	1.949	1.0	105	0.00
54	2,4-Dinitrotoluene	0.474	0.478	-0.8	106	0.00
55	2,3,4,6-Tetrachlorophenol	0.459	0.467	-1.7	106	0.00
56	Diethylphthalate	1.742	1.674	3.9	104	0.00
57 S	Fluorene-d10	1.445	1.446	-0.1	106	0.00
58	Fluorene	1.588	1.569	1.2	103	0.00
59	4-Chlorophenyl-phenylether	0.838	0.826	1.4	103	0.00
60	4-Nitroaniline	0.372	0.280	24.7#	78	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.111	0.100	9.9	99	0.00
63	4,6-Dinitro-2-methylphenol	0.120	0.110	8.3	97	0.00
64	N-Nitrosodiphenylamine	0.566	0.588	-3.9	102	0.00
65	4-Bromophenyl-phenylether	0.227	0.239	-5.3	107	0.00
66	Hexachlorobenzene	0.256	0.268	-4.7	105	0.00
67	Atrazine	0.228	0.238	-4.4	102	0.00
68 C	Pentachlorophenol	0.158	0.168	-6.3	111	0.00
69	Phenanthrene	1.076	1.084	-0.7	101	0.00
70 S	Anthracene-d10	0.927	0.956	-3.1	103	0.00
71	Anthracene	1.104	1.118	-1.3	100	0.00
72	Carbazole	0.951	0.987	-3.8	99	0.00
73	Di-n-butylphthalate	1.163	1.258	-8.2	107	0.00
74 C	Fluoranthene	1.305	1.393	-6.7	100	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
76 S	Pyrene-d10	0.844	0.896	-6.2	99	0.00
77	Pyrene	1.106	1.151	-4.1	97	0.00
78	Butylbenzylphthalate	0.430	0.487	-13.3	104	0.00
79	3,3'-Dichlorobenzidine	0.386	0.325	15.8	68	0.00
80	Benzo(a)anthracene	1.182	1.211	-2.5	95	0.00
81	Bis(2-ethylhexyl)phthalate	0.636	0.713	-12.1	102	0.00
82	Chrysene	1.080	1.103	-2.1	94	0.00
83 I	Perylene-d12	1.000	1.000	0.0	95	0.00
84	Di-n-octyl phthalate	1.032	1.140	-10.5	99	0.00
85	Benzo(b)fluoranthene	1.205	1.214	-0.7	93	0.00
86	Benzo(k)fluoranthene	1.131	1.136	-0.4	94	0.00
87 S	Benzo(a)pyrene-d12	0.915	0.932	-1.9	95	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM081216\
Data File : BM007047.D
Acq On : 12 Aug 2016 17:28
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02060

Quant Time: Aug 15 15:53:53 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM081116.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 12 23:03:04 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.142	1.160	-1.6	94	0.00
89	Indeno(1,2,3-cd)pyrene	1.400	1.411	-0.8	93	0.00
90	Dibenzo(a,h)anthracene	1.175	1.192	-1.4	93	0.00
91	Benzo(a,h,i)perylene	1.183	1.189	-0.5	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0