

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081219\
 Data File : BM022083.D
 Acq On : 13 Aug 2019 23:05
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.456

Quant Time: Aug 14 07:22:46 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM080319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 07:18:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	64	0.00
2 I	Naphthalene-d8	0.400	0.400	0.0	76	0.01
3	Naphthalene	0.400	0.392	2.0	75	0.02
4 SURR	2-Methylnaphthalene-d10	0.400	0.398	0.5	77	0.02
5	2-Methylnaphthalene	0.400	0.404	-1.0	78	0.02
6 I	Acenaphthene-d10	0.400	0.400	0.0	74	0.00
7	Acenaphthylene	0.400	0.381	4.8	69	0.00
8 C	Acenaphthene	0.400	0.414	-3.5	76	0.00
9	Fluorene	0.400	0.390	2.5	71	0.01
10 I	Phenanthrene-d10	0.400	0.400	0.0	70	0.00
11	Pentachlorophenol	0.400	0.388	3.0	75	0.00
12	Phenanthrene	0.400	0.394	1.5	68	0.00
13	Anthracene	0.400	0.401	-0.3	70	0.01
14 SURR	Fluoranthene-d10	0.400	0.367	8.3	63	0.00
15 C	Fluoranthene	0.400	0.363	9.3	63	0.00
16 I	Chrysene-d12	0.400	0.400	0.0	66	0.00
17	Pyrene	0.400	0.391	2.3	63	0.00
18	Benzo(a)anthracene	0.400	0.397	0.8	65	0.00
19	Chrysene	0.400	0.384	4.0	63	0.00
20 I	Perylene-d12	0.400	0.400	0.0	61	0.00
21	Benzo(b)fluoranthene	0.400	0.436	-9.0	65	0.00
22	Benzo(k)fluoranthene	0.400	0.431	-7.7	65	0.00
23 C	Benzo(a)pyrene	0.400	0.435	-8.7	66	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.442	-10.5	67	0.00
25	Dibenzo(a,h)anthracene	0.400	0.435	-8.7	66	0.00
26	Benzo(g,h,i)perylene	0.400	0.428	-7.0	64	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0