


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM081220\  
 Data File : BM027038.D  
 Acq On : 12 Aug 2020 08:40  
 Operator : JU/CG  
 Sample : DFTPP24  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP24

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SOM-EPA-BM080720MA.M  
 Title : SVOA CALIBRATION  
 Last Update : Wed Aug 12 11:04:55 2020



AutoFind: Scans 2465, 2466, 2467; Background Corrected with Scan 2459

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 30.6      | 70842   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.8       | 2118    | PASS             |
| 69          | 198          | 0.00         | 100          | 50.2      | 116437  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.5       | 547     | PASS             |
| 127         | 198          | 10           | 80           | 46.1      | 106728  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.6       | 1485    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 231722  | PASS             |
| 199         | 198          | 5            | 9            | 6.8       | 15692   | PASS             |
| 275         | 198          | 10           | 60           | 30.1      | 69845   | PASS             |
| 365         | 198          | 1            | 100          | 5.2       | 11959   | PASS             |
| 441         | 198          | 0.01         | 100          | 12.8      | 29586   | PASS             |
| 442         | 198          | 50           | 100          | 84.0      | 194666  | PASS             |
| 443         | 442          | 15           | 24           | 19.0      | 37021   | PASS             |

| m/z    | Abundance |
|--------|-----------|
| 36.90  | 282.0     |
| 38.00  | 909.0     |
| 39.00  | 3804.0    |
| 40.00  | 478.0     |
| 43.90  | 703.0     |
| 49.10  | 341.0     |
| 50.00  | 13507.0   |
| 51.00  | 47240.0   |
| 52.10  | 2296.0    |
| 52.90  | 179.0     |
| 55.00  | 367.0     |
| 56.00  | 1488.0    |
| 57.00  | 3835.0    |
| 58.10  | 226.0     |
| 61.00  | 810.0     |
| 61.90  | 1109.0    |
| 63.00  | 3005.0    |
| 63.90  | 452.0     |
| 65.00  | 1530.0    |
| 68.00  | 1416.0    |
| 69.00  | 79048.0   |
| 70.00  | 454.0     |
| 73.00  | 1182.0    |
| 74.00  | 9093.0    |
| 75.00  | 13280.0   |
| 76.10  | 4436.0    |
| 77.10  | 82320.0   |
| 78.10  | 5889.0    |
| 79.00  | 6216.0    |
| 80.00  | 4696.0    |
| 81.00  | 6513.0    |
| 82.00  | 1816.0    |
| 83.00  | 1190.0    |
| 83.90  | 185.0     |
| 85.00  | 959.0     |
| 86.00  | 2088.0    |
| 86.90  | 747.0     |
| 87.90  | 305.0     |
| 91.00  | 1730.0    |
| 92.00  | 1334.0    |
| 93.00  | 11437.0   |
| 94.00  | 1032.0    |
| 96.10  | 350.0     |
| 96.90  | 185.0     |
| 98.00  | 9280.0    |
| 99.00  | 7063.0    |
| 100.00 | 713.0     |
| 101.00 | 3545.0    |
| 102.90 | 1115.0    |
| 104.00 | 2266.0    |
| 105.00 | 2074.0    |
| 105.90 | 559.0     |
| 107.00 | 26120.0   |
| 108.00 | 3532.0    |
| 108.90 | 661.0     |
| 110.00 | 42208.0   |
| 111.00 | 6170.0    |
| 111.90 | 1094.0    |
| 112.90 | 291.0     |
| 115.90 | 1776.0    |
| 117.00 | 30184.0   |
| 118.10 | 1737.0    |
| 120.00 | 344.0     |
| 122.00 | 2020.0    |
| 123.00 | 2952.0    |
| 123.90 | 1131.0    |
| 125.00 | 963.0     |
| 126.10 | 431.0     |
| 127.00 | 77912.0   |
| 128.00 | 6138.0    |
| 129.00 | 33728.0   |
| 130.00 | 2770.0    |
| 131.10 | 798.0     |
| 132.00 | 199.0     |
| 133.00 | 204.0     |
| 133.90 | 1031.0    |
| 135.00 | 3311.0    |
| 136.00 | 1017.0    |
| 137.00 | 1899.0    |
| 137.70 | 367.0     |
| 138.80 | 172.0     |
| 140.00 | 490.0     |
| 141.00 | 4814.0    |
| 142.00 | 1335.0    |
| 143.00 | 1597.0    |
| 144.00 | 294.0     |
| 146.00 | 1208.0    |
| 146.90 | 2909.0    |
| 148.00 | 7016.0    |
| 149.00 | 1492.0    |
| 150.10 | 185.0     |
| 151.00 | 586.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP24

|        |          |
|--------|----------|
| 151.90 | 519.0    |
| 152.90 | 1686.0   |
| 153.90 | 1509.0   |
| 155.00 | 2806.0   |
| 156.00 | 4241.0   |
| 157.10 | 925.0    |
| 158.00 | 884.0    |
| 159.00 | 836.0    |
| 160.00 | 1667.0   |
| 161.00 | 2599.0   |
| 162.00 | 875.0    |
| 163.20 | 199.0    |
| 163.80 | 302.0    |
| 165.00 | 2110.0   |
| 166.00 | 1892.0   |
| 167.00 | 11793.0  |
| 168.00 | 6146.0   |
| 169.00 | 1171.0   |
| 169.90 | 310.0    |
| 170.90 | 374.0    |
| 172.00 | 1069.0   |
| 173.00 | 1321.0   |
| 174.00 | 2651.0   |
| 175.00 | 4343.0   |
| 176.00 | 1046.0   |
| 177.00 | 1911.0   |
| 177.90 | 740.0    |
| 179.00 | 8788.0   |
| 180.00 | 6317.0   |
| 181.00 | 2392.0   |
| 182.10 | 476.0    |
| 182.70 | 240.0    |
| 183.10 | 153.0    |
| 183.90 | 684.0    |
| 185.00 | 4266.0   |
| 186.00 | 30240.0  |
| 187.00 | 8535.0   |
| 188.00 | 994.0    |
| 189.00 | 2539.0   |
| 190.10 | 275.0    |
| 191.00 | 918.0    |
| 191.90 | 2724.0   |
| 193.00 | 3612.0   |
| 194.00 | 514.0    |
| 195.10 | 397.0    |
| 196.00 | 5153.0   |
| 197.00 | 1117.0   |
| 197.90 | 191872.0 |
| 198.90 | 11147.0  |
| 200.00 | 957.0    |
| 200.90 | 332.0    |
| 201.50 | 946.0    |
| 202.90 | 1638.0   |
| 204.00 | 9121.0   |
| 205.00 | 14023.0  |
| 206.00 | 57472.0  |
| 207.00 | 8673.0   |
| 208.00 | 2193.0   |
| 209.00 | 886.0    |
| 210.00 | 1023.0   |
| 210.90 | 2009.0   |
| 211.90 | 461.0    |
| 213.00 | 190.0    |
| 214.90 | 787.0    |
| 216.00 | 1202.0   |
| 217.00 | 17208.0  |
| 218.00 | 2246.0   |
| 220.00 | 249.0    |
| 221.00 | 7790.0   |
| 221.90 | 1843.0   |
| 223.00 | 4020.0   |
| 224.00 | 34032.0  |
| 225.00 | 9213.0   |
| 226.00 | 766.0    |
| 227.00 | 16496.0  |
| 228.00 | 2385.0   |
| 229.00 | 3010.0   |
| 230.10 | 516.0    |
| 231.00 | 934.0    |
| 232.80 | 220.0    |
| 234.00 | 849.0    |
| 234.90 | 976.0    |
| 236.00 | 903.0    |
| 237.00 | 1015.0   |
| 238.00 | 295.0    |
| 239.00 | 620.0    |
| 240.00 | 551.0    |
| 241.00 | 978.0    |
| 242.00 | 1618.0   |
| 243.00 | 1925.0   |
| 244.00 | 25464.0  |
| 245.00 | 3110.0   |
| 246.00 | 6095.0   |
| 247.00 | 1139.0   |
| 248.00 | 363.0    |
| 249.00 | 821.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP24

|        |          |
|--------|----------|
| 249.80 | 252.0    |
| 250.90 | 197.0    |
| 251.90 | 291.0    |
| 253.00 | 695.0    |
| 254.10 | 947.0    |
| 255.00 | 145024.0 |
| 256.00 | 21288.0  |
| 257.00 | 1794.0   |
| 258.00 | 10280.0  |
| 258.90 | 1584.0   |
| 259.90 | 241.0    |
| 261.00 | 249.0    |
| 264.20 | 219.0    |
| 265.00 | 4555.0   |
| 266.00 | 556.0    |
| 271.00 | 414.0    |
| 271.90 | 513.0    |
| 273.00 | 4398.0   |
| 274.00 | 11026.0  |
| 275.00 | 61200.0  |
| 276.00 | 8283.0   |
| 277.00 | 6301.0   |
| 277.90 | 1195.0   |
| 279.10 | 255.0    |
| 281.10 | 495.0    |
| 281.90 | 227.0    |
| 283.00 | 751.0    |
| 284.10 | 494.0    |
| 285.00 | 1227.0   |
| 286.00 | 195.0    |
| 288.90 | 344.0    |
| 290.00 | 163.0    |
| 292.00 | 300.0    |
| 293.00 | 1250.0   |
| 293.80 | 312.0    |
| 294.10 | 370.0    |
| 294.90 | 372.0    |
| 296.00 | 24416.0  |
| 296.90 | 2486.0   |
| 301.10 | 320.0    |
| 302.00 | 454.0    |
| 303.10 | 2490.0   |
| 303.90 | 568.0    |
| 308.10 | 485.0    |
| 309.00 | 320.0    |
| 309.90 | 297.0    |
| 314.10 | 1107.0   |
| 315.00 | 2835.0   |
| 315.90 | 1369.0   |
| 317.00 | 241.0    |
| 321.00 | 589.0    |
| 322.10 | 305.0    |
| 323.10 | 7024.0   |
| 323.90 | 1131.0   |
| 327.00 | 1641.0   |
| 328.10 | 681.0    |
| 329.00 | 162.0    |
| 331.90 | 649.0    |
| 333.10 | 726.0    |
| 334.00 | 4260.0   |
| 335.00 | 994.0    |
| 340.10 | 174.0    |
| 341.00 | 702.0    |
| 342.00 | 242.0    |
| 346.00 | 1788.0   |
| 347.10 | 310.0    |
| 352.00 | 2152.0   |
| 353.00 | 1593.0   |
| 354.10 | 2388.0   |
| 355.00 | 435.0    |
| 359.20 | 215.0    |
| 365.00 | 12370.0  |
| 365.90 | 1575.0   |
| 366.90 | 155.0    |
| 371.00 | 509.0    |
| 372.00 | 3580.0   |
| 373.00 | 731.0    |
| 383.00 | 1319.0   |
| 383.90 | 333.0    |
| 390.00 | 576.0    |
| 391.00 | 425.0    |
| 391.90 | 162.0    |
| 401.00 | 167.0    |
| 402.00 | 1527.0   |
| 403.10 | 1957.0   |
| 404.10 | 849.0    |
| 405.00 | 265.0    |
| 421.00 | 1617.0   |
| 422.00 | 1806.0   |
| 423.00 | 11482.0  |
| 424.10 | 2509.0   |
| 425.00 | 267.0    |
| 440.10 | 194.0    |
| 441.10 | 32128.0  |
| 442.10 | 211328.0 |
| 443.10 | 39576.0  |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP24

444.103861.0

445.10303.0

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP24

| m/z    | Abundance |
|--------|-----------|
| 36.90  | 532.0     |
| 38.10  | 1310.0    |
| 39.10  | 7020.0    |
| 40.10  | 659.0     |
| 44.00  | 730.0     |
| 45.00  | 240.0     |
| 49.20  | 681.0     |
| 50.10  | 23072.0   |
| 51.10  | 85232.0   |
| 52.10  | 4947.0    |
| 55.00  | 571.0     |
| 56.00  | 2866.0    |
| 57.00  | 6220.0    |
| 57.90  | 388.0     |
| 61.00  | 1580.0    |
| 62.00  | 1584.0    |
| 63.00  | 4817.0    |
| 64.00  | 596.0     |
| 65.00  | 2170.0    |
| 66.00  | 238.0     |
| 68.00  | 2470.0    |
| 69.00  | 142976.0  |
| 70.10  | 557.0     |
| 73.10  | 1787.0    |
| 74.00  | 17008.0   |
| 75.00  | 22760.0   |
| 76.10  | 7855.0    |
| 77.10  | 142528.0  |
| 78.00  | 9803.0    |
| 79.00  | 11089.0   |
| 80.00  | 8159.0    |
| 81.00  | 11299.0   |
| 82.00  | 3243.0    |
| 83.00  | 1960.0    |
| 84.90  | 1627.0    |
| 86.00  | 3447.0    |
| 87.00  | 1551.0    |
| 88.10  | 759.0     |
| 89.10  | 289.0     |
| 91.00  | 2561.0    |
| 92.00  | 2710.0    |
| 93.00  | 20368.0   |
| 94.00  | 1464.0    |
| 95.00  | 272.0     |
| 96.10  | 584.0     |
| 97.20  | 281.0     |
| 98.00  | 15959.0   |
| 99.00  | 10011.0   |
| 100.00 | 1009.0    |
| 101.00 | 5729.0    |
| 102.10 | 260.0     |
| 103.00 | 1477.0    |
| 104.00 | 4009.0    |
| 104.90 | 3352.0    |
| 106.00 | 1129.0    |
| 107.00 | 43992.0   |
| 108.00 | 6202.0    |
| 109.00 | 1412.0    |
| 110.00 | 67864.0   |
| 111.00 | 11459.0   |
| 112.00 | 1516.0    |
| 112.90 | 643.0     |
| 116.10 | 2308.0    |
| 117.00 | 47872.0   |
| 118.00 | 3029.0    |
| 118.90 | 395.0     |
| 120.00 | 655.0     |
| 121.10 | 231.0     |
| 122.00 | 3004.0    |
| 123.00 | 4207.0    |
| 124.00 | 2066.0    |
| 125.00 | 2050.0    |
| 126.10 | 907.0     |
| 127.00 | 128480.0  |
| 128.00 | 9959.0    |
| 129.00 | 55936.0   |
| 130.00 | 4873.0    |
| 130.90 | 877.0     |
| 132.00 | 489.0     |
| 133.00 | 215.0     |
| 134.00 | 1943.0    |
| 135.00 | 4607.0    |
| 135.90 | 1659.0    |
| 137.00 | 2718.0    |
| 137.90 | 611.0     |
| 138.90 | 349.0     |
| 139.90 | 829.0     |
| 141.00 | 7643.0    |
| 142.00 | 2974.0    |
| 142.90 | 2017.0    |
| 144.00 | 481.0     |
| 144.90 | 453.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP24

|        |          |
|--------|----------|
| 146.00 | 1346.0   |
| 147.00 | 4655.0   |
| 148.00 | 10543.0  |
| 149.00 | 2016.0   |
| 149.90 | 769.0    |
| 151.20 | 900.0    |
| 152.00 | 317.0    |
| 153.00 | 2526.0   |
| 154.00 | 1994.0   |
| 155.00 | 5354.0   |
| 156.00 | 7308.0   |
| 157.00 | 1228.0   |
| 158.00 | 1491.0   |
| 158.90 | 1252.0   |
| 160.00 | 3372.0   |
| 161.00 | 3858.0   |
| 162.10 | 1140.0   |
| 163.00 | 407.0    |
| 164.10 | 435.0    |
| 165.00 | 3283.0   |
| 166.00 | 2274.0   |
| 167.00 | 18520.0  |
| 168.00 | 10339.0  |
| 169.10 | 1652.0   |
| 170.00 | 497.0    |
| 171.00 | 739.0    |
| 172.00 | 1740.0   |
| 173.00 | 2111.0   |
| 174.00 | 3795.0   |
| 175.00 | 6874.0   |
| 176.10 | 1336.0   |
| 177.00 | 3527.0   |
| 178.10 | 1006.0   |
| 179.00 | 13125.0  |
| 180.00 | 9248.0   |
| 181.00 | 3578.0   |
| 181.90 | 781.0    |
| 183.10 | 304.0    |
| 183.90 | 1067.0   |
| 185.00 | 6648.0   |
| 186.00 | 42552.0  |
| 187.00 | 13980.0  |
| 188.10 | 1591.0   |
| 189.00 | 3909.0   |
| 189.90 | 589.0    |
| 190.90 | 1550.0   |
| 192.00 | 4018.0   |
| 193.00 | 4575.0   |
| 194.00 | 885.0    |
| 195.00 | 763.0    |
| 196.00 | 8573.0   |
| 197.10 | 2098.0   |
| 198.00 | 283136.0 |
| 198.90 | 20032.0  |
| 200.00 | 1623.0   |
| 201.50 | 1014.0   |
| 203.00 | 2272.0   |
| 204.00 | 12141.0  |
| 205.00 | 20192.0  |
| 206.00 | 83136.0  |
| 207.00 | 12423.0  |
| 208.00 | 3189.0   |
| 209.10 | 1200.0   |
| 210.00 | 1328.0   |
| 211.00 | 3781.0   |
| 213.00 | 246.0    |
| 214.90 | 1122.0   |
| 216.00 | 1701.0   |
| 217.00 | 24088.0  |
| 218.00 | 3236.0   |
| 220.30 | 621.0    |
| 221.10 | 10127.0  |
| 221.80 | 3760.0   |
| 223.00 | 5191.0   |
| 224.00 | 48864.0  |
| 225.00 | 11854.0  |
| 226.10 | 1441.0   |
| 227.00 | 23920.0  |
| 228.00 | 3363.0   |
| 229.00 | 4482.0   |
| 229.90 | 667.0    |
| 231.00 | 1619.0   |
| 231.90 | 154.0    |
| 232.90 | 376.0    |
| 234.00 | 1726.0   |
| 234.90 | 1816.0   |
| 236.00 | 1165.0   |
| 237.00 | 1964.0   |
| 237.90 | 253.0    |
| 239.00 | 982.0    |
| 239.90 | 735.0    |
| 241.00 | 1030.0   |
| 242.00 | 2477.0   |
| 243.10 | 2751.0   |
| 244.00 | 37352.0  |
| 245.00 | 4995.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP24

|        |          |
|--------|----------|
| 246.00 | 9944.0   |
| 247.00 | 2080.0   |
| 248.00 | 445.0    |
| 248.90 | 1378.0   |
| 249.90 | 272.0    |
| 250.90 | 223.0    |
| 252.10 | 347.0    |
| 253.00 | 1176.0   |
| 254.20 | 1479.0   |
| 255.00 | 206208.0 |
| 256.00 | 30000.0  |
| 256.90 | 2365.0   |
| 258.00 | 14028.0  |
| 259.00 | 2664.0   |
| 259.90 | 306.0    |
| 260.90 | 253.0    |
| 263.90 | 338.0    |
| 265.00 | 6009.0   |
| 266.00 | 907.0    |
| 270.00 | 255.0    |
| 270.90 | 508.0    |
| 272.00 | 609.0    |
| 273.00 | 6612.0   |
| 274.00 | 16205.0  |
| 275.00 | 86896.0  |
| 276.00 | 11196.0  |
| 277.00 | 9060.0   |
| 278.00 | 1740.0   |
| 281.00 | 414.0    |
| 281.90 | 342.0    |
| 283.00 | 1146.0   |
| 283.90 | 583.0    |
| 285.00 | 1330.0   |
| 286.00 | 306.0    |
| 288.90 | 408.0    |
| 289.80 | 189.0    |
| 290.10 | 174.0    |
| 290.90 | 225.0    |
| 292.10 | 264.0    |
| 292.90 | 2003.0   |
| 293.90 | 560.0    |
| 295.00 | 570.0    |
| 296.00 | 30560.0  |
| 297.00 | 3416.0   |
| 297.90 | 239.0    |
| 300.80 | 303.0    |
| 302.00 | 444.0    |
| 303.00 | 3297.0   |
| 304.10 | 1206.0   |
| 305.00 | 174.0    |
| 307.90 | 332.0    |
| 309.00 | 255.0    |
| 310.00 | 312.0    |
| 312.90 | 327.0    |
| 314.00 | 1070.0   |
| 315.00 | 3867.0   |
| 316.00 | 1477.0   |
| 317.00 | 299.0    |
| 321.00 | 1313.0   |
| 322.00 | 535.0    |
| 323.10 | 8906.0   |
| 324.00 | 1595.0   |
| 325.90 | 153.0    |
| 327.00 | 1822.0   |
| 328.00 | 827.0    |
| 332.00 | 788.0    |
| 333.00 | 895.0    |
| 334.00 | 5976.0   |
| 335.10 | 1086.0   |
| 336.00 | 211.0    |
| 341.00 | 1081.0   |
| 342.10 | 283.0    |
| 346.00 | 2030.0   |
| 347.10 | 310.0    |
| 351.00 | 181.0    |
| 352.00 | 2946.0   |
| 353.10 | 1901.0   |
| 354.10 | 2038.0   |
| 355.10 | 775.0    |
| 359.10 | 204.0    |
| 364.20 | 245.0    |
| 365.00 | 14226.0  |
| 366.00 | 2424.0   |
| 370.10 | 153.0    |
| 371.00 | 886.0    |
| 372.10 | 4434.0   |
| 373.00 | 1295.0   |
| 373.90 | 220.0    |
| 383.10 | 1231.0   |
| 384.00 | 406.0    |
| 390.00 | 378.0    |
| 390.90 | 378.0    |
| 392.10 | 188.0    |
| 401.10 | 277.0    |
| 402.00 | 1786.0   |
| 403.10 | 2802.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP24

|        |          |
|--------|----------|
| 404.00 | 885.0    |
| 405.00 | 192.0    |
| 415.00 | 191.0    |
| 421.00 | 1842.0   |
| 422.00 | 2164.0   |
| 423.10 | 13128.0  |
| 424.10 | 3467.0   |
| 425.00 | 498.0    |
| 441.10 | 35488.0  |
| 442.10 | 233920.0 |
| 443.10 | 45488.0  |
| 444.10 | 4176.0   |
| 445.00 | 322.0    |

**Instrument :**  
BNA\_M  
**ClientSampleId :**  
DFTPP24

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 358.0     |
| 38.00  | 1398.0    |
| 39.00  | 6579.0    |
| 39.90  | 772.0     |
| 41.10  | 181.0     |
| 44.00  | 972.0     |
| 45.00  | 177.0     |
| 48.90  | 684.0     |
| 50.10  | 22264.0   |
| 51.10  | 80056.0   |
| 52.00  | 4867.0    |
| 53.20  | 252.0     |
| 55.00  | 707.0     |
| 56.00  | 2923.0    |
| 57.00  | 6226.0    |
| 57.90  | 250.0     |
| 61.00  | 1692.0    |
| 62.10  | 1757.0    |
| 63.00  | 4585.0    |
| 64.10  | 653.0     |
| 65.00  | 2005.0    |
| 67.00  | 189.0     |
| 68.10  | 2468.0    |
| 69.00  | 127288.0  |
| 70.00  | 630.0     |
| 73.00  | 1506.0    |
| 74.00  | 14718.0   |
| 75.00  | 20672.0   |
| 76.00  | 7119.0    |
| 77.10  | 128024.0  |
| 78.10  | 8514.0    |
| 79.00  | 9113.0    |
| 80.00  | 7764.0    |
| 81.00  | 10459.0   |
| 82.00  | 2037.0    |
| 83.00  | 1766.0    |
| 83.90  | 325.0     |
| 85.00  | 1961.0    |
| 86.00  | 3599.0    |
| 87.00  | 1268.0    |
| 88.00  | 444.0     |
| 88.90  | 259.0     |
| 91.00  | 1998.0    |
| 91.90  | 2646.0    |
| 93.00  | 18016.0   |
| 94.00  | 1422.0    |
| 95.00  | 248.0     |
| 96.10  | 646.0     |
| 97.00  | 489.0     |
| 98.00  | 14054.0   |
| 99.00  | 11084.0   |
| 99.90  | 802.0     |
| 101.00 | 4763.0    |
| 101.90 | 207.0     |
| 103.00 | 1557.0    |
| 104.00 | 2964.0    |
| 105.00 | 3426.0    |
| 106.10 | 1284.0    |
| 107.00 | 39656.0   |
| 108.00 | 5594.0    |
| 109.10 | 1187.0    |
| 110.00 | 60032.0   |
| 111.00 | 10448.0   |
| 111.90 | 1184.0    |
| 113.10 | 484.0     |
| 114.80 | 253.0     |
| 116.00 | 2249.0    |
| 117.00 | 40488.0   |
| 118.00 | 3339.0    |
| 118.90 | 508.0     |
| 119.90 | 428.0     |
| 122.00 | 2456.0    |
| 123.00 | 3933.0    |
| 124.00 | 1643.0    |
| 125.00 | 1424.0    |
| 125.90 | 368.0     |
| 127.00 | 113792.0  |
| 128.00 | 8335.0    |
| 129.00 | 48664.0   |
| 130.00 | 4290.0    |
| 131.00 | 735.0     |
| 131.90 | 418.0     |
| 133.00 | 434.0     |
| 134.00 | 1568.0    |
| 135.00 | 3865.0    |
| 135.90 | 1854.0    |
| 137.00 | 2127.0    |
| 137.90 | 445.0     |
| 139.00 | 264.0     |
| 139.90 | 788.0     |
| 141.00 | 7047.0    |
| 142.00 | 2104.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP24

|        |          |
|--------|----------|
| 143.00 | 1505.0   |
| 143.90 | 233.0    |
| 144.90 | 615.0    |
| 146.00 | 1342.0   |
| 147.00 | 3820.0   |
| 148.00 | 10370.0  |
| 149.10 | 1699.0   |
| 149.90 | 507.0    |
| 151.00 | 970.0    |
| 151.70 | 549.0    |
| 153.00 | 2135.0   |
| 154.00 | 1748.0   |
| 155.00 | 4341.0   |
| 156.00 | 5887.0   |
| 157.10 | 643.0    |
| 158.00 | 1610.0   |
| 159.00 | 1090.0   |
| 160.00 | 2323.0   |
| 161.00 | 3347.0   |
| 162.00 | 1160.0   |
| 162.90 | 290.0    |
| 164.10 | 250.0    |
| 165.00 | 2959.0   |
| 166.00 | 2613.0   |
| 167.00 | 15192.0  |
| 168.00 | 8708.0   |
| 168.90 | 1664.0   |
| 169.90 | 559.0    |
| 170.90 | 797.0    |
| 171.90 | 1327.0   |
| 173.00 | 1565.0   |
| 174.00 | 2868.0   |
| 175.00 | 5583.0   |
| 176.00 | 1338.0   |
| 177.00 | 2278.0   |
| 178.00 | 1057.0   |
| 179.00 | 11799.0  |
| 180.00 | 7335.0   |
| 181.00 | 3641.0   |
| 182.10 | 427.0    |
| 182.70 | 175.0    |
| 184.10 | 720.0    |
| 185.10 | 4869.0   |
| 186.00 | 34368.0  |
| 187.00 | 9845.0   |
| 188.00 | 1216.0   |
| 189.00 | 2778.0   |
| 190.00 | 374.0    |
| 191.00 | 1106.0   |
| 192.00 | 3460.0   |
| 193.00 | 4287.0   |
| 194.00 | 1014.0   |
| 194.90 | 316.0    |
| 196.00 | 6689.0   |
| 196.80 | 1090.0   |
| 197.10 | 1241.0   |
| 198.00 | 220160.0 |
| 199.00 | 15899.0  |
| 200.00 | 1320.0   |
| 201.60 | 1126.0   |
| 203.00 | 2202.0   |
| 204.00 | 9993.0   |
| 205.00 | 17408.0  |
| 206.00 | 64976.0  |
| 207.00 | 10089.0  |
| 208.00 | 2802.0   |
| 209.00 | 1044.0   |
| 210.00 | 1173.0   |
| 211.00 | 3095.0   |
| 213.10 | 183.0    |
| 215.00 | 679.0    |
| 216.00 | 1827.0   |
| 217.00 | 20096.0  |
| 217.90 | 2252.0   |
| 219.00 | 259.0    |
| 221.00 | 7458.0   |
| 222.00 | 2348.0   |
| 223.00 | 3971.0   |
| 224.00 | 37696.0  |
| 225.10 | 9313.0   |
| 226.00 | 1204.0   |
| 227.00 | 19048.0  |
| 228.00 | 2592.0   |
| 229.00 | 3545.0   |
| 230.00 | 509.0    |
| 231.00 | 1321.0   |
| 232.00 | 231.0    |
| 232.90 | 282.0    |
| 234.00 | 1126.0   |
| 235.00 | 1468.0   |
| 236.00 | 1098.0   |
| 237.00 | 1712.0   |
| 238.00 | 284.0    |
| 239.00 | 717.0    |
| 240.00 | 535.0    |
| 241.10 | 889.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP24

|        |          |
|--------|----------|
| 242.00 | 2022.0   |
| 243.00 | 1939.0   |
| 244.00 | 26400.0  |
| 245.00 | 4337.0   |
| 246.00 | 6816.0   |
| 247.00 | 1288.0   |
| 247.80 | 376.0    |
| 249.00 | 1040.0   |
| 250.00 | 185.0    |
| 250.80 | 239.0    |
| 252.10 | 324.0    |
| 252.90 | 536.0    |
| 254.10 | 1476.0   |
| 255.00 | 146816.0 |
| 256.00 | 21752.0  |
| 257.00 | 2399.0   |
| 258.00 | 10509.0  |
| 258.90 | 1492.0   |
| 259.90 | 378.0    |
| 260.90 | 374.0    |
| 264.10 | 275.0    |
| 265.00 | 4251.0   |
| 266.00 | 613.0    |
| 267.00 | 174.0    |
| 270.00 | 177.0    |
| 271.00 | 374.0    |
| 271.90 | 385.0    |
| 273.00 | 4950.0   |
| 274.00 | 11164.0  |
| 275.00 | 61440.0  |
| 276.00 | 8620.0   |
| 277.00 | 6694.0   |
| 278.00 | 1049.0   |
| 279.00 | 223.0    |
| 281.10 | 380.0    |
| 281.90 | 260.0    |
| 282.90 | 603.0    |
| 284.10 | 502.0    |
| 285.00 | 973.0    |
| 288.90 | 219.0    |
| 292.00 | 278.0    |
| 293.00 | 1322.0   |
| 294.00 | 284.0    |
| 294.90 | 450.0    |
| 296.00 | 21952.0  |
| 297.00 | 3190.0   |
| 298.00 | 279.0    |
| 300.90 | 289.0    |
| 302.00 | 567.0    |
| 303.00 | 2419.0   |
| 303.90 | 632.0    |
| 307.90 | 237.0    |
| 309.90 | 233.0    |
| 312.90 | 322.0    |
| 314.00 | 1016.0   |
| 315.00 | 2522.0   |
| 316.00 | 1188.0   |
| 316.90 | 206.0    |
| 320.90 | 633.0    |
| 321.80 | 379.0    |
| 323.00 | 5571.0   |
| 324.00 | 1053.0   |
| 327.00 | 1468.0   |
| 328.10 | 889.0    |
| 332.10 | 660.0    |
| 333.10 | 929.0    |
| 334.00 | 3702.0   |
| 335.10 | 1108.0   |
| 335.80 | 302.0    |
| 341.00 | 707.0    |
| 341.90 | 193.0    |
| 346.00 | 1545.0   |
| 346.90 | 284.0    |
| 352.00 | 1485.0   |
| 353.00 | 1306.0   |
| 354.00 | 1582.0   |
| 354.90 | 390.0    |
| 358.90 | 217.0    |
| 365.00 | 9282.0   |
| 366.00 | 1362.0   |
| 371.00 | 519.0    |
| 372.10 | 2897.0   |
| 373.10 | 845.0    |
| 383.10 | 485.0    |
| 384.00 | 261.0    |
| 389.80 | 218.0    |
| 391.10 | 179.0    |
| 391.90 | 208.0    |
| 401.00 | 238.0    |
| 402.00 | 1161.0   |
| 403.00 | 1716.0   |
| 404.10 | 729.0    |
| 404.90 | 286.0    |
| 421.10 | 1074.0   |
| 422.10 | 1195.0   |
| 423.10 | 7854.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP24

|        |          |
|--------|----------|
| 424.00 | 2503.0   |
| 425.00 | 249.0    |
| 440.00 | 203.0    |
| 441.10 | 21144.0  |
| 442.10 | 138752.0 |
| 443.10 | 26000.0  |
| 444.00 | 2448.0   |
| 445.00 | 245.0    |

**Instrument :**  
BNA\_M  
**ClientSampleId :**  
DFTPP24

Scan 2459 (17.451 min): BM027038.D  
DFTPP24

| m/z    | Abundance |
|--------|-----------|
| 40.10  | 419.0     |
| 44.00  | 726.0     |
| 73.00  | 321.0     |
| 191.10 | 200.0     |
| 207.00 | 1112.0    |
| 208.00 | 202.0     |
| 281.00 | 301.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP24