

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081224\
 Data File : BM047152.D
 Acq On : 12 Aug 2024 11:22
 Operator : RC/JU
 Sample : PB162520BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB162520BL

Quant Time: Aug 12 12:05:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Aug 11 23:47:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.375	152	224656	20.000	ng	0.00
21) Naphthalene-d8	10.128	136	856520	20.000	ng	0.00
39) Acenaphthene-d10	14.027	164	555199	20.000	ng	0.00
64) Phenanthrene-d10	16.792	188	1169777	20.000	ng	0.00
76) Chrysene-d12	21.027	240	971602	20.000	ng	0.00
86) Perylene-d12	23.721	264	1088063	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.028	112	1733060	137.842	ng	0.00
7) Phenol-d6	6.587	99	2281499	131.694	ng	0.00
23) Nitrobenzene-d5	8.516	82	1477007	86.831	ng	0.00
42) 2,4,6-Tribromophenol	15.539	330	850103	132.052	ng	0.00
45) 2-Fluorobiphenyl	12.639	172	3088156	85.799	ng	0.00
79) Terphenyl-d14	19.456	244	4695356	103.551	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081224\
Data File : BM047152.D
Acq On : 12 Aug 2024 11:22
Operator : RC/JU
Sample : PB162520BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB162520BL

Quant Time: Aug 12 12:05:51 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM081024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Aug 11 23:47:58 2024
Response via : Initial Calibration

