

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM081315\
 Data File : BM002391.D
 Acq On : 13 Aug 2015 10:14
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02030

Quant Time: Aug 14 01:57:49 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080715MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 13 05:08:41 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	194#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	171#	0.00
3	Naphthalene	1.053	1.041	1.1	166#	0.00
4	2-Methylnaphthalene	0.752	0.709	5.7	158#	0.00
5	1-Methylnaphthalene	0.741	0.686	7.4	154#	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	141	0.00
7 S	Acenaphthylene-d8	2.072	2.082	-0.5	138	0.00
8	Acenaphthylene	2.151	2.171	-0.9	138	0.00
9 C	Acenaphthene	1.447	1.451	-0.3	137	0.00
10 S	Fluorene-d10	1.441	1.361	5.6	130	0.00
11	Fluorene	1.629	1.531	6.0	129	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
13	Phenanthrene	1.151	1.161	-0.9	109	0.00
14 S	Anthracene-d10	0.970	0.966	0.4	107	0.00
15	Anthracene	1.189	1.181	0.7	107	0.00
16 C	Fluoranthene	1.275	1.037	18.7	87	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	78	0.00
18 S	Pyrene-d10	0.970	1.085	-11.9	85	0.00
19	Pyrene	1.293	1.441	-11.4	84	0.00
20	Benzo(a)anthracene	1.221	1.226	-0.4	76	0.00
21	Chrysene	1.157	1.151	0.5	76	0.00
22 I	Perylene-d12	1.000	1.000	0.0	77	0.00
23	Benzo(b)fluoranthene	1.243	1.243	0.0	76	0.00
24	Benzo(k)fluoranthene	1.194	1.194	0.0	74	0.00
25 S	Benzo(a)pyrene-d12	1.062	1.067	-0.5	76	0.00
26 C	Benzo(a)pyrene	1.189	1.195	-0.5	75	0.00
27	Indeno(1,2,3-cd)pyrene	1.363	1.418	-4.0	76	0.00
28	Dibenzo(a,h)anthracene	1.155	1.198	-3.7	76	0.00
29	Benzo(g,h,i)perylene	1.150	1.190	-3.5	75	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0